

RECEIVED

FEB 09 2016

Florida Department of Law Enforcement

Alcohol Testing Program

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORMOND BEACH P.D.
Time of Inspection: 11:13

Date of Inspection: 02/08/2016

Serial Number: 80-001156
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG511701 Exp: 04/27/2017
0.000	0.052	0.082	0.208	0.078
0.000	0.053	0.082	0.209	0.078
0.000	0.053	0.083	0.209	0.077
0.000	0.053	0.083	0.208	0.078
0.000	0.053	0.083	0.208	0.078
0.000	0.053	0.083	0.209	0.078
0.000	0.053	0.083	0.209	0.078
0.000	0.053	0.083	0.208	0.078
0.000	0.053	0.082	0.209	0.078
0.000	0.053	0.083	0.208	0.077
Standard Deviations	0.0003	0.0004	0.0005	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

JAKE L SHANAHAN

02/08/2016
Date