

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

AUG 15 2016

FDLE

Alcohol Testing

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

BK

Agency: COLLIER COUNTY SO
Time of Inspection: 11:38

Date of Inspection: 08/04/2016

Serial Number: 80-001202
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG600504 Exp: 01/05/2018
0.000	0.051	0.081	0.203	0.079
0.000	0.052	0.081	0.203	0.079
0.000	0.052	0.081	0.203	0.079
0.000	0.051	0.082	0.203	0.079
0.000	0.052	0.082	0.203	0.079
0.000	0.051	0.081	0.203	0.079
0.000	0.051	0.082	0.203	0.078
0.000	0.052	0.082	0.204	0.079
0.000	0.051	0.082	0.204	0.079
0.000	0.052	0.081	0.204	0.078
Standard Deviations	0.0005	0.0005	0.0004	0.0004

0927

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

08/04/2016
Date