

RECEIVED

✓ BCC

Florida Department of Law Enforcement MAR 10 2016
 Alcohol Testing Program FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: AUBURNDALE PD
 Time of Inspection: 15:47

Date of Inspection: 03/09/2016

Serial Number: 80-001231
 Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

BK

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG605301 Exp: 02/22/2018
0.000	0.051	0.081	0.198	0.084
0.000	0.051	0.081	0.199	0.083
0.000	0.051	0.082	0.199	0.083
0.000	0.051	0.081	0.200	0.082
0.000	0.052	0.082	0.199	0.082
0.000	0.051	0.082	0.200	0.081
0.000	0.051	0.081	0.203	0.081
0.000	0.051	0.082	0.201	0.081
0.000	0.053	0.082	0.200	0.081
0.000	0.052	0.082	0.200	0.081

Standard Deviations	0.0006	0.0005	0.0013	0.0011
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0008 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Handwritten Signature]

DANIELLE M BELL

Signature and Printed Name

03/09/2016
Date