

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

NOV 01 2016

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: EDGEWATER P.D.
Time of Inspection: 12:20

Date of Inspection: 10/31/2016

Serial Number: 80-001243
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.049	0.079	0.195	0.080
0.000	0.049	0.080	0.197	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.050	0.080	0.197	0.081
0.000	0.050	0.079	0.197	0.081
0.000	0.050	0.080	0.197	0.081
0.000	0.049	0.080	0.198	0.081
0.000	0.050	0.080	0.197	0.081
0.000	0.049	0.080	0.198	0.081
0.000	0.050	0.080	0.197	0.081

Standard Deviations	0.0005	0.0004	0.0008	0.0003
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

MS

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

JAKE L SHANAHAN

10/31/2016
Date

Florida Department of Law Enforcement OCT 06 2016 *VSC*
Alcohol Testing Program FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: EDGEWATER P.D.
 Time of Inspection: 09:50

Date of Inspection: 10/06/2016

Serial Number: 80-001243
 Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG619605 Exp: 07/14/2018
0.000	0.051	0.080	0.198	0.076
0.000	0.050	0.080	0.199	0.077
0.000	0.051	0.080	0.198	0.076
0.000	0.051	0.080	0.198	0.077
0.000	0.051	0.080	0.198	0.077
0.000	0.051	0.081	0.198	0.076
0.000	0.051	0.081	0.198	0.077
0.000	0.051	0.081	0.198	0.077
0.000	0.051	0.081	0.199	0.077
0.000	0.051	0.080	0.199	0.077

Standard Deviations	0.0003	0.0005	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP 7512

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 Signature and Printed Name JAKE L SHANAHAN

10/06/2016
 Date

RECEIVED *WJS*

Florida Department of Law Enforcement Alcohol Testing Program *PDM*

FEB 09 2016

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: EDGEWATER P.D.
Time of Inspection: 11:37

Date of Inspection: 02/07/2016

Serial Number: 80-001243
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

BK

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG511701 Exp: 04/27/2017
0.000	0.051	0.081	0.206	0.078
0.000	0.052	0.082	0.207	0.078
0.000	0.052	0.082	0.207	0.079
0.000	0.052	0.082	0.207	0.079
0.000	0.052	0.082	0.207	0.079
0.000	0.052	0.082	0.207	0.079
0.000	0.052	0.082	0.207	0.079
0.000	0.052	0.083	0.206	0.079
0.000	0.052	0.083	0.207	0.079
0.000	0.053	0.083	0.206	0.080

Standard Deviations	0.0004	0.0006	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

02/07/2016
Date