

Florida Department of Law Enforcement

JAN 09 2017 *EC*

Alcohol Testing Program

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 09:53

Date of Inspection: 12/26/2016

Serial Number: 80-001250
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG605301 Exp: 02/22/2018
0.000	0.049	0.079	0.200	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.202	0.078
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.080	0.202	0.078
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.080	0.203	0.079
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.080	0.202	0.079

Standard Deviations	0.0003	0.0004	0.0008	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera

 Signature and Printed Name

12/26/2016
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JUN 15 2016 *JK*

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000 *JK*

Agency: BROWARD COUNTY SO
Time of Inspection: 12:32

Date of Inspection: 06/07/2016

Serial Number: 80-001250
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG600504 Exp: 01/05/2018
0.000	0.050	0.079	0.202	0.080
0.000	0.050	0.079	0.203	0.080
0.000	0.050	0.079	0.203	0.079
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.051	0.080	0.200	0.079
0.000	0.050	0.079	0.200	0.079

Standard Deviations	0.0003	0.0004	0.0010	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Rivera

DAVID E REYES-RIVERA
Signature and Printed Name

06/07/2016
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001250	Broward County Sheriffs Office	06/07/2016	<i>Will</i>

SK

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

BROWARD COUNTY SO
IntoxiLizer - Alcohol Analyzer
Model 8000 SN 80-001250
06/07/2016
Software: 8100.27

Test g/210L Time

Air Blank 0.000 10:02
Control Test 0.050 10:03
Air Blank 0.000 10:03
Control Test 0.051 10:04
Air Blank 0.000 10:04
Control Test 0.050 10:05
Air Blank 0.000 10:05
Control Test Status
Average 0.0533
Std Dev 0.0006
Rel Std Dev(%) 1.1471

Will
Operator's Signature

BROWARD COUNTY SO
IntoxiLizer - Alcohol Analyzer
Model 8000 SN 80-001250
06/07/2016
Software: 8100.27

Test g/210L Time

Air Blank 0.000 10:07
Control Test 0.080 10:08
Air Blank 0.000 10:08
Control Test 0.079 10:09
Air Blank 0.000 10:09
Control Test 0.080 10:10
Air Blank 0.000 10:10
Control Test Status
Average 0.0797
Std Dev 0.0006
Rel Std Dev(%) 0.7247

Will
Operator's Signature

BROWARD COUNTY SO
IntoxiLizer - Alcohol Analyzer
Model 8000 SN 80-001250
06/07/2016
Software: 8100.27

Test g/210L Time

Air Blank 0.000 10:13
Control Test 0.200 10:13
Air Blank 0.000 10:14
Control Test 0.199 10:15
Air Blank 0.000 10:15
Control Test 0.200 10:16
Air Blank 0.000 10:16
Control Test Status
Average 0.1997
Std Dev 0.0006
Rel Std Dev(%) 0.2892

Will
Operator's Signature

BROWARD COUNTY SO
IntoxiLizer - Alcohol Analyzer
Model 9000 SN 80-001250
06/07/2016
Software: 8100.27

Test g/210L Time

Air Blank 0.000 10:19
Control Test 0.080 10:19
Air Blank 0.000 10:20
Control Test 0.080 10:20
Air Blank 0.000 10:20
Control Test 0.080 10:21
Air Blank 0.000 10:21
Control Test Status
Average 0.0800
Std Dev 0.0008
Rel Std Dev(%) 0.0000

Will
Operator's Signature

Will

Florida Department of Law Enforcement FEB 26 2016 *BC*
 Alcohol Testing Program *Qgm* FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
 Time of Inspection: 11:32

Date of Inspection: 02/09/2016

Serial Number: 80-001250
 Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG507503 Exp: 03/16/2017
0.000	0.050	0.080	0.202	0.081
0.000	0.050	0.080	0.202	0.080
0.000	0.050	0.079	0.202	0.080
0.000	0.050	0.080	0.202	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.079	0.202	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.051	0.080	0.201	0.079
0.000	0.051	0.080	0.201	0.079
0.000	0.050	0.081	0.202	0.080

Standard Deviations	0.0004	0.0005	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Rivera DAVID E REYES-RIVERA
 Signature and Printed Name

02/09/2016
Date