

RECEIVED ✓

# Florida Department of Law Enforcement Alcohol Testing Program

JAN 27 2016

FDLE

Alcohol Testing Program  
8000

## DEPARTMENT INSPECTION REPORT - INTOXILYZER

Agency: PINELLAS COUNTY SO  
Time of Inspection: 17:02

Date of Inspection: 01/26/2016

Serial Number: 80-001274  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG511701 Exp: 04/27/2017
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.203	0.080
0.000	0.050	0.080	0.203	0.080
0.000	0.051	0.080	0.204	0.080
0.000	0.051	0.080	0.204	0.079
0.000	0.050	0.080	0.203	0.079
0.000	0.051	0.080	0.203	0.079
0.000	0.051	0.080	0.204	0.079
0.000	0.051	0.080	0.204	0.079
0.000	0.051	0.080	0.204	0.079
0.000	0.050	0.081	0.203	0.079

Standard Deviations	0.0005	0.0003	0.0009	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIELLE M BELL

Signature and Printed Name

01/26/2016  
Date

JAN 27 2016 *BC*

# Florida Department of Law Enforcement

## Alcohol Testing Program *DM*

FDLE  
Alcohol Testing Program

### DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO  
Time of Inspection: 12:03

Date of Inspection: 01/25/2016

Serial Number: 80-001274  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

*BC*

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG511701 Exp: 04/27/2017
0.000	0.052	0.082	0.204	0.081
0.000	0.053	0.082	0.206	0.080
0.000	0.053	0.083	0.206	0.080
0.000	0.053	0.082	0.206	0.080
0.000	0.053	0.082	0.206	0.079
0.000	0.054	0.082	0.206	0.080
0.000	0.054	0.083	0.206	0.080
0.000	0.054	0.083	0.206	0.080
0.000	0.054	0.083	0.206	0.080
0.000	0.054	0.083	0.206	0.080
0.000	0.054	0.083	0.206	0.079
Standard Deviations	0.0006	0.0005	0.0006	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Danielle M Bell*

DANIELLE M BELL

Signature and Printed Name

01/25/2016  
Date