

RECEIVED

✓
BCE

JUN 10 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

BK

Agency: UNION COUNTY SO
Time of Inspection: 13:01

Date of Inspection: 06/10/2016

Serial Number: 80-001288
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG600504 Exp: 01/05/2018
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.080	0.200	0.078
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.078

Standard Deviations	0.0000	0.0005	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

MB

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy


PATRICK J MURPHY

Signature and Printed Name

06/10/2016
Date

UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001288
06/10/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:05
Control Test	0.048	10:06
Air Blank	0.000	10:06
Control Test	0.048	10:07
Air Blank	0.000	10:07
Control Test	0.049	10:08
Air Blank	0.000	10:08
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	


Operator's Signature


UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001288
06/10/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:09
Control Test	0.079	10:10
Air Blank	0.000	10:11
Control Test	0.078	10:11
Air Blank	0.000	10:12
Control Test	0.078	10:13
Air Blank	0.000	10:13
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	


Operator's Signature

UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001288
06/10/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:15
Control Test	0.198	10:15
Air Blank	0.000	10:16
Control Test	0.197	10:17
Air Blank	0.000	10:17
Control Test	0.197	10:18
Air Blank	0.000	10:18
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	


Operator's Signature

UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001288
06/10/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:21
Control Test	0.078	10:21
Air Blank	0.000	10:22
Control Test	0.078	10:22
Air Blank	0.000	10:23
Control Test	0.078	10:23
Air Blank	0.000	10:23
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

065


Operator's Signature


13R