

RECEIVED

JUN 28 2016

1300

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BRADFORD COUNTY SO
Time of Inspection: 13:56

Date of Inspection: 06/24/2016

Serial Number: 80-001290
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG605301 Exp: 02/22/2018
0.000	0.049	0.080	0.197	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.081	0.200	0.078
0.000	0.051	0.081	0.200	0.078
0.000	0.051	0.080	0.199	0.078
0.000	0.050	0.081	0.201	0.079
0.000	0.051	0.080	0.201	0.078
0.000	0.051	0.081	0.200	0.078
0.000	0.050	0.081	0.200	0.078
0.000	0.051	0.081	0.200	0.078

Standard Deviations	0.0006	0.0005	0.0011	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PATRICK J MURPHY

Signature and Printed Name

06/24/2016
Date

RECEIVED ✓ B23

Florida Department of Law Enforcement Alcohol Testing Program

FEB 23 2016

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BRADFORD COUNTY SO
Time of Inspection: 15:46

Date of Inspection: 02/22/2016

Serial Number: 80-001290
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AGS11701 Exp: 04/27/2017
0.000	0.050	0.081	0.201	0.080
0.000	0.051	0.081	0.204	0.080
0.000	0.051	0.081	0.202	0.080
0.000	0.051	0.082	0.202	0.079
0.000	0.051	0.082	0.201	0.079
0.000	0.051	0.082	0.202	0.079
0.000	0.051	0.081	0.202	0.079
0.000	0.051	0.081	0.201	0.079
0.000	0.051	0.082	0.202	0.079
0.000	0.050	0.081	0.202	0.080

Standard Deviations	0.0004	0.0005	0.0008	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JENNIFER H JACOBS
Signature and Printed Name

02/22/2016
Date