

RECEIVED

Florida Department of Law Enforcement SEP 26 2016

Alcohol Testing Program

FDLE Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MEXICO BEACH PD
Time of Inspection: 14:46

Date of Inspection: 09/22/2016

Serial Number: 80-001306
Software: 8100.27

Table with 6 columns: Check or Test, YES, NO, Check or Test, YES, NO. Rows include Diagnostic Check, Minimum Sample Volume Check, Alcohol Free Subject Test, Interferent Detect Test, etc.

Table with 5 columns: Alcohol Free Test (g/210L), 0.05g/210L Test (g/210L), 0.08g/210L Test (g/210L), 0.20g/210L Test (g/210L), 0.08 g/210L Dry Gas Std Test (g/210L). Rows show test results for multiple samples.

Table with 5 columns: Standard Deviations, 0.0006, 0.0005, 0.0011, 0.0006

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

Handwritten mark resembling a circled 'e' or similar symbol.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Handwritten signature of Phillip W. Suber.

PHILLIP W SUBER

Signature and Printed Name

09/22/2016
Date