

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
SEP 23 2016
FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARTIN COUNTY SO
Time of Inspection: 12:50

Date of Inspection: 09/20/2016

Serial Number: 80-001439
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG612405 Exp: 05/03/2018
/ 0.000	0.050	0.079	0.197	0.079
/ 0.000	0.051	0.080	0.198	0.079
/ 0.000	0.051	0.081	0.199	0.080
/ 0.000	0.051	0.081	0.199	0.080
/ 0.000	0.053	0.081	0.200	0.080
/ 0.000	0.053	0.081	0.200	0.080
/ 0.000	0.053	0.081	0.200	0.080
/ 0.000	0.053	0.081	0.201	0.080
/ 0.000	0.054	0.082	0.201	0.082
/ 0.000	0.053	0.081	0.202	0.081

Standard Deviations	0.0013	0.0007	0.0014	0.0008
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0010 Number of Simulators Used: 5

Remarks:

00: Ambient Fail-MOVED AWAY FROM INSTRUMENT WITH MA

PLS

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

09/20/2016
Date