

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

SEP 07 2016

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PATRICK AFB SFS
Time of Inspection: 12:07

Date of Inspection: 09/07/2016

Serial Number: 80-001471
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG61240S Exp: 05/03/2018
/ 0.000	0.049	0.078	0.195	0.080
/ 0.000	0.051	0.080	0.196	0.080
/ 0.000	0.051	0.080	0.197	0.080
/ 0.000	0.050	0.080	0.197	0.080
/ 0.000	0.051	0.080	0.197	0.080
/ 0.000	0.051	0.080	0.196	0.080
/ 0.000	0.051	0.080	0.197	0.081
/ 0.000	0.051	0.080	0.197	0.081
/ 0.000	0.051	0.080	0.197	0.080
/ 0.000	0.051	0.081	0.197	0.080

Standard Deviations	0.0006	0.0007	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

00: Ambient Fail MA VAPORS PRESENT RETEST.

*OWS
BR*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

09/07/2016
Date