

# Florida Department of Law Enforcement Alcohol Testing Program

DEC 09 2016

FDLE  
Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

*fac*

Agency: ALACHUA PD  
Time of Inspection: 10:03

Date of Inspection: 12/09/2016

Serial Number: 80-001720  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.082 / 0.083			
0.000	0.083 / 0.083			
0.000	0.083 / 0.083			
0.000	0.083 / 0.083			
0.000	0.083 / 0.083			
0.000	0.083 / 0.083			
0.000	0.084 / 0.083			
0.000	0.084 / 0.084			
0.000	0.084 / 0.084			
0.000	0.084 / 0.084			
0.000	0.084 / 0.084			

Standard Deviations	0.0006 / 0.0004		
---------------------	-----------------	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: \_\_\_\_\_ Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance REPEAT, Control Outside Toler. WRONG SIMULATOR CONNECTED Non-compliance  
:WRONG SIMULATOR FOR 05 WILL REPEAT DI.

*Did not realize/notice wrong simulator. Will complete another Department inspection w/ correct simulator and will be in compliance once completed.*

*(Signature)*

*bk*

*(Signature)*

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*(Signature)*

JAKE L. SHANAHAN

Signature and Printed Name

12/09/2016  
Date

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

DEC 09 2016

FDLE  
Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

*BK*

Agency: ALACHUA PD  
Time of Inspection: 11:30

Date of Inspection: 12/09/2016

Serial Number: 80-001720  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG619605 Exp: 07/14/2018
0.000	0.051	0.082	0.202	0.082
0.000	0.051	0.083	0.203	0.082
0.000	0.051	0.084	0.204	0.082
0.000	0.051	0.084	0.203	0.082
0.000	0.051	0.083	0.204	0.083
0.000	0.051	0.084	0.203	0.083
0.000	0.051	0.083	0.203	0.083
0.000	0.052	0.084	0.203	0.083
0.000	0.052	0.084	0.204	0.083
0.000	0.052	0.084	0.204	0.083

Standard Deviations	0.0004	0.0007	0.0006	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

*[Handwritten initials]*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*[Handwritten Signature]*

JAKE L SHANAHAN

Signature and Printed Name

12/09/2016  
Date