

RECEIVED

Florida Department of Law Enforcement JUL 06 2016

Alcohol Testing Program

FDLE Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 13:48

Date of Inspection: 07/05/2016

Serial Number: 80-001742
Software: 8100.27

Table with 6 columns: Check or Test, YES, NO, Check or Test, YES, NO. Rows include Diagnostic Check (Pre-Inspection): OK, Minimum Sample Volume Check: OK, Alcohol Free Subject Test: 0.000, Interferent Detect Test: Interferent Detect.

Table with 5 columns: Alcohol Free Test (g/210L), 0.05g/210L Test (g/210L), 0.08g/210L Test (g/210L), 0.20g/210L Test (g/210L), 0.08 g/210L Dry Gas Std Test (g/210L). Multiple rows of test results.

Table with 5 columns: Standard Deviations, 0.0004, 0.0006, 0.0010, 0.0000

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

Handwritten initials 'WMB'

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Handwritten signature of Phillip W. Suber

PHILLIP W SUBER

Signature and Printed Name

07/05/2016
Date