

Florida Department of Law Enforcement OCT 10 2016 *zsc*
Alcohol Testing Program FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000 *BK*

Agency: PORT ST LUCIE PD
 Time of Inspection: 11:43

Date of Inspection: 10/07/2016

Serial Number: 80-001962
 Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG619605 Exp: 07/14/2018
/ 0.000	0.051	0.080	0.197	0.081
/ 0.000	0.052	0.082	0.200	0.080
/ 0.000	0.052	0.082	0.199	0.079
/ 0.000	0.052	0.082	0.199	0.080
/ 0.000	0.053	0.082	0.199	0.080
/ 0.000	0.052	0.082	0.199	0.079
/ 0.000	0.051	0.082	0.199	0.079
/ 0.000	0.053	0.083	0.200	0.079
/ 0.000	0.052	0.082	0.198	0.079
/ 0.000	0.052	0.082	0.199	0.079

Standard Deviations	0.0006	0.0007	0.0008	0.0007
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

A F / M A: DID NOT PROVIDE ADEQUATE SAMPLE. 00: Ambient Fail- ALLOWED TIME FOR MA TO CLEAR.

W/S

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

10/07/2016
Date