

RECEIVED

Florida Department of Law Enforcement OCT 18 2016

Alcohol Testing Program

FDLE Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI GARDENS PD
Time of Inspection: 10:08

Date of Inspection: 10/12/2016

Serial Number: 80-002988
Software: 8100.27

Table with 6 columns: Check or Test, YES, NO, Check or Test, YES, NO. Rows include Diagnostic Check, Minimum Sample Volume Check, Alcohol Free Subject Test, Interferent Detect Test, and Diagnostic Check (Post-Inspection).

Table with 5 columns: Alcohol Free Test (g/210L), 0.05g/210L Test, 0.08g/210L Test, 0.20g/210L Test, 0.08 g/210L Dry Gas Std Test. Contains 12 rows of test results.

Table with 5 columns: Standard Deviations, 0.0004, 0.0005, 0.0005, 0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name: DAVID E REYES-RIVERA

10/12/2016
Date