

RECEIVED *Val*

# Florida Department of Law Enforcement Alcohol Testing Program *PJM*

FEB 02 2016

FDLE

Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.  
Time of Inspection: 12:11

Date of Inspection: 02/01/2016

Serial Number: 80-004587  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG411701 Exp: 02/24/2017
0.000	0.049	0.080	0.203	0.000 / 0.078
0.000	0.050	0.080	0.202	0.000 / 0.078
0.000	0.050	0.082	0.203	0.078 / 0.078
0.000	0.050	0.081	0.203	0.077 / 0.077
0.000	0.050	0.081	0.202	0.077 / 0.077
0.000	0.050	0.081	0.204	0.077 / 0.078
0.000	0.051	0.080	0.203	0.077 / 0.076
0.000	0.050	0.080	0.203	0.076 / 0.078
0.000	0.051	0.080	0.202	0.076 / 0.077
0.000	0.051	0.081	0.204	0.077 / 0.077

Standard Deviations	0.0006	0.0006	0.0007	0.0324 / 0.0006
---------------------	--------	--------	--------	-----------------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance DGS NOT CONNECTED AT FIRST

*SMB*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

02/01/2016  
Date