

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUN 08 2016

FDLE

Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OCALA POLICE DEPT  
Time of Inspection: 12:42

Date of Inspection: 06/07/2016

Serial Number: 80-005454  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG600504 Exp: 01/05/2018
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.080	0.201	0.081
0.000	0.051	0.080	0.201	0.080
0.000	0.051	0.081	0.200	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.051	0.080	0.200	0.079
0.000	0.051	0.080	0.200	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.050	0.080	0.201	0.080
0.000	0.052	0.081	0.201	0.080

Standard Deviations	0.0006	0.0006	0.0005	0.0006
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

PWS

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PATRICK J MURPHY  
Signature and Printed Name

06/07/2016  
Date

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005454  
06/07/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:22
Control Test	0.050	10:22
Air Blank	0.000	10:23
Control Test	0.050	10:23
Air Blank	0.000	10:24
Control Test	0.049	10:25
Air Blank	0.000	10:25
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

  
-----  
Operator's Signature

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005454  
06/07/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:28
Control Test	0.079	10:29
Air Blank	0.000	10:29
Control Test	0.079	10:30
Air Blank	0.000	10:30
Control Test	0.079	10:31
Air Blank	0.000	10:32
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

  
-----  
Operator's Signature

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005454  
06/07/2016  
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	10:34
Control Test	0.199	10:35
Air Blank	0.000	10:35
Control Test	0.200	10:36
Air Blank	0.000	10:36
Control Test	0.200	10:37
Air Blank	0.000	10:38
Control Test Stats		
Average	0.1997	
Std Dev	0.0086	
Rel Std Dev(%)	0.2892	

  
-----  
Operator's Signature

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005454  
06/07/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:48
Control Test	0.080	12:49
Air Blank	0.000	12:49
Control Test	0.080	12:49
Air Blank	0.000	12:50
Control Test	0.080	12:50
Air Blank	0.000	12:51
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

D65

  
-----  
Operator's Signature

*ms*  
*BK*