

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JUN 15 2016

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: US COAST GUARD MIAMI
Time of Inspection: 12:08

Date of Inspection: 06/07/2016

Serial Number: 80-006085
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG600504 Exp: 01/05/2018
0.000	0.050	0.081	0.201	0.080
0.000	0.051	0.081	0.202	0.080
0.000	0.051	0.082	0.202	0.080
0.000	0.051	0.082	0.201	0.080
0.000	0.052	0.081	0.202	0.080
0.000	0.051	0.082	0.201	0.080
0.000	0.051	0.082	0.202	0.080
0.000	0.051	0.082	0.203	0.079
0.000	0.051	0.082	0.203	0.080
0.000	0.051	0.081	0.202	0.079

BSM

Standard Deviations	0.0004	0.0005	0.0007	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes-Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

06/07/2016
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006085	United States Coast Guard	06/07/2016	<i>ALL</i>

AK

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

US COAST GUARD M19M1
Intoxilizer - Alcohol Analyzer
Model 8000- SN 80-106085
06/07/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:48
Control Test	0.049	09:48
Air Blank	0.000	09:49
Control Test	0.050	09:49
Air Blank	0.000	09:50
Control Test	0.049	09:51
Air Blank	0.000	09:51
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1793	

ALL
Operator's Signature

US COAST GUARD M19M1
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-096085
06/07/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:53
Control Test	0.080	09:53
Air Blank	0.000	09:54
Control Test	0.079	09:55
Air Blank	0.000	09:55
Control Test	0.079	09:56
Air Blank	0.000	09:56
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

ALL
Operator's Signature

US COAST GUARD M19M1
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-106085
06/07/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:58
Control Test	0.200	09:59
Air Blank	0.000	09:59
Control Test	0.200	10:00
Air Blank	0.000	10:00
Control Test	0.201	10:01
Air Blank	0.000	10:02
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

ALL
Operator's Signature

US COAST GUARD M19M1
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-106085
06/07/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:04
Control Test	0.079	10:04
Air Blank	0.000	10:05
Control Test	0.080	10:05
Air Blank	0.000	10:05
Control Test	0.080	10:06
Air Blank	0.000	10:06
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

ALL
Operator's Signature

RAM