

RECEIVED ✓  
VBC

Florida Department of Law Enforcement JUN 08 2016

Alcohol Testing Program

FDLE  
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000 BK

Agency: OCALA POLICE DEPT  
Time of Inspection: 13:04

Date of Inspection: 06/07/2016

Serial Number: 80-006214  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG600504 Exp: 01/05/2018
0.000	0.049	0.078	0.199	0.079
0.000	0.050	0.078	0.201	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.200	0.079
0.000	0.050	0.078	0.200	0.079
0.000	0.050	0.079	0.200	0.078
0.000	0.049	0.079	0.200	0.078

Standard Deviations	0.0004	0.0005	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

PWS

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Patrick J Murphy*

PATRICK J MURPHY

Signature and Printed Name

06/07/2016  
Date

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006214  
06/07/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:35
Control Test	0.048	10:36
Air Blank	0.000	10:37
Control Test	0.048	10:37
Air Blank	0.000	10:38
Control Test	0.048	10:39
Air Blank	0.000	10:39
Control Test Stats		
Average	0.0486	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006214  
06/07/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:22
Control Test	0.077	10:23
Air Blank	0.000	10:23
Control Test	0.078	10:24
Air Blank	0.000	10:24
Control Test	0.078	10:25
Air Blank	0.000	10:26
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006214  
06/07/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:28
Control Test	0.199	10:29
Air Blank	0.000	10:29
Control Test	0.198	10:30
Air Blank	0.000	10:30
Control Test	0.197	10:31
Air Blank	0.000	10:32
Control Test Stats		
Average	0.1988	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

  
Operator's Signature

  
Operator's Signature

  
Operator's Signature

OCALA POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006214  
06/07/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:00
Control Test	0.079	13:07
Air Blank	0.000	13:07
Control Test	0.079	13:07
Air Blank	0.000	13:08
Control Test	0.079	13:08
Air Blank	0.000	13:09
Control Test Stats		
Average	0.0796	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DBS

  
Operator's Signature

3WS  
4BK