

RECEIVED
JBC

Florida Department of Law Enforcement Alcohol Testing Program

MAR 28 2016

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE CITY PD
Time of Inspection: 13:47

Date of Inspection: 03/26/2016

Serial Number: 80-006221
Software: 8100.27

BK

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG605301 Exp: 02/22/2018
0.000	0.049	0.079	0.202	0.079
0.000	0.049	0.079	0.202	0.079
0.000	0.049	0.079	0.202	0.079
0.000	0.049	0.080	0.202	0.079
0.000	0.049	0.080	0.202	0.079
0.000	0.049	0.079	0.202	0.079
0.000	0.049	0.080	0.202	0.079
0.000	0.050	0.079	0.202	0.079
0.000	0.049	0.079	0.202	0.079
0.000	0.050	0.079	0.202	0.079

Standard Deviations	0.0004	0.0004	0.0000	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

PKS

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

03/26/2016
Date

Florida Department of Law Enforcement RECEIVED

Alcohol Testing Program FEB 16 2016

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000 FDLE

Alcohol Testing Program

Agency: LAKE CITY PD Date of Inspection: 02/04/2016 Serial Number: 80-006221
 Time of Inspection: 13:49 Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG511701 Exp: 04/27/2017
0.000	0.049	0.077	0.200	0.080
0.000	0.049	0.078	0.201	0.079
0.000	0.049	0.078	0.201	0.079
0.000	0.049	0.078	0.201	0.080
0.000	0.049	0.078	0.201	0.079
0.000	0.049	0.078	0.201	0.080
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.201	0.079
Standard Deviations	0.0000	0.0003	0.0005	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JENNIFER H JACOBS
Signature and Printed Name

02/04/2016
Date