

RECEIVED

Florida Department of Law Enforcement

MAR 03 2016

Alcohol Testing Program

FDLE Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: VOLUSIA COUNTY S.O.  
Time of Inspection: 13:37

Date of Inspection: 03/03/2016

Serial Number: 80-006262  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AGS11701 Exp: 04/27/2017
/ 0.000	0.050	0.079	0.198	0.079
/ 0.000	0.051	0.080	0.201	0.079
/ 0.000	0.051	0.081	0.201	0.080
/ 0.000	0.051	0.081	0.201	0.080
/ 0.000	0.051	0.080	0.201	0.080
/ 0.000	0.051	0.080	0.201	0.080
/ 0.000	0.051	0.081	0.201	0.079
/ 0.000	0.052	0.081	0.201	0.080
/ 0.000	0.052	0.081	0.202	0.080
/ 0.000	0.052	0.081	0.202	0.080

Standard Deviations	0.0006	0.0007	0.0011	0.0004
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

00: Ambient Fail TOO MUCH MA RETEST.

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JAKE L SHANAHAN

Signature and Printed Name

03/03/2016  
Date