

RECEIVED

Florida Department of Law Enforcement JUL 06 2016

Alcohol Testing Program

FDLE Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO
Time of Inspection: 13:01

Date of Inspection: 07/05/2016

Serial Number: 80-006435
Software: 8100.27

Table with 6 columns: Check or Test, YES, NO, Check or Test, YES, NO. Rows include Diagnostic Check, Minimum Sample Volume, Alcohol Free Subject, and Interferent Detect Test.

Table with 5 columns: Alcohol Free Test (g/210L), 0.05g/210L Test, 0.08g/210L Test, 0.20g/210L Test, 0.08 g/210L Dry Gas Std Test. Multiple rows of numerical data.

Table with 5 columns: Standard Deviations, 0.0003, 0.0003, 0.0004, 0.0000

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Handwritten signature of Phillip W. Suber

PHILLIP W SUBER

Signature and Printed Name

07/05/2016
Date