

RECEIVED

SEP 28 2016

FDLE

Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 10:33

Date of Inspection: 09/19/2016

Serial Number: 80-006471
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG600504 Exp: 01/05/2018
0.000	0.049	0.079	0.201	0.078
0.000	0.050	0.080	0.202	0.079
0.000	0.049	0.080	0.202	0.078
0.000	0.049	0.080	0.202	0.079
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.080	0.203	0.079
0.000	0.049	0.080	0.203	0.079
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.079	0.202	0.078

egm

Standard Deviations	0.0005	0.0004	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

09/19/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED *Bre*

SEP 28 2016

FDLE

DEPARTMENT INSPECTION REPORT - INTOXILYZER ~~81000~~ Testing Program *BK*

Agency: MONROE COUNTY SO
Time of Inspection: 08:47

Date of Inspection: 09/19/2016

Serial Number: 80-006471
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

OGM

Standard Deviations				
---------------------	--	--	--	--


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks:

A F / M A: RAN INADVERTEDLY STOPPED BLOWING BEFORE DETECTION. 3 INST IN A ROW, RAN OUT OF BREATH
 First stage of the retest for the Alcohol Free / Mouth Alcohol test WAS NOT performed I had mouth alcohol and registered slope NOT MET when it was supposed to be alcohol free. Will restart the test, inspector error not attributed to instrument performance.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

DAVID E REYES-RIVERA

09/19/2016
Date