

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAY 26 2016 *ccc*

FDLE  
Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 11:37

Date of Inspection: 05/25/2016

Serial Number: 80-006629

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG600504 Exp: 01/05/2018
0.000	0.049	0.079	0.201	0.077 / 0.081
0.000	0.049	0.080	0.202	0.077 / 0.080
0.000	0.049	0.080	0.203	0.075 / 0.080
0.000	0.049	0.079	0.203	0.076 / 0.079
0.000	0.049	0.079	0.202	0.074 / 0.079
0.000	0.049	0.080	0.202	0.070 / 0.079
0.000	0.050	0.080	0.202	0.000 / 0.079
0.000	0.049	0.079	0.202	0.000 / 0.079
0.000	0.050	0.079	0.202	0.000 / 0.079
0.000	0.049	0.080	0.202	0.000 / 0.079

*PJM*

Standard Deviations	0.0004	0.0005	0.0005	0.0386 / 0.0006
---------------------	--------	--------	--------	-----------------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

**Remarks:**

08: Control Outside Tolerance DGS CYL EMPTY.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*David E Reyes-Rivera*

DAVID E REYES-RIVERA

Signature and Printed Name

05/25/2016  
Date