

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

OCT 03 2016 *Bus*

FDLE

Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLMES BEACH PD  
Time of Inspection: 13:11

Date of Inspection: 10/03/2016

Serial Number: 80-006692  
Software: 8100.27 *BK*

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG612405 Exp: 05/03/2018
0.000	0.049	0.079	0.195	0.077
0.000	0.050	0.080	0.196	0.077
0.000	0.049	0.079	0.196	0.077
0.000	0.049	0.079	0.195	0.077
0.000	0.049	0.080	0.196	0.077
0.000	0.049	0.079	0.196	0.077
0.000	0.049	0.079	0.195	0.077
0.000	0.049	0.080	0.196	0.077
0.000	0.050	0.080	0.196	0.077
0.000	0.050	0.079	0.195	0.077

Standard Deviations	0.0004	0.0005	0.0005	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

**Remarks:**

A F / M A: DID NOT PROVIDE ADEQUATE SAMPLE. *OW*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

10/03/2016  
Date



Florida Department of  
Law Enforcement

## REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: INTOXILYZER 8000

SERIAL NUMBER: 80-006692

OWNING AGENCY: HOLMES BEACH POLICE DEPT.

DATE OF DEPARTMENT INSPECTION: 10/3/2016

AGENCY INSPECTOR: SGT. MICHAEL PILATO

ADDRESS: 5801 MARINA DRIVE

CITY, STATE, ZIP: HOLMES BEACH, FL. 34217

TELEPHONE NUMBER: 941-708-5804 / 941-932-5557(c)

FAX NUMBER: \_\_\_\_\_

EMAIL ADDRESS (if available): pilatom@holmesbeach.org

***For Program Office Use Only:***

- Registration Issued
- Instrument Added to Evidentiary Instrument Spreadsheet
- Instrument Added to Evidentiary Instrument Database
- Instrument Added to Monthly Statistics Database
- Contact Information Added to Contact Spreadsheet