

Bell, Danielle

From: Deborah Reasoner <DReasoner@clermontfl.org>
Sent: Tuesday, June 28, 2016 1:21 PM
To: (awilliams@sao5.org); Chambers, Rebecca; Bell, Danielle; winterspringsbar@flhsmv.gov
Subject: Clermont PD Intox inspection May and June 2016
Attachments: May 2016 inspection.pdf; June 2016 inspection.pdf

Hello everyone,

I can't find where I sent you all the May inspection form, so I am attaching it with June's.

For June's inspection, I had an issue with the simulator head for the .20 solution. I switched simulator heads between the 1st and 2nd test (this is the reason for the reading on the 2nd test) but the instrument was not in compliance per FAC. I did contact the Department inspector for my agency, Danielle Bell, via telephone at approximately 1015 hours to advise her of the situation. A second inspection was then completed, utilizing only 3 simulator heads, and the instrument was in compliance.

I am attaching both of the June inspection forms for your review/files. The bad simulator head is being packaged to send out to determine if it can be repaired, until that time, it is not in service for inspections.

As always, should you have any questions, please feel free to contact me.

Enjoy your day.

Officer Debbie Reasoner
Clermont Police Department
Traffic Enforcement Unit
DRE/THI
865 W. Montrose St., Clermont, FL 34711
Tel 352-394-5588, ext. 141 | Fax 352-394-1644
Cell 352-348-0137
dreasoner@clermontfl.org
www.cityofclermontfl.gov



Our mission: To preserve and enhance the quality of life for the Clermont community by providing exceptional services.

Website: www.clermontfl.gov

PLEASE NOTE: Under Florida law, e-mail addresses are public records. If you do not want your email address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing.