



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FDLE S/N 80-000206

Date In 5/24/16 Date Out 9/26/16  Ship  P/U  H/D  CMI  EE

**Intake** Performed By TV

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By SP

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 208  
 Flow Verification (L/s)  
 Flow Column # ATP102  
 32mm 167 (.139 - .169)  
 36mm 183 (.156 - .190)  
 53mm 250 (.228 - .278)  
 103mm 515 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7-14-17
0.08	SD1011	201601F 1-26-18
0.20	SD1025	201604C 4-5-18
0.08 DGS	N/A	AG612405 5-3-18

**Flow Calibration** Performed By SP

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # 17mm  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

RECEIVED  
SEP 28 2016  
FDLE  
Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By SP

Barometric Pressure Gauge 1017  
 ID# 28427 Instrument 1015

Mouth Alcohol Solution Lot # 2015-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration  
 Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: QC - KUB

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kildland  
Quality Control Review

9/28/16  
Date

STABILITY CHECKS- INSTRUMENT # 80-000206 - FDUE - 9/26/14 SP

FDLE  
Intoxilyzer - Alconol Analyzer  
Model: 8000 SN 80-000206  
09/26/2016  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	10:12
Control Test	0.048	10:13
Air Blank	0.000	10:13
Control Test	0.047	10:14
Air Blank	0.000	10:15
Control Test	0.047	10:15
Air Blank	0.000	10:16
Control Test Stats		
Average	0.0473	
Std Dev	0.0016	
Rel Std Dev(%)	1.2196	

SP  
Operator's Signature

FDLE  
Intoxilyzer - Alconol Analyzer  
Model: 8000 SN 80-000206  
09/26/2016  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	10:03
Control Test	0.077	10:04
Air Blank	0.000	10:04
Control Test	0.078	10:05
Air Blank	0.000	10:06
Control Test	0.078	10:06
Air Blank	0.000	10:07
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

SP  
Operator's Signature

FDLE  
Intoxilyzer - Alconol Analyzer  
Model: 8000 SN 80-000206  
09/26/2016  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	10:08
Control Test	0.198	10:08
Air Blank	0.000	10:09
Control Test	0.200	10:10
Air Blank	0.000	10:10
Control Test	0.201	10:11
Air Blank	0.000	10:11
Control Test Stats		
Average	0.1997	
Std Dev	0.0015	
Rel Std Dev(%)	0.7650	

SP  
Operator's Signature

FDLE  
Intoxilyzer - Alconol Analyzer  
Model: 8000 SN 80-000206  
09/26/2016  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	10:17
Control Test	0.079	10:17
Air Blank	0.000	10:17
Control Test	0.079	10:18
Air Blank	0.000	10:18
Control Test	0.079	10:19
Air Blank	0.000	10:19
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AKB

SP  
Operator's Signature

SK