



Alcohol Testing Program

### INSTRUMENT PROCESSING SHEET

Agency FDLE

S/N 80-000227

Date In \_\_\_\_\_ Date Out \_\_\_\_\_

Ship  P/U  H/D  CMI  EE

**Intake** Performed By RUB

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
 Case  Handle  
 Dry Gas Holder  Feet  
 Keyboard/Plug  Back/Plugs  
 Screws tight  Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other \_\_\_\_\_

Notes: \_\_\_\_\_

**Quality Checks** Performed By RUB

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 243  
 Flow Verification (L/s)  
 Flow Column # ATP105  
 32mm 0.152 (.139 - .169)  
 36mm 0.167 (.156 - .190)  
 53mm 0.238 (.228 - .278)  
 103mm 0.511 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 216932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3962	2015D7A 7/14/17
0.08	SD3964	2016D1F 1/26/18
0.20	SD3933	2016D4C 4/5/18
0.08 DGS	N/A	AG1621605 9/22/18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 58mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

RECEIVED  
DEC 08 2016  
FDLE  
Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By RUB

Barometric Pressure 1016 Gauge  
 ID# 216932 1016 Instrument

Mouth Alcohol Solution Lot # 2015-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2880
Interferent	G2834
0.05	SD3962
0.08	SD3964
0.20	SD3933

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration

Optical Bench Cal  
 Post-Stability Tests  
 Other FORM 40

Notes: QC OK, SP

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland  
Quality Control Review

12/8/16  
Date

Stability Checks # 80-000227 FDLE (Evidential Loader) 12/07/16 QMS

DOS

FILE  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000227  
12/07/2016  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	12:24
Control Test	0.050	12:25
Air Blank	0.000	12:25
Control Test	0.051	12:26
Air Blank	0.000	12:27
Control Test	0.050	12:27
Air Blank	0.000	12:28
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

FILE  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000227  
12/07/2016  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	11:46
Control Test	0.079	11:47
Air Blank	0.000	11:48
Control Test	0.079	11:48
Air Blank	0.000	11:49
Control Test	0.080	11:50
Air Blank	0.000	11:50
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

FILE  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000227  
12/07/2016  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	11:52
Control Test	0.199	11:52
Air Blank	0.000	11:53
Control Test	0.199	11:54
Air Blank	0.000	11:54
Control Test	0.198	11:55
Air Blank	0.000	11:55
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

FILE  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000227  
12/07/2016  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	11:31
Control Test	0.080	11:31
Air Blank	0.000	11:32
Control Test	0.080	11:32
Air Blank	0.000	11:32
Control Test	0.080	11:33
Air Blank	0.000	11:33
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

QMS  
Operator's Signature

QMS  
Operator's Signature

QMS  
Operator's Signature

QMS  
Operator's Signature

SP

BK