

INSTRUMENT PROCESSING SHEET

Agency FDLE S/N 90-000228
 Date In 12/16/15 Date Out 1/6/2016 Ship P/U H/D EE

RECEIVED
 JAN 07 2016
 FDLE
 Alcohol Testing Program

Intake Performed By TR

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By 99

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 206
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm .148 (.139 - .169)
 36mm .164 (.156 - .190)
 53mm .234 (.228 - .278)
 103mm .500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28662

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 07/14/2017
0.08	SD1011	2015026 02/24/2017
0.20	SD1025	201505A 05/12/2017
0.08 DGS	N/A	A6511701 04/27/2017

Flow Calibration Performed By 99

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By 99

Barometric Pressure 1022 Gauge
 ID# 28662 1021 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Optical Bench Cal
<input checked="" type="checkbox"/> Pre-Stability Tests	<input type="checkbox"/> Post-Stability Tests
<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Other _____

Notes: 2016 Annual
QC - BK

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy
 Quality Control Review

1/7/16
 Date

Stability Checks - Instrument # 80-000228 FDLE 1/6/2016

ASM

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000
01/06/2016
Software: 8100.27

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000
01/06/2016
Software: 8100.27

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000
01/06/2016
Software: 8100.27

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000
01/06/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	15:02
Control Test	0.052	15:03
Air Blank	0.000	15:04
Control Test	0.051	15:04
Air Blank	0.000	15:05
Control Test	0.051	15:06
Air Blank	0.000	15:06
Control Test Stats		
Average	0.0513	
Std Dev	0.0006	
Rel Std Dev(%)	1.1247	

Test	9/21/0L	Time
Air Blank	0.000	15:07
Control Test	0.080	15:08
Air Blank	0.000	15:09
Control Test	0.079	15:09
Air Blank	0.000	15:10
Control Test	0.080	15:11
Air Blank	0.000	15:11
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Test	9/21/0L	Time
Air Blank	0.000	15:12
Control Test	0.202	15:13
Air Blank	0.000	15:13
Control Test	0.202	15:14
Air Blank	0.000	15:15
Control Test	0.202	15:16
Air Blank	0.000	15:16
Control Test Stats		
Average	0.2020	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	9/21/0L	Time
Air Blank	0.000	14:59
Control Test	0.081	15:00
Air Blank	0.000	15:00
Control Test	0.082	15:01
Air Blank	0.000	15:01
Control Test	0.082	15:01
Air Blank	0.000	15:02
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

AK

D6S

AS
Operator's Signature

AS
Operator's Signature

AS
Operator's Signature

AS
Operator's Signature