



INSTRUMENT PROCESSING SHEET

Agency Marion County SD S/N 80-000752
Date In 6/30/16 Date Out 7/5/2016

Ship P/U H/D CMI EEE

RECEIVED
JUL 06 2016
FDLE
Alcohol Testing Program

Intake Performed By DD
 Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case ok Handle ok
 Dry Gas Holder ok Feet ok
 Keyboard/Plug ok Back/Plugs ok
 Screws tight ok Breath Hose ok

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By AVG
 Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 1.2
 Flow Verification (L/s)
Flow Column # ATP102
32mm 160 (.139 - .169)
36mm 171 (.156 - .190)
53mm 242 (.228 - .278)
103mm 511 (.447 - .547)
 Barometric Pressure Check
Gauge ID # 28427
 Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	503962	201507A 7/14/17
0.08	503964	201601F 1/22/18
0.20	64444	201505A 5/12/17
0.08 DGS	N/A	AG105301 2/22/18

Flow Calibration Performed By _____
 Flow Calibration N/A
 Flow Calibration Complete
Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
Flow Column # _____
32mm _____ (.139 - .169)
36mm _____ (.156 - .190)
53mm _____ (.228 - .278)
103mm _____ (.447 - .547)

Maintenance Performed By _____
 Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____
 Optical Bench Calibration N/A
 Optical Bench Calibration Complete
Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Notes: CC-PWS

Department Inspection Performed By AVG
 Barometric Pressure 1017 Gauge
ID# 28427 1013 Instrument

Mouth Alcohol Solution Lot # 2016-A
Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	62880
Interferent	62824
0.05	503962
0.08	503964
0.20	64444

Attachments
 Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place Into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

L. Brett Kurbelant
Quality Control Review

7/5/16
Date

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000752
07/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:13
Control Test	0.049	11:13
Air Blank	0.000	11:14
Control Test	0.049	11:14
Air Blank	0.000	11:15
Control Test	0.049	11:16
Air Blank	0.000	11:16
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000752
07/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:08
Control Test	0.078	11:08
Air Blank	0.000	11:09
Control Test	0.078	11:09
Air Blank	0.000	11:10
Control Test	0.079	11:11
Air Blank	0.000	11:11
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000752
07/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:18
Control Test	0.198	11:19
Air Blank	0.000	11:19
Control Test	0.199	11:20
Air Blank	0.000	11:20
Control Test	0.198	11:21
Air Blank	0.000	11:21
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

P Murphy

Operator's Signature

P Murphy

Operator's Signature

P Murphy

Operator's Signature

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000752
07/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:27
Control Test	0.081	11:27
Air Blank	0.000	11:27
Control Test	0.082	11:28
Air Blank	0.000	11:28
Control Test	0.081	11:29
Air Blank	0.000	11:29
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

DGS

P Murphy

Operator's Signature

*pwr
1314*