

## INSTRUMENT PROCESSING SHEET

Agency Okaloosa CSO S/N 80-000763  
 Date In 8/5/16 Date Out 8/8/16  Ship  P/U  H/D  CMI  EE

**RECEIVED**  
**AUG 10 2016**  
**FDLE**  
**Alcohol Testing Program**

**Intake** Performed By [Signature]

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
ok Case ok Handle  
ok Dry Gas Holder ok Feet  
ok Keyboard/Plug ok Back/Plugs  
ok Screws tight ok Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other \_\_\_\_\_

Notes: \_\_\_\_\_

**Quality Checks** Performed By [Signature]

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 220  
 Flow Verification (L/s)  
 Flow Column # API105  
 32mm 148 (.139 - .169)  
 36mm 167 (.156 - .190)  
 53mm 242 (.228 - .278)  
 103mm 500 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G1739	201507A 7/14/17
0.08	G8149	201601F 1/26/18
0.20	G11621	201604C 4/5/18
0.08 DGS	N/A	AG612405 5/3/18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 53mm  
 15L/min - 53mm  
 30L/min - 103mm  
 R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure 1011 Gauge  
 ID# 28427 1008 Instrument

Mouth Alcohol Solution Lot # 2015-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2879
Interferent	G8144
0.05	G1739
0.08	G8149
0.20	G11621

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration  
 Optical Bench Cal  
 Post-Stability Tests  
 Other Form 40

Notes: PA/DC OK [Signature]

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Brett Hickland  
 Quality Control Review

8/9/16  
 Date

Stability Tests

- Okaloosa CSO

#80-000763

8/8/16

BK

OKALOOSA COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000763  
 08/08/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:49
Control Test	0.050	12:49
Air Blank	0.000	12:50
Control Test	0.051	12:51
Air Blank	0.000	12:51
Control Test	0.050	12:52
Air Blank	0.000	12:52
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

*Handwritten signature*

*Handwritten signature*

Operator's Signature

OKALOOSA COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000763  
 08/08/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:54
Control Test	0.080	12:55
Air Blank	0.000	12:55
Control Test	0.080	12:56
Air Blank	0.000	12:57
Control Test	0.080	12:57
Air Blank	0.000	12:58
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*Handwritten signature*

Operator's Signature

OKALOOSA COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000763  
 08/08/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:44
Control Test	0.200	12:45
Air Blank	0.000	12:45
Control Test	0.200	12:46
Air Blank	0.000	12:47
Control Test	0.202	12:47
Air Blank	0.000	12:48
Control Test Stats		
Average	0.2007	
Std Dev	0.0012	
Rel Std Dev(%)	0.5754	

*Handwritten signature*

Operator's Signature

OKALOOSA COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000763  
 08/08/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:59
Control Test	0.078	12:59
Air Blank	0.000	13:00
Control Test	0.079	13:00
Air Blank	0.000	13:01
Control Test	0.079	13:01
Air Blank	0.000	13:01
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

*Handwritten signature*

*Handwritten signature*

Operator's Signature