

INSTRUMENT PROCESSING SHEET

Agency Lafayette County S/N 80-000773
 Date In 11/10/16 Date Out 11/15/16 Ship P/U H/D CMI EE

Intake Performed By <u>[Signature]</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement <input type="checkbox"/> Electronics <input type="checkbox"/> Other _____ Visual Inspection: <input checked="" type="checkbox"/> Case <u>ok</u> Handle <u>ok</u> <input checked="" type="checkbox"/> Dry Gas Holder <u>ok</u> Feet <u>ok</u> <input checked="" type="checkbox"/> Keyboard/Plug <u>ok</u> Back/Plugs <u>ok</u> <input checked="" type="checkbox"/> Screws tight <u>ok</u> Breath Hose <u>ok</u> Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>206</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>A7P102</u> 32mm <u>151</u> (.139 - .169) 36mm <u>177</u> (.156 - .190) 53mm <u>246</u> (.228 - .278) 103mm <u>508</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG626605 9/22/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AG626605 9/22/18	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification Program Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD1011	201601F 1/26/18															
0.20	SD1025	201604C 4/5/18															
0.08 DGS	N/A	AG626605 9/22/18															
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____															

RECEIVED
 NOV 16 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1017</u> Gauge ID# <u>28427</u> <u>1014</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	SD1019																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	SD1025																																																												
Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____																																																													

Notes: QC & DGS

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Whiteland
 Quality Control Review

11/16/16
 Date

80-000773
 Stabilizing Curve
 11/15/16 RD

ANALYZER 8000
 Instrument Initialization
 07:56 11/15/2016

LAFAYETTE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000773
 11/15/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:21
Control Test	0.078	08:21
Air Blank	0.000	08:22
Control Test	0.078	08:22
Air Blank	0.000	08:22
Control Test	0.078	08:22
Air Blank	0.000	08:23
Control Test	0.078	08:23
Air Blank	0.000	08:23
Control Test	0.000	08:23
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

RDG

Operator's Signature

LAFAYETTE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000773
 11/15/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:32
Control Test	0.048	08:33
Air Blank	0.000	08:33
Control Test	0.049	08:34
Air Blank	0.000	08:35
Control Test	0.049	08:35
Air Blank	0.000	08:35
Control Test	0.000	08:36
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

Operator's Signature

LAFAYETTE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000773
 11/15/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:37
Control Test	0.199	08:38
Air Blank	0.000	08:38
Control Test	0.199	08:39
Air Blank	0.000	08:39
Control Test	0.198	08:40
Air Blank	0.000	08:41
Control Test	0.000	08:41
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

Operator's Signature

LAFAYETTE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000773
 11/15/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:42
Control Test	0.078	08:42
Air Blank	0.000	08:42
Control Test	0.079	08:43
Air Blank	0.000	08:43
Control Test	0.079	08:43
Air Blank	0.000	08:43
Control Test	0.000	08:44
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

RDG BK

Operator's Signature