



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency NAS - Jacksonville

S/N 80-000778

Date In 12/5/16

Date Out 12/9/16

Ship P/U H/D CMI EE

Intake Performed By <u>[Signature]</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>262</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATD/02</u> 32mm <u>161</u> (.139 - .169) 36mm <u>183</u> (.156 - .190) 53mm <u>264</u> (.228 - .278) 103mm <u>519</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>DR2035</td> <td>201603D 3/8/18</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/20/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG626605 9/22/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	DR2035	201603D 3/8/18	0.08	SD1011	201601F 1/20/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AG626605 9/22/18	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # <u>DEC 09 2016</u> <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	DR2035	201603D 3/8/18															
0.08	SD1011	201601F 1/20/18															
0.20	SD1025	201604C 4/5/18															
0.08 DGS	N/A	AG626605 9/22/18															
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____															

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																																
0.000		N/A	N/A																																																
0.040																																																			
0.100																																																			
0.200																																																			
0.400																																																			
0.080 DGS	N/A																																																		
Simulator	Serial Number	Lot Number	Expiration																																																
0.05																																																			
0.08																																																			
0.20																																																			
0.08 DGS	N/A																																																		

Department Inspection Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1026</u> Gauge ID# <u>28427</u> <u>1027</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>DR2035</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table>		Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	DR2035	0.08	SD1011	0.20	SD1025
Simulator	Serial Number												
0.00	SD1019												
Interferent	SD1021												
0.05	DR2035												
0.08	SD1011												
0.20	SD1025												
Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Other _____													

Notes: During the 105 run on Dept Insp - I inadvertently connected the wrong simulator corrected by connecting the correct one and repeating - In Compliance

[Signature]

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

80-000778
 Stability Checks
 12/9/16

INTOXILYZER 8000
 Instrument Initialization
 07:59 12/09/2016


MHS JACKSONVILLE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000778
 12/09/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:34
Control Test	0.078	08:35
Air Blank	0.000	08:35
Control Test	0.078	08:36
Air Blank	0.000	08:37
Control Test	0.079	08:37
Air Blank	0.000	08:38
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	


 Operator's Signature

MHS JACKSONVILLE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000778
 12/09/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:40
Control Test	0.049	08:40
Air Blank	0.000	08:41
Control Test	0.049	08:41
Air Blank	0.000	08:42
Control Test	0.048	08:43
Air Blank	0.000	08:43
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	


 Operator's Signature

MHS JACKSONVILLE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000778
 12/09/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:45
Control Test	0.201	08:46
Air Blank	0.000	08:46
Control Test	0.199	08:47
Air Blank	0.000	08:47
Control Test	0.199	08:48
Air Blank	0.000	08:49
Control Test Stats		
Average	0.1997	
Std Dev	0.0012	
Rel Std Dev(%)	0.5783	


 Operator's Signature

RMS


 Operator's Signature

RSK