



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Sumner County S.O. S/N 80-000780

Date In 12/27/16 Date Out 12/28/16 Ship P/U H/D CMI EE

Intake Performed By RWB

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By RWB

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 231
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 161 (.139 - .169)
 36mm 179 (.156 - .190)
 53mm 264 (.228 - .278)
 103mm 508 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28602

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>SD1011</u>	<u>201601P</u> <u>1/26/18</u>
0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	N/A	<u>AG619605</u> <u>7/14/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____

5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____

Post Calibration _____

Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
JAN 03 2017

FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By RWB

Barometric Pressure 1020 Gauge
 ID# 28602 1020 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>DR2035</u>
0.08	<u>SD1011</u>
0.20	<u>SD1025</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC'd RWB

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
Quality Control Review

1/3/17
Date

80-000780
 Stable. Very Checks
 12/28/16

INTOXILYZER 8000
 Instrument Initialization
 08:10 12/28/2016

SUWANNEE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000780
 12/28/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:35
Control Test	0.082	08:36
Air Blank	0.000	08:36
Control Test	0.082	08:36
Air Blank	0.000	08:37
Control Test	0.082	08:37
Air Blank	0.000	08:38
Control Test Stats		
Average	0.0820	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Operator's Signature

SUWANNEE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000780
 12/28/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:46
Control Test	0.195	08:47
Air Blank	0.000	08:48
Control Test	0.196	08:48
Air Blank	0.000	08:49
Control Test	0.196	08:50
Air Blank	0.000	08:50
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	

Operator's Signature

SUWANNEE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000780
 12/28/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:43
Control Test	0.082	08:43
Air Blank	0.000	08:44
Control Test	0.082	08:44
Air Blank	0.000	08:45
Control Test	0.082	08:45
Air Blank	0.000	08:45
Control Test Stats		
Average	0.0820	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

SUWANNEE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000780
 12/28/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:51
Control Test	0.048	08:52
Air Blank	0.000	08:52
Control Test	0.048	08:53
Air Blank	0.000	08:53
Control Test	0.048	08:54
Air Blank	0.000	08:55
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Operator's Signature

BK