



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Golf Breeze PD S/N 70-000787

Date In 9/14/16 Date Out 9/15/16 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By PWD

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 132
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 152 (.139 - .169)
 36mm 167 (.156 - .190)
 53mm 238 (.228 - .278)
 103mm 515 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201109F 1/26/18
0.20	SD1025	201604C 4/5/18
0.08 DGS	N/A	AG 612405 5/3/18

Flow Calibration Performed By [Signature]

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
SEP 16 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By PWD

Barometric Pressure 1018 Gauge
 ID# 28427 1014 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QA/QC OK. WSPM 9/16/16

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett H. [Signature]
Quality Control Review

9/16/16
Date

Stability Tests - Gulf Breeze PD #80-000787 9/16/16

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000787
09/15/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:39
Control Test	0.048	10:39
Air Blank	0.000	10:40
Control Test	0.048	10:41
Air Blank	0.000	10:41
Control Test	0.048	10:42
Air Blank	0.000	10:42
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

WJM

PWS

Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000787
09/15/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:31
Control Test	0.078	10:32
Air Blank	0.000	10:33
Control Test	0.079	10:33
Air Blank	0.000	10:34
Control Test	0.079	10:34
Air Blank	0.000	10:35
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

PWS

Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000787
09/15/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:50
Control Test	0.196	10:51
Air Blank	0.000	10:51
Control Test	0.196	10:52
Air Blank	0.000	10:53
Control Test	0.196	10:53
Air Blank	0.000	10:54
Control Test Stats		
Average	0.1960	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

PWS

Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000787
09/15/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:43
Control Test	0.081	10:44
Air Blank	0.000	10:44
Control Test	0.081	10:45
Air Blank	0.000	10:45
Control Test	0.082	10:45
Air Blank	0.000	10:45
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

PWS

RSK

PWS

Operator's Signature



INSTRUMENT PROCESSING SHEET

Agency Gulf Breeze PD S/N 80-000787

Date In 6/6/16 Date Out 6/9/16 Ship P/U A/D CMI EE

Intake Performed By DP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By PHS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 131
 Flow Verification (L/s)
 Flow Column # HT1102
 32mm 144 (.139 - .169)
 36mm 164 (.156 - .190)
 53mm 238 (.228 - .278)
 103mm 511 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>G2403</u>	<u>201507A</u> <u>7/14/17</u>
0.08	<u>503964</u>	<u>201601F</u> <u>1/26/18</u>
0.20	<u>G4444</u>	<u>201505A</u> <u>5/12/17</u>
0.08 DGS	<u>N/A</u>	<u>AG600504</u> <u>1/5/18</u>

Flow Calibration Performed By PHS

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED JUN 10 2016 FDLE Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		<u>N/A</u>	<u>N/A</u>
0.040			
0.100			
0.200			
0.400			
0.080 DGS	<u>N/A</u>		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	<u>N/A</u>		

Department Inspection Performed By PHS

Barometric Pressure 1013 Gauge
 ID# 28427 1018 Instrument

Mouth Alcohol Solution Lot # 2015-1A
 Acetone Stock Solution Lot # 2016-13

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2834</u>
0.05	<u>G2403</u>
0.08	<u>503964</u>
0.20	<u>G4444</u>

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: DR/ACC OK GYM

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
Quality Control Review

6/10/16
Date