

INSTRUMENT PROCESSING SHEET

Agency Gulf County SO S/N 80-000795
 Date In 9/21/16 Date Out 9/30/16 Ship P/U H/D CMI EE

Intake Performed By <u>DP</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	Quality Checks Performed By <u>PWS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>237</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>152</u> (.139 - .169) 36mm <u>167</u> (.156 - .190) 53mm <u>234</u> (.228 - .278) 103mm <u>500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>SD1018</u></td> <td><u>201507A</u> <u>7/14/17</u></td> </tr> <tr> <td>0.08</td> <td><u>SD1011</u></td> <td><u>201601F</u> <u>1/26/18</u></td> </tr> <tr> <td>0.20</td> <td><u>SD1025</u></td> <td><u>201604C</u> <u>4/5/18</u></td> </tr> <tr> <td>0.08 DGS</td> <td><u>N/A</u></td> <td><u>AG612405</u> <u>5/3/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>SD1018</u>	<u>201507A</u> <u>7/14/17</u>	0.08	<u>SD1011</u>	<u>201601F</u> <u>1/26/18</u>	0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>	0.08 DGS	<u>N/A</u>	<u>AG612405</u> <u>5/3/18</u>	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # <u>DGT 03 2016</u> <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	<u>SD1018</u>	<u>201507A</u> <u>7/14/17</u>															
0.08	<u>SD1011</u>	<u>201601F</u> <u>1/26/18</u>															
0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>															
0.08 DGS	<u>N/A</u>	<u>AG612405</u> <u>5/3/18</u>															
Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____																	

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td><u>N/A</u></td> <td><u>N/A</u></td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td><u>N/A</u></td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		<u>N/A</u>	<u>N/A</u>	0.040				0.100				0.200				0.400				0.080 DGS	<u>N/A</u>			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	<u>N/A</u>			Department Inspection Performed By <u>PWS</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1017</u> Gauge ID# <u>28427</u> <u>1017</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td><u>SD1019</u></td> </tr> <tr> <td>Interferent</td> <td><u>SD1021</u></td> </tr> <tr> <td>0.05</td> <td><u>SD1018</u></td> </tr> <tr> <td>0.08</td> <td><u>SD1011</u></td> </tr> <tr> <td>0.20</td> <td><u>SD1025</u></td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	Simulator	Serial Number	0.00	<u>SD1019</u>	Interferent	<u>SD1021</u>	0.05	<u>SD1018</u>	0.08	<u>SD1011</u>	0.20	<u>SD1025</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		<u>N/A</u>	<u>N/A</u>																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	<u>N/A</u>																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	<u>N/A</u>																																																												
Simulator	Serial Number																																																												
0.00	<u>SD1019</u>																																																												
Interferent	<u>SD1021</u>																																																												
0.05	<u>SD1018</u>																																																												
0.08	<u>SD1011</u>																																																												
0.20	<u>SD1025</u>																																																												

Notes: QC: SF

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Wickland
 Quality Control Review

10/3/16
 Date

I HEREBY CERTIFY THE AUTHENTICITY OF THE SEAL APPLIED HERETO.

Stability

Tests - Gulf CSO

80-000795

9/30/16

GULF COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000795
09/30/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:40
Control Test	0.049	08:41
Air Blank	0.000	08:41
Control Test	0.049	08:42
Air Blank	0.000	08:43
Control Test	0.049	08:43
Air Blank	0.000	08:44
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

PWS

Operator's Signature

GULF COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000795
09/30/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:45
Control Test	0.079	08:45
Air Blank	0.000	08:46
Control Test	0.079	08:47
Air Blank	0.000	08:47
Control Test	0.079	08:48
Air Blank	0.000	08:48
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

PWS

Operator's Signature

GULF COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000795
09/30/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:49
Control Test	0.199	08:50
Air Blank	0.000	08:50
Control Test	0.197	08:51
Air Blank	0.000	08:52
Control Test	0.198	08:52
Air Blank	0.000	08:53
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

PWS

Operator's Signature

GULF COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000795
09/30/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:54
Control Test	0.077	08:54
Air Blank	0.000	08:55
Control Test	0.078	08:55
Air Blank	0.000	08:55
Control Test	0.078	08:56
Air Blank	0.000	08:56
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

PWS

PWS

Operator's Signature

SP 13K

I HEREBY CERTIFY THE AUTHENTICITY OF THE SIGNATURES HEREIN.



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Gulf County SO S/N 80-000795

Date In 9/21/16 Date Out 9/30/16 Ship P/U H/D CMI EE

Intake <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <input checked="" type="checkbox"/> Case <u>OK</u> <input checked="" type="checkbox"/> Handle <u>OK</u> <input checked="" type="checkbox"/> Dry Gas Holder <u>OK</u> <input checked="" type="checkbox"/> Feet <u>OK</u> <input checked="" type="checkbox"/> Keyboard/Plug <u>OK</u> <input checked="" type="checkbox"/> Back/Plugs <u>OK</u> <input checked="" type="checkbox"/> Screws tight <u>OK</u> <input checked="" type="checkbox"/> Breath Hose <u>OK</u> Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	Quality Checks Performed By <u>JWS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>237</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>152</u> (.139 - .169) 36mm <u>167</u> (.156 - .190) 53mm <u>234</u> (.228 - .278) 103mm <u>500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>SD1018</u></td> <td><u>201507A</u> <u>7/14/17</u></td> </tr> <tr> <td>0.08</td> <td><u>SD1011</u></td> <td><u>201601F</u> <u>1/26/18</u></td> </tr> <tr> <td>0.20</td> <td><u>SD1025</u></td> <td><u>201604C</u> <u>4/5/18</u></td> </tr> <tr> <td>0.08 DGS</td> <td><u>N/A</u></td> <td><u>AG62405</u> <u>5/3/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>SD1018</u>	<u>201507A</u> <u>7/14/17</u>	0.08	<u>SD1011</u>	<u>201601F</u> <u>1/26/18</u>	0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>	0.08 DGS	<u>N/A</u>	<u>AG62405</u> <u>5/3/18</u>	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration <u>RECEIVED</u> <input type="checkbox"/> Flow Calibration Complete Flow Column # <u>OCT 03 2016</u> <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____
Simulator	Serial #	Lot #/Exp															
0.05	<u>SD1018</u>	<u>201507A</u> <u>7/14/17</u>															
0.08	<u>SD1011</u>	<u>201601F</u> <u>1/26/18</u>															
0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>															
0.08 DGS	<u>N/A</u>	<u>AG62405</u> <u>5/3/18</u>															

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td><u>N/A</u></td> <td><u>N/A</u></td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td><u>N/A</u></td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		<u>N/A</u>	<u>N/A</u>	0.040				0.100				0.200				0.400				0.080 DGS	<u>N/A</u>			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	<u>N/A</u>			Department Inspection Performed By <u>JWS</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1017</u> Gauge ID# <u>28427</u> <u>1017</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td><u>SD1019</u></td> </tr> <tr> <td>Interferent</td> <td><u>SD1021</u></td> </tr> <tr> <td>0.05</td> <td><u>SD1018</u></td> </tr> <tr> <td>0.08</td> <td><u>SD1011</u></td> </tr> <tr> <td>0.20</td> <td><u>SD1025</u></td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	Simulator	Serial Number	0.00	<u>SD1019</u>	Interferent	<u>SD1021</u>	0.05	<u>SD1018</u>	0.08	<u>SD1011</u>	0.20	<u>SD1025</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		<u>N/A</u>	<u>N/A</u>																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	<u>N/A</u>																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	<u>N/A</u>																																																												
Simulator	Serial Number																																																												
0.00	<u>SD1019</u>																																																												
Interferent	<u>SD1021</u>																																																												
0.05	<u>SD1018</u>																																																												
0.08	<u>SD1011</u>																																																												
0.20	<u>SD1025</u>																																																												

Notes: QC: SF

 Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kuckland

Quality Control Review

10/3/16
Date

Stability

Tests - Gulf CSO

80-000795

9/30/16

GULF COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000795
 09/30/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	08:40
Control Test	0.049	08:41
Air Blank	0.000	08:41
Control Test	0.049	08:42
Air Blank	0.000	08:43
Control Test	0.049	08:43
Air Blank	0.000	08:44
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

PWS

Operator's Signature

GULF COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000795
 09/30/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	08:45
Control Test	0.079	08:45
Air Blank	0.000	08:46
Control Test	0.079	08:47
Air Blank	0.000	08:47
Control Test	0.079	08:48
Air Blank	0.000	08:48
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

PWS

Operator's Signature

GULF COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000795
 09/30/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	08:49
Control Test	0.199	08:50
Air Blank	0.000	08:50
Control Test	0.197	08:51
Air Blank	0.000	08:52
Control Test	0.198	08:52
Air Blank	0.000	08:53
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

PWS

Operator's Signature

GULF COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000795
 09/30/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	08:54
Control Test	0.077	08:54
Air Blank	0.000	08:55
Control Test	0.078	08:55
Air Blank	0.000	08:55
Control Test	0.078	08:56
Air Blank	0.000	08:56
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

PWS

PWS

Operator's Signature

SP 13K