

INSTRUMENT PROCESSING SHEET

Agency Hillsborough County SO

S/N 80-000808

Date In 3/8/16

Date Out 3/10/16

Ship P/U H/D CMI EE

Intake Performed By <u>FB</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>RMB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>177</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32mm <u>0.152</u> (.139 - .169) 36mm <u>0.167</u> (.156 - .190) 53mm <u>0.234</u> (.228 - .278) 103mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>G4444</td> <td>201505A 5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG1605301 2/22/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	G4444	201505A 5/12/17	0.08 DGS	N/A	AG1605301 2/22/18	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 103mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD1011	201601F 1/26/18															
0.20	G4444	201505A 5/12/17															
0.08 DGS	N/A	AG1605301 2/22/18															
Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																	
Suggested Service _____ _____																	

RECEIVED
 MAR 21 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By <u>RMB</u> <input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1013</u> ID# <u>26932</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>DR1275</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD3962</td> <td>15108</td> <td>8/18/17</td> </tr> <tr> <td>0.100</td> <td>G2078</td> <td>15001</td> <td>5/20/17</td> </tr> <tr> <td>0.200</td> <td>G2408</td> <td>15104</td> <td>5/27/17</td> </tr> <tr> <td>0.400</td> <td>SD3933</td> <td>15105</td> <td>6/10/17</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>09014080A1</td> <td>5/1/16</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A</td> <td>7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F</td> <td>01/26/18</td> </tr> <tr> <td>0.20</td> <td>G4444</td> <td>201505A</td> <td>5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG1605301</td> <td>2/22/18</td> </tr> </tbody> </table> Notes: <u>Optical bench calibration completed to bring values closer to nominal. RMB</u> <u>GC-TBK</u> _____ _____	Simulator	Serial Number	Lot Number	Expiration	0.000	DR1275	N/A	N/A	0.040	SD3962	15108	8/18/17	0.100	G2078	15001	5/20/17	0.200	G2408	15104	5/27/17	0.400	SD3933	15105	6/10/17	0.080 DGS	N/A	09014080A1	5/1/16	Simulator	Serial Number	Lot Number	Expiration	0.05	SD1018	201507A	7/14/17	0.08	SD1011	201601F	01/26/18	0.20	G4444	201505A	5/12/17	0.08 DGS	N/A	AG1605301	2/22/18	Department Inspection Performed By <u>RMB</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1013</u> Gauge ID# <u>28427</u> <u>1013</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SN1022</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>G4444</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	Simulator	Serial Number	0.00	SN1022	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	G4444
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	DR1275	N/A	N/A																																																										
0.040	SD3962	15108	8/18/17																																																										
0.100	G2078	15001	5/20/17																																																										
0.200	G2408	15104	5/27/17																																																										
0.400	SD3933	15105	6/10/17																																																										
0.080 DGS	N/A	09014080A1	5/1/16																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05	SD1018	201507A	7/14/17																																																										
0.08	SD1011	201601F	01/26/18																																																										
0.20	G4444	201505A	5/12/17																																																										
0.08 DGS	N/A	AG1605301	2/22/18																																																										
Simulator	Serial Number																																																												
0.00	SN1022																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	G4444																																																												
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																													

Patrick Murphy
Quality Control Review

3/21/16
Date

Pre-

Stability Checks 80-000808 Hillsborough County S.O. 3/9/16 *DMB*

DMB

AK

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000808
03/09/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:16
Control Test	0.052	15:17
Air Blank	0.000	15:17
Control Test	0.050	15:18
Air Blank	0.000	15:19
Control Test	0.050	15:19
Air Blank	0.000	15:20
Control Test Stats		
Average	0.0507	
Std Dev	0.0012	
Rel Std Dev(%)	2.2790	

DMB

Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000808
03/09/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:25
Control Test	0.080	15:26
Air Blank	0.000	15:27
Control Test	0.080	15:27
Air Blank	0.000	15:28
Control Test	0.081	15:29
Air Blank	0.000	15:29
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DMB

Operator's Signature

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000808
03/09/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:59
Control Test	0.207	15:00
Air Blank	0.000	15:01
Control Test	0.201	15:02
Air Blank	0.000	15:02
Control Test	0.201	15:03
Air Blank	0.000	15:04
Control Test Stats		
Average	0.2030	
Std Dev	0.0035	
Rel Std Dev(%)	1.7065	

DMB

Operator's Signature

*Kink in tubing. Corrected after 1st Sample. *DMB**

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000808
03/09/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:40
Control Test	0.202	15:40
Air Blank	0.000	15:41
Control Test	0.203	15:42
Air Blank	0.000	15:42
Control Test	0.203	15:43
Air Blank	0.000	15:43
Control Test Stats		
Average	0.2027	
Std Dev	0.0006	
Rel Std Dev(%)	0.2849	

DMB

Operator's Signature

file-

Stability Checks 80-000808 Hillsborough County S.O. 3/9/16 RMB

PJAM

AK

DES

DES

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000808
03/09/2016
Software: 8100.27

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000808
03/09/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:05
Control Test	0.083	15:05
Air Blank	0.000	15:06
Control Test	0.083	15:06
Air Blank	0.000	15:06
Control Test	0.083	15:07
Air Blank	0.000	15:07
Control Test Stats		
Average	0.0830	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	15:22
Control Test	0.082	15:22
Air Blank	0.000	15:23
Control Test	0.082	15:23
Air Blank	0.000	15:23
Control Test	0.083	15:24
Air Blank	0.000	15:24
Control Test Stats		
Average	0.0823	
Std Dev	0.0006	
Rel Std Dev(%)	0.7012	

RMB

Operator's Signature

RMB

Operator's Signature

Calibration Data 80-000808 Hillsborough County S.D. 3/15/16 *DBS*

POEM

HILLSBOROUGH CO SD
Intoxilyzer - Alcotest Analyzer
Model 8800
03/15/2016 10:45:55

Auto Calibration
Max Power Res Value = 48
Auto Range Res Value = 27

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12734, Sum Io = 13973

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0890 (-0.0190)
Sample #2 = 0.1040 (0.0450)
Sample #3 = 0.0850 (0.0560)
Sample #4 = 0.0950 (0.1010)
Avg % Abs = 0.0947 (0.0673)
STD DEV = 0.0095 (0.0297)
REL STD DEV = 10.040 (44.065)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1400 (-0.0190)
Sample #2 = 0.1520 (-0.0090)
Sample #3 = 0.1200 (0.0050)
Sample #4 = 0.1290 (0.0210)
Avg % Abs = 0.1337 (0.0057)
STD DEV = 0.0165 (0.0150)
REL STD DEV = 12.346 (264.902)

Channel 1 Data:
Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12704, Sum Io = 13959

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.5340 (-0.0040)
Sample #2 = 1.5220 (-0.0070)
Sample #3 = 1.5120 (0.0130)
Sample #4 = 1.5260 (0.0140)
Avg % Abs = 1.5200 (0.0067)
STD DEV = 0.0072 (0.0118)
REL STD DEV = 0.474 (177.694)

Channel 1 Data:
Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12701, Sum Io = 13957

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.6290 (-0.0140)
Sample #2 = 3.6140 (0.0110)
Sample #3 = 3.5920 (0.0280)
Sample #4 = 3.6300 (0.0210)
Avg % Abs = 3.6120 (0.0200)
STD DEV = 0.0191 (0.0085)
REL STD DEV = 0.528 (42.720)

Channel 1 Data:
Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12693, Sum Io = 13953

Channel 1 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.095
Std Dev = 0.01 Rel Std Dev = 10.04
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.770
Std Dev = 0.02 Rel Std Dev = 1.99
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.819
Std Dev = 0.02 Rel Std Dev = 0.89

Channel 2 Data:
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.523
Std Dev = 0.02 Rel Std Dev = 0.61
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 6.668
Std Dev = 0.01 Rel Std Dev = 0.19
Zero Order Coef = -203.31
First Order Coef = 2647.19
Second Order Coef = 35.74
Standard Deviation = 44.033665

Channel 1 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.134
Std Dev = 0.02 Rel Std Dev = 12.35
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.520
Std Dev = 0.01 Rel Std Dev = 0.47
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.612
Std Dev = 0.02 Rel Std Dev = 0.53
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.936
Std Dev = 0.01 Rel Std Dev = 0.20
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 12.910
Std Dev = 0.01 Rel Std Dev = 0.07
Zero Order Coef = -137.43
First Order Coef = 1296.34
Second Order Coef = 14.64
Standard Deviation = 34.330444

Channel 2 Data:
Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 3098.00
Sample #1 = 3098.00
Sample #2 = 3066.00
Sample #3 = 3100.00
Sample #4 = 3117.00
Average Result = 3094.3333
STD DEV = 25.9679
REL STD DEV = 0.839

***** CHANNEL 2
Sample #1 = 3274.00
Sample #2 = 3286.00
Sample #3 = 3328.00
Sample #4 = 3295.00
Average Result = 3303.0000
STD DEV = 22.1133
REL STD DEV = 0.669

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1013
3 um H2O Adjust (mg/l*10.000) = 715
9 um H2O Adjust (mg/l*10.000) = 506
**** AUTO CAL PASS

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0010
0.040 0.039 3.0010
0.100 0.099 0.0007
0.200 0.201 -0.0009
0.400 0.400 0.0002

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0008
0.040 0.039 0.0008
0.100 0.099 0.0005
0.200 0.201 -0.0007
0.400 0.400 0.0002

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0008
0.040 0.039 0.0008
0.100 0.099 0.0005
0.200 0.201 -0.0007
0.400 0.400 0.0002

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0008
0.040 0.039 0.0008
0.100 0.099 0.0005
0.200 0.201 -0.0007
0.400 0.400 0.0002

Post Cal

Stability Checks

80-00808 Hillsborough County S.O. 3/15/16 DMB

DDM

SSD

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
03/15/2016
Software: 8100.27

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
03/15/2016
Software: 8100.27

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
03/15/2016
Software: 8100.27

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
03/15/2016
Software: 8100.27

Test	9/21/0L	Time	Test	9/21/0L	Time	Test	9/21/0L	Time	Test	9/21/0L	Time
Air Blank	0.000	11:49	Air Blank	0.000	11:55	Air Blank	0.000	12:00	Air Blank	0.000	12:07
Control Test	0.049	11:50	Control Test	0.079	11:55	Control Test	0.198	12:01	Control Test	0.078	12:07
Air Blank	0.000	11:50	Air Blank	0.000	11:56	Air Blank	0.000	12:02	Air Blank	0.000	12:08
Control Test	0.049	11:51	Control Test	0.078	11:57	Control Test	0.198	12:02	Control Test	0.079	12:08
Air Blank	0.000	11:52	Air Blank	0.000	11:57	Air Blank	0.000	12:03	Air Blank	0.000	12:09
Control Test	0.050	11:52	Control Test	0.078	11:58	Control Test	0.200	12:03	Control Test	0.079	12:09
Air Blank	0.000	11:53	Air Blank	0.000	11:58	Air Blank	0.000	12:04	Air Blank	0.000	12:09
Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats		
Average	0.0493		Average	0.0783		Average	0.1987		Average	0.0787	
Std Dev	0.0005		Std Dev	0.0006		Std Dev	0.0012		Std Dev	0.0006	
Rel Std Dev(%)	1.1703		Rel Std Dev(%)	0.7370		Rel Std Dev(%)	0.5812		Rel Std Dev(%)	0.7339	

KK

DMB

Operator's Signature

DMB

Operator's Signature

DMB

Operator's Signature

DMB

Operator's Signature