



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Dixie County SO

S/N 80-000809

Date In 1/14/16

Date Out 1/20/2016

Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>[Signature]</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>gg</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>219</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>AMP103</u> 32mm <u>.144</u> (.139 - .169) 36mm <u>.164</u> (.156 - .190) 53mm <u>.234</u> (.228 - .278) 103mm <u>.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 07/14/2017</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201502G 02/24/2017</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201505A 05/12/2017</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>A651701 04/23/2017</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 07/14/2017	0.08	SD1011	201502G 02/24/2017	0.20	SD1025	201505A 05/12/2017	0.08 DGS	N/A	A651701 04/23/2017	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 07/14/2017															
0.08	SD1011	201502G 02/24/2017															
0.20	SD1025	201505A 05/12/2017															
0.08 DGS	N/A	A651701 04/23/2017															
		<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____															

RECEIVED  
JAN 21 2016  
Alcohol Testing Program  
FDLE

<b>Optical Bench Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			<b>Department Inspection</b> Performed By <u>gg</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1027</u> Gauge ID# <u>28427</u> <u>1026</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2015-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1027</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD1027	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	SD1027																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	SD1025																																																												
Notes: <u>QC-BK</u> _____ _____ _____	<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____																																																												
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																													

Patrick Murphy  
Quality Control/Review

1/21/16  
Date

Stability Checks - Instrument # 80-000809 Dixie County So 1/20/2016 88

888M

DIXIE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
01/20/2016  
Software: 8100.27  
SN 80-000809

DIXIE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
01/20/2016  
Software: 8100.27  
SN 80-000805

DIXIE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
01/20/2016  
Software: 8100.27  
SN 80-000809

DIXIE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
01/20/2016  
Software: 8100.27  
SN 80-000809

Test	9/21/0L	Time	Test	9/21/0L	Time	Test	9/21/0L	Time	Test	9/21/0L	Time
Air Blank	0.000	10:09	Air Blank	0.000	10:04	Air Blank	0.000	09:55	Air Blank	0.000	09:55
Control Test	0.050	10:09	Control Test	0.203	10:05	Control Test	0.080	09:55	Control Test	0.080	09:55
Air Blank	0.000	10:10	Air Blank	0.000	10:05	Air Blank	0.000	09:56	Air Blank	0.000	09:56
Control Test	0.051	10:11	Control Test	0.203	10:06	Control Test	0.079	09:56	Control Test	0.079	09:56
Air Blank	0.000	10:11	Air Blank	0.000	10:06	Air Blank	0.000	09:57	Air Blank	0.000	09:57
Control Test	0.051	10:12	Control Test	0.204	10:07	Control Test	0.080	09:57	Control Test	0.080	09:57
Air Blank	0.000	10:12	Air Blank	0.000	10:08	Air Blank	0.000		Air Blank	0.000	
Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats		
Average	0.0507		Average	0.2033		Average	0.0797		Average	0.0797	
Std Dev	0.0006		Std Dev	0.0006		Std Dev	0.0006		Std Dev	0.0006	
Rel Std Dev(%)	1.1395		Rel Std Dev(%)	0.2839		Rel Std Dev(%)	0.7247		Rel Std Dev(%)	0.7247	

AK

D65

Operator's Signature

Operator's Signature

Operator's Signature

Operator's Signature