



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Dixie County SO S/N 80-000816
Date In 2/1/16 Date Out 2/4/2016 Ship P/U H/D CMI EE

Intake Performed By TP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By gg

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 171
 Flow Verification (L/s)
 Flow Column # ATP 163
 32mm .148 (.139 - .169)
 36mm .160 (.156 - .190)
 53mm .246 (.228 - .278)
 103mm .527 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 07/14/2017
0.08	SD1011	201502G 02/21/2017
0.20	SD1025	201505A 05/12/2017
0.08 DGS	N/A	AG511701 01/27/2017

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
FEB 05 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By gg

Barometric Pressure 1018 Gauge
 ID# 28427 1018 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC-215

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Derrick Murphy
Quality Control Review

2/5/16
Date

Stability Checks - Instrument # 80-000810 Dixie County SO 2/4/2014 289

DDM

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000810
12/24/2016
Software: 8100.27

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000810
12/24/2016
Software: 8100.27

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000810
12/24/2016
Software: 8100.27

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000810
12/24/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:37
Control Test	0.050	10:37
Air Blank	0.000	10:38
Control Test	0.049	10:39
Air Blank	0.000	10:39
Control Test	0.050	10:40
Air Blank	0.000	10:40
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

Test	9/21/0L	Time
Air Blank	0.000	10:42
Control Test	0.079	10:42
Air Blank	0.000	10:43
Control Test	0.079	10:44
Air Blank	0.000	10:44
Control Test	0.079	10:45
Air Blank	0.000	10:45
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	9/21/0L	Time
Air Blank	0.000	10:47
Control Test	0.201	10:47
Air Blank	0.000	10:48
Control Test	0.202	10:48
Air Blank	0.000	10:49
Control Test	0.202	10:50
Air Blank	0.000	10:50
Control Test Stats		
Average	0.2017	
Std Dev	0.0006	
Rel Std Dev(%)	0.2863	

Test	9/21/0L	Time
Air Blank	0.000	10:33
Control Test	0.000	10:33
Air Blank	0.000	10:33
Control Test	0.079	10:34
Air Blank	0.000	10:34
Control Test	0.000	10:35
Air Blank	0.000	10:35
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

BK

D65

Operator's Signature

Operator's Signature

Operator's Signature

Operator's Signature