

INSTRUMENT PROCESSING SHEET

Agency Perry Police Dept. S/N 80-000811
 Date In 6/16/16 Date Out 6/23/16 Ship P/U H/D CMI EE

Intake Performed By DR

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By DR

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 206
 Flow Verification (L/s)
 Flow Column # AT702
 32mm 152 (.139 - .169)
 36mm 171 (.156 - .190)
 53mm 250 (.228 - .278)
 103mm 523 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp.
0.05	<u>SD3962</u>	<u>201507A</u> <u>7/14/17</u>
0.08	<u>SD3964</u>	<u>201601F</u> <u>1/26/18</u>
0.20	<u>G4444</u>	<u>201505A</u> <u>5/12/17</u>
0.08 DGS	<u>N/A</u>	<u>AG600504</u> <u>1/5/18</u>

Flow Calibration Performed By DR

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 JUN 27 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		<u>N/A</u>	<u>N/A</u>
0.040			
0.100			
0.200			
0.400			
0.080 DGS	<u>N/A</u>		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	<u>N/A</u>		

Department Inspection Performed By DR

Barometric Pressure 1020 Gauge
 ID# 28427 1017 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-13

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2834</u>
0.05	<u>SD3962</u>
0.08	<u>SD3964</u>
0.20	<u>G4444</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: OK DR 6/23/16

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Hankland
 Quality Control Review

6/27/16
 Date

Stability Tests - Perry PD #80-000811 6/23/16

PERRY POLICE DEPT
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000811
06/23/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:10
Control Test	0.048	11:11
Air Blank	0.000	11:12
Control Test	0.047	11:12
Air Blank	0.000	11:13
Control Test	0.048	11:14
Air Blank	0.000	11:14
Control Test Stats		
Average	0.0477	
Std Dev	0.0006	
Rel Std Dev(%)	1.2112	

QWS

Operator's Signature

PERRY POLICE DEPT
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000811
06/23/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:19
Control Test	0.078	11:20
Air Blank	0.000	11:20
Control Test	0.078	11:21
Air Blank	0.000	11:21
Control Test	0.078	11:22
Air Blank	0.000	11:23
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Q

QWS

Operator's Signature

PERRY POLICE DEPT
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000811
06/23/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:23
Control Test	0.196	11:24
Air Blank	0.000	11:25
Control Test	0.196	11:25
Air Blank	0.000	11:26
Control Test	0.198	11:26
Air Blank	0.000	11:27
Control Test Stats		
Average	0.1967	
Std Dev	0.0012	
Rel Std Dev(%)	0.5871	

QWS

Operator's Signature

PERRY POLICE DEPT
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000811
06/23/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:28
Control Test	0.079	11:28
Air Blank	0.000	11:29
Control Test	0.080	11:29
Air Blank	0.000	11:30
Control Test	0.080	11:30
Air Blank	0.000	11:30
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

QWS

AK

QWS

Operator's Signature