



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Hernando County

S/N 80-000814

Date In 7/19/16

Date Out 8/1/16

Ship P/U H/D CMI EE

Intake <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u>	Quality Checks <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>193</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>0.140</u> (.139 - .169) 36mm <u>0.156</u> (.156 - .190) 53mm <u>0.234</u> (.228 - .278) 103mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 100mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)															
Notes: _____	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201505A 5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405 5/3/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201505A 5/12/17	0.08 DGS	N/A	AG612405 5/3/18	Maintenance <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD1011	201601F 1/26/18															
0.20	SD1025	201505A 5/12/17															
0.08 DGS	N/A	AG612405 5/3/18															

RECEIVED
AUG 02 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection <input checked="" type="checkbox"/> Barometric Pressure <u>1017</u> Gauge ID# <u>28427</u> <u>1015</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	SD1019																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	SD1025																																																												

Notes: QA/OC OK GSPM 8/9/16

 Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Wickland
Quality Control Review

8/2/16
Date

Stability Checks # 80-000814 Hernando County S.O. 8/1/16 *RMS*

DS

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000814
08/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:22
Control Test	0.049	18:23
Air Blank	0.000	18:23
Control Test	0.049	18:24
Air Blank	0.000	18:24
Control Test	0.050	18:25
Air Blank	0.000	18:25
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

RMS

Operator's Signature

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000814
08/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:26
Control Test	0.080	18:27
Air Blank	0.000	18:28
Control Test	0.080	18:28
Air Blank	0.000	18:29
Control Test	0.081	18:30
Air Blank	0.000	18:30
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

RMS

Operator's Signature

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000814
08/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:31
Control Test	0.199	18:32
Air Blank	0.000	18:32
Control Test	0.199	18:33
Air Blank	0.000	18:34
Control Test	0.199	18:34
Air Blank	0.000	18:35
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

RMS

Operator's Signature

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000814
08/01/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:37
Control Test	0.080	18:38
Air Blank	0.000	18:38
Control Test	0.081	18:39
Air Blank	0.000	18:39
Control Test	0.080	18:39
Air Blank	0.000	18:40
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

RMS

Operator's Signature

RMS