

INSTRUMENT PROCESSING SHEET

Agency Hickborough County SO S/N 80-000830
 Date In 3/8/16 Date Out 3/15/16 Ship P/U H/D CMI EE

Intake Performed By <u>SP</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>RWB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>146</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32mm <u>0.144</u> (.139 - .169) 36mm <u>0.164</u> (.156 - .190) 53mm <u>0.230</u> (.228 - .278) 103mm <u>0.496</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>64444</td> <td>201505A 5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG605301 2/22/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	64444	201505A 5/12/17	0.08 DGS	N/A	AG605301 2/22/18	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/A Program) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD1011	201601F 1/26/18															
0.20	64444	201505A 5/12/17															
0.08 DGS	N/A	AG605301 2/22/18															
Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____ _____																	

RECEIVED
 MAR 21 2016
 FDLE
 Alcohol Test (L/A) Program

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____																							
Simulator	Serial Number	Lot Number	Expiration																				
0.000		N/A	N/A																				
0.040																							
0.100																							
0.200																							
0.400																							
0.080 DGS	N/A																						
<input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																				
0.05																							
0.08																							
0.20																							
0.08 DGS	N/A																						

Department Inspection Performed By <u>CMR</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1014</u> Gauge ID# <u>28427</u> <u>1010</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-B</u>	
Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	64444
Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Other _____	

Notes: QC - BK

- Instrument Complies with Chapter 11D-8, FAC
- Instrument Does Not Comply with Chapter 11D-8, FAC
- Return to/Place into Evidentiary Use
- Remain Out of Evidentiary Use
- Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy
 Quality Control Review

3/21/16
 Date

Stability Checks 80-000830 Hillsborough County S.O. 3/9/16 RMB

RMB

DS
 HILLSBOROUGH CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000830
 03/09/2016
 Software: 8100.27

HILLSBOROUGH CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000830
 03/09/2016
 Software: 8100.27

HILLSBOROUGH CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000830
 03/09/2016
 Software: 8100.27

HILLSBOROUGH CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000830
 03/09/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:51
Control Test	0.079	15:52
Air Blank	0.000	15:52
Control Test	0.080	15:52
Air Blank	0.000	15:53
Control Test	0.079	15:53
Air Blank	0.000	15:54
Control Test	0.079	15:54
Average	0.0733	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Test	g/210L	Time
Air Blank	0.000	15:46
Control Test	0.197	15:47
Air Blank	0.000	15:47
Control Test	0.197	15:48
Air Blank	0.000	15:49
Control Test	0.197	15:49
Air Blank	0.000	15:50
Control Test	0.197	15:50
Average	0.1970	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	15:34
Control Test	0.079	15:35
Air Blank	0.000	15:36
Control Test	0.079	15:36
Air Blank	0.000	15:37
Control Test	0.079	15:37
Air Blank	0.000	15:38
Control Test	0.079	15:38
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	15:40
Control Test	0.049	15:41
Air Blank	0.000	15:41
Control Test	0.049	15:42
Air Blank	0.000	15:43
Control Test	0.050	15:43
Air Blank	0.000	15:44
Control Test	0.0493	15:44
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

SK

RMB
 Operator's Signature

RMB
 Operator's Signature

RMB
 Operator's Signature

RMB
 Operator's Signature