

## INSTRUMENT PROCESSING SHEET

Agency Hillsborough County S/N 80-000833  
 Date In 10/14/16 Date Out 10/27/16  Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>DB</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: OK Case OK Handle OK Dry Gas Holder OK Feet OK Keyboard/Plug OK Back/Plugs OK Screws tight OK Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>DB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>SDS</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>0.160</u> (.139 - .169) 36mm <u>0.175</u> (.156 - .190) 53mm <u>0.242</u> (.228 - .278) 103mm <u>0.527</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AP1619605 7/14/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AP1619605 7/14/18	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD1011	201601F 1/26/18															
0.20	SD1025	201604C 4/5/18															
0.08 DGS	N/A	AP1619605 7/14/18															
<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																	
<b>Suggested Service</b> _____ _____																	

RECEIVED  
 OCT 27 2016  
 ALCOHOL TESTING PROGRAM  
 FDLE

<b>Optical Bench Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			<b>Department Inspection</b> Performed By <u>DB</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1025</u> Gauge ID# <u>28427</u> <u>1025</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	SD1019																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	SD1025																																																												
<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____																																																													

Notes: Brought to FDLE after battery change at Enforcement Electronics DB  
DB-SP

- Instrument Complies with Chapter 11D-8, FAC
- Instrument Does Not Comply with Chapter 11D-8, FAC
- Return to/Place into Evidentiary Use
- Remain Out of Evidentiary Use
- Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland  
 Quality Control Review

10/27/16  
 Date

Stability Checks # 80-000833 Hillsborough County S.O. 10/27/16 *DBS*

*DBS*

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000833  
10/27/2016  
Software: 8100.27

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000833  
10/27/2016  
Software: 8100.27

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000833  
10/27/2016  
Software: 8100.27

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000833  
10/27/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:58
Control Test	0.049	10:59
Air Blank	0.000	10:59
Control Test	0.050	11:00
Air Blank	0.000	11:00
Control Test	0.050	11:01
Air Blank	0.000	11:02
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

Test	9/21/0L	Time
Air Blank	0.000	11:04
Control Test	0.079	11:04
Air Blank	0.000	11:05
Control Test	0.080	11:05
Air Blank	0.000	11:06
Control Test	0.079	11:06
Air Blank	0.000	11:07
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Test	9/21/0L	Time
Air Blank	0.000	11:10
Control Test	0.197	11:10
Air Blank	0.000	11:11
Control Test	0.197	11:12
Air Blank	0.000	11:12
Control Test	0.196	11:13
Air Blank	0.000	11:13
Control Test Stats		
Average	0.1967	
Std Dev	0.0006	
Rel Std Dev(%)	0.2936	

Test	9/21/0L	Time
Air Blank	0.000	11:15
Control Test	0.080	11:16
Air Blank	0.000	11:16
Control Test	0.080	11:17
Air Blank	0.000	11:17
Control Test	0.080	11:17
Air Blank	0.000	11:18
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*BSK*

*DBS*  
Operator's Signature

*DBS*  
Operator's Signature

*DBS*  
Operator's Signature

*DBS*  
Operator's Signature

*SP*

## INSTRUMENT PROCESSING SHEET

Agency Hillsborough County SO S/N 80-000833  
 Date In 6/22/16 Date Out 7/14/16  Ship  P/U  H/D  CMI  EE

**Intake** Performed By CS

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
 Case  Handle  
 Dry Gas Holder  Feet  
 Keyboard/Plug  Back/Plugs  
 Screws tight  Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By RMS

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 212  
 Flow Verification (L/s)  
 Flow Column # APP 102  
 32mm 0.160 (.139 - .169)  
 36mm 0.175 (.156 - .190)  
 53mm 0.242 (.228 - .278)  
 103mm 0.515 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks 6/23/16

Simulator	Serial #	Lot #/Exp
0.05	SD3962	201507A 7/14/17
0.08	SD3964	201601F 7/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AG1600504 1/5/18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm  
 R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service** \_\_\_\_\_

RECEIVED  
 JUL 15 2016  
 FDLE  
 Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By RMS

Barometric Pressure  
 ID# 28427 Gauge 1019  
 Instrument 1019

Mouth Alcohol Solution Lot # 2015-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

**Attachments**

Form 41  
 Pre-Stability Tests v. 2  
 Flow Calibration  
 Optical Bench Cal  
 Post-Stability Tests  
 Other Form 40

Notes: Simulators & Lot #s for 2<sup>nd</sup> set of  
Quality Checks are indicated on Stability  
Check Results Sheets @MS  
Lab environment was incredibly cold. AdH checks  
needed @MS  
QC @ 7/15/16

Brett Kerkhove  
 Quality Control Review

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

7/15/16  
 Date

Stability Checks 80-000833 Hillsborough County S.O. 6/23/16 ~~8008~~ BK

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
06/23/2016  
Software: 8100.27

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
06/23/2016  
Software: 8100.27

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
06/23/2016  
Software: 8100.27

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
06/23/2016  
Software: 8100.27

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
06/23/2016  
Software: 8100.27

Test	9/210L	Time	9/210L	Time
Air Blank	0.000	16:55	Air Blank	0.000
Control Test	0.049	16:56	Control Test	0.196
Air Blank	0.000	16:57	Air Blank	0.000
Control Test	0.049	16:57	Control Test	0.195
Air Blank	0.000	16:58	Air Blank	0.000
Control Test	0.050	16:59	Control Test	0.195
Air Blank	0.000	16:59	Air Blank	0.000
Control Test Stats			Control Test Stats	
Average	0.0493		Average	0.1953
Std Dev	0.0016		Std Dev	0.0006
Rel Std Dev(%)	1.1703		Rel Std Dev(%)	0.2956

Test	9/210L	Time	9/210L	Time
Air Blank	0.000	16:50	Air Blank	0.000
Control Test	0.192	16:50	Control Test	0.192
Air Blank	0.000	16:51	Air Blank	0.000
Control Test	0.195	16:52	Control Test	0.195
Air Blank	0.000	16:52	Air Blank	0.000
Control Test	0.196	16:53	Control Test	0.195
Air Blank	0.000	16:54	Air Blank	0.000
Control Test Stats			Control Test Stats	
Average	0.1943		Average	0.1940
Std Dev	0.0021		Std Dev	0.0017
Rel Std Dev(%)	1.0712		Rel Std Dev(%)	0.8928

Test	9/210L	Time	9/210L	Time
Air Blank	0.000	17:00	Air Blank	0.000
Control Test	0.078	17:01	Control Test	0.192
Air Blank	0.000	17:02	Air Blank	0.000
Control Test	0.078	17:02	Control Test	0.195
Air Blank	0.000	17:03	Air Blank	0.000
Control Test	0.079	17:04	Control Test	0.196
Air Blank	0.000	17:04	Air Blank	0.000
Control Test Stats			Control Test Stats	
Average	0.0783		Average	0.1940
Std Dev	0.0006		Std Dev	0.0017
Rel Std Dev(%)	0.7370		Rel Std Dev(%)	0.8928

Test	9/210L	Time	9/210L	Time
Air Blank	0.000	16:55	Air Blank	0.000
Control Test	0.049	16:56	Control Test	0.192
Air Blank	0.000	16:57	Air Blank	0.000
Control Test	0.049	16:57	Control Test	0.195
Air Blank	0.000	16:58	Air Blank	0.000
Control Test	0.050	16:59	Control Test	0.195
Air Blank	0.000	16:59	Air Blank	0.000
Control Test Stats			Control Test Stats	
Average	0.0493		Average	0.1940
Std Dev	0.0016		Std Dev	0.0017
Rel Std Dev(%)	1.1703		Rel Std Dev(%)	0.8928

Test	9/210L	Time	9/210L	Time
Air Blank	0.000	17:06	Air Blank	0.000
Control Test	0.192	17:07	Control Test	0.196
Air Blank	0.000	17:08	Air Blank	0.000
Control Test	0.195	17:08	Control Test	0.195
Air Blank	0.000	17:09	Air Blank	0.000
Control Test	0.195	17:10	Control Test	0.195
Air Blank	0.000	17:10	Air Blank	0.000
Control Test Stats			Control Test Stats	
Average	0.1940		Average	0.1953
Std Dev	0.0017		Std Dev	0.0006
Rel Std Dev(%)	0.8928		Rel Std Dev(%)	0.2956

1st Conf. Test - tubing had not warmed up yet.

Changed tubing on simulator.

*[Signature]*  
Operator's Signature

*[Signature]*  
Operator's Signature

*[Signature]*  
Operator's Signature

*[Signature]*  
Operator's Signature

*[Signature]*  
Operator's Signature

Stability Checks # 80-000833

Hillsborough County S.O. 6/23/16  
BK

003

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000833  
06/23/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	17:15
Control Test	0.080	17:16
Air Blank	0.000	17:16
Control Test	0.080	17:17
Air Blank	0.000	17:17
Control Test	0.080	17:17
Air Blank	0.000	17:18
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

④



Operator's Signature

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000833  
07/06/2016  
Software: 8100.27

0.05

*DMS*

SD1018

201507A

7/14/17

Test	g/210L	Time
Air Blank	0.000	15:57
Control Test	0.046	15:58
Air Blank	0.000	15:58
Control Test	0.047	15:59
Air Blank	0.000	15:59
Control Test	0.047	16:00
Air Blank	0.000	16:01
Control Test Stats		
Average	0.0467	
Std Dev	0.0006	
Rel Std Dev(%)	1.2372	

0.05 Retest *DMS*

*DMS*

Operator's Signature

HILLSBOROUGH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000833  
07/06/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:01
Control Test	0.048	16:02
Air Blank	0.000	16:03
Control Test	0.048	16:03
Air Blank	0.000	16:04
Control Test	0.048	16:05
Air Blank	0.000	16:05
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Retested after allowing  
cold tubing to warm.  
(Replaced just prior to test.)

*DMS*

*✓*

*134*

*DMS*

Operator's Signature

Note: Lab environment was extremely cold and tubing on simulators was having trouble heating as normal. Environmental conditions corrected during subsequent Department inspection.

*DMS*

Stability Checks Day #2

80-000833

Hillsborough County S.O.

7/6/16 ~~DBS~~

BK

D.08 7/21/16 Retest

80

HILLSBOROUGH CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000833  
 07/06/2016  
 Software: 8100.27

HILLSBOROUGH CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000833  
 07/06/2016  
 Software: 8100.27

HILLSBOROUGH CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000833  
 07/06/2016  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:06
Control Test	0.076	16:07
Air Blank	0.000	16:07
Control Test	0.077	16:08
Air Blank	0.000	16:09
Control Test	0.078	16:09
Air Blank	0.000	16:10
Control Test Stats		
Average	0.0770	
Std Dev	0.0010	
Rel Std Dev(%)	1.2987	

Test	g/210L	Time
Air Blank	0.000	16:27
Control Test	0.079	16:27
Air Blank	0.000	16:28
Control Test	0.079	16:29
Air Blank	0.000	16:29
Control Test	0.079	16:30
Air Blank	0.000	16:30
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	16:16
Control Test	0.196	16:16
Air Blank	0.000	16:17
Control Test	0.197	16:17
Air Blank	0.000	16:18
Control Test	0.196	16:19
Air Blank	0.000	16:19
Control Test Stats		
Average	0.1963	
Std Dev	0.0006	
Rel Std Dev(%)	0.2941	

Test	g/210L	Time
Air Blank	0.000	16:12
Control Test	0.078	16:12
Air Blank	0.000	16:13
Control Test	0.079	16:13
Air Blank	0.000	16:14
Control Test	0.079	16:14
Air Blank	0.000	16:14
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7739	

*DBS*  
Operator's Signature

*DBS*  
Operator's Signature

*DBS*  
Operator's Signature

*DBS*  
Operator's Signature

Cold tubing  
Retest.

0.08 ~~DBS~~  
SD1011  
201601F  
01/26/18

0.20  
SD1025  
201505A  
5/12/17

0.08 ~~DBS~~  
AG605301  
2/22/18



Alcohol Testing Program

**INSTRUMENT PROCESSING SHEET**

Agency Hillsborough County SO S/N 80-000833

Date In 2/5/16

Date Out 2/23/16

Ship  P/U  H/D  CMI  EE

<b>Intake</b> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>DWS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>215</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATR103</u> 32mm <u>0.144</u> (.139 - .169) 36mm <u>0.167</u> (.156 - .190) 53mm <u>0.238</u> (.228 - .278) 103mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>G11739</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>G8149</td> <td>201502G 2/24/17</td> </tr> <tr> <td>0.20</td> <td>G11621</td> <td>201505A 5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>10511701 4/27/17</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	G11739	201507A 7/14/17	0.08	G8149	201502G 2/24/17	0.20	G11621	201505A 5/12/17	0.08 DGS	N/A	10511701 4/27/17	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	G11739	201507A 7/14/17															
0.08	G8149	201502G 2/24/17															
0.20	G11621	201505A 5/12/17															
0.08 DGS	N/A	10511701 4/27/17															
		<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____															

RECEIVED  
FEB 23 2016  
FDLE  
Alcohol Testing Program

<b>Optical Bench Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																													
0.000		N/A	N/A																																													
0.040																																																
0.100																																																
0.200																																																
0.400																																																
0.080 DGS	N/A																																															
Simulator	Serial Number	Lot Number	Expiration																																													
0.05																																																
0.08																																																
0.20																																																
0.08 DGS	N/A																																															

<b>Department Inspection</b> Performed By <u>DWS</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1010</u> Gauge ID# <u>26932</u> <u>1009</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2015-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>G2879</td> </tr> <tr> <td>Interferent</td> <td>G8144</td> </tr> <tr> <td>0.05</td> <td>G11739</td> </tr> <tr> <td>0.08</td> <td>G8149</td> </tr> <tr> <td>0.20</td> <td>G11621</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	G2879	Interferent	G8144	0.05	G11739	0.08	G8149	0.20	G11621
Simulator	Serial Number											
0.00	G2879											
Interferent	G8144											
0.05	G11739											
0.08	G8149											
0.20	G11621											
<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Other _____												

Notes: QC-98  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy  
Quality Control Review

2/23/16  
Date



Stability Checks 80-000833 Hillsborough County S.O. 2/23/16 *RMS*

*RMS*

*RMS*

HILLSBOROUGH CO SO  
Intoxilyzer - Alcotest Analyzer  
Model 8000  
02/23/2016  
Software: 8100.27  
SN 80-000833

HILLSBOROUGH CO SO  
Intoxilyzer - Alcotest Analyzer  
Model 8000  
02/23/2016  
Software: 8100.27  
SN 80-000833

HILLSBOROUGH CO SO  
Intoxilyzer - Alcotest Analyzer  
Model 8000  
02/23/2016  
Software: 8100.27  
SN 80-000833

HILLSBOROUGH CO SO  
Intoxilyzer - Alcotest Analyzer  
Model 8000  
02/23/2016  
Software: 8100.27  
SN 80-000833

Test	9/21/0L	9/21/0L	Time
Air Blank	0.000	0.000	10:28
Control Test	0.051	0.060	10:28
Air Blank	0.000	0.000	10:29
Control Test	0.051	0.060	10:29
Air Blank	0.000	0.000	10:29
Control Test	0.051	0.061	10:29
Air Blank	0.000	0.001	10:30
Control Test	0.000	0.009	10:30
Control Test Stats			
Average	0.0510	0.0603	
Std Dev	0.0090	0.0096	
Rel Std Dev(%)	0.0090	0.7787	

Test	9/21/0L	9/21/0L	Time
Air Blank	0.000	0.000	10:22
Control Test	0.204	0.204	10:23
Air Blank	0.000	0.000	10:23
Control Test	0.204	0.204	10:24
Air Blank	0.000	0.000	10:25
Control Test	0.204	0.204	10:25
Air Blank	0.000	0.000	10:26
Control Test Stats			
Average	0.2040	0.2040	
Std Dev	0.0000	0.0000	
Rel Std Dev(%)	0.0000	0.0000	

Test	9/21/0L	9/21/0L	Time
Air Blank	0.000	0.000	10:17
Control Test	0.081	0.081	10:18
Air Blank	0.000	0.000	10:19
Control Test	0.081	0.081	10:19
Air Blank	0.000	0.000	10:20
Control Test	0.080	0.080	10:21
Air Blank	0.000	0.000	10:21
Control Test Stats			
Average	0.0807	0.0807	
Std Dev	0.0096	0.0096	
Rel Std Dev(%)	0.7157	0.7157	

Test	9/21/0L	9/21/0L	Time
Air Blank	0.000	0.000	10:31
Control Test	0.051	0.051	10:32
Air Blank	0.000	0.000	10:32
Control Test	0.051	0.051	10:33
Air Blank	0.000	0.000	10:34
Control Test	0.051	0.051	10:34
Air Blank	0.000	0.000	10:35
Control Test Stats			
Average	0.0510	0.0510	
Std Dev	0.0090	0.0090	
Rel Std Dev(%)	0.0090	0.0090	

*RMS*

Operator's Signature

*RMS*

Operator's Signature

*RMS*

Operator's Signature

*RMS*

Operator's Signature