



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Panama City Beach PD S/N 90-000838
 Date In 8/1/16 Date Out 8/2/16 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	Performed By <u>DS</u>	Quality Checks	Performed By <u>DS</u>	Flow Calibration	Performed By <u>DS</u>												
☐ Registration ☒ Annual ☐ Return from CMI ☐ Return from Enforcement ☐ Electronics ☐ Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: ☐ Power cord ☐ Printer Cable ☒ Other <u>Static Bag</u> Notes: _____ _____ _____ _____	☒ Breath Tube Screen ☒ Replace O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>190</u> ☒ Flow Verification (L/s) Flow Column # <u>ADP02</u> 32mm <u>156</u> (.139 - .169) 36mm <u>175</u> (.156 - .190) 53mm <u>242</u> (.228 - .278) 103mm <u>519</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG605301 2/22/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AG605301 2/22/18	☒ Flow Calibration N/A ☐ Flow Calibration Complete Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 7/14/17															
0.08	SD1011	201601F 1/26/18															
0.20	SD1025	201604C 4/5/18															
0.08 DGS	N/A	AG605301 2/22/18															
		Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____		Suggested Service _____ _____ _____													

RECEIVED
 AUG 03 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration	Performed By _____	Department Inspection	Performed By <u>DS</u>																																																										
☒ Optical Bench Calibration N/A ☐ Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			☒ Barometric Pressure <u>1020</u> Gauge ID# <u>28427</u> <u>1018</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	SD1019																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	SD1025																																																												
Attachments ☒ Form 41 ☒ Pre-Stability Tests ☐ Flow Calibration ☐ Optical Bench Cal ☐ Post-Stability Tests ☐ Other _____																																																													

Notes: QA/AC OK RSM 8/3/16

☒ Instrument Complies with Chapter 11D-8, FAC
☐ Instrument Does Not Comply with Chapter 11D-8, FAC
☒ Return to/Place into Evidentiary Use
☐ Remain Out of Evidentiary Use
☒ Conduct an Agency Inspection Before Evidentiary Use

Brett Kunkland

Quality Control Review

8/3/16
 Date

Stability Tests - Panama City Beach PD #80-000838 8/2/16 BK

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000
08/02/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:38
Control Test	0.048	09:39
Air Blank	0.000	09:39
Control Test	0.049	09:40
Air Blank	0.000	09:40
Control Test	0.049	09:41
Air Blank	0.000	09:42
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

Operator's Signature

Operator's Signature

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000
08/02/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:46
Control Test	0.079	09:46
Air Blank	0.000	09:47
Control Test	0.079	09:48
Air Blank	0.000	09:48
Control Test	0.080	09:49
Air Blank	0.000	09:49
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Operator's Signature

Operator's Signature

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000
08/02/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:33
Control Test	0.195	09:34
Air Blank	0.000	09:34
Control Test	0.196	09:35
Air Blank	0.000	09:36
Control Test	0.196	09:36
Air Blank	0.000	09:37
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	

Operator's Signature

Operator's Signature

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000
08/02/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:51
Control Test	0.080	09:51
Air Blank	0.000	09:52
Control Test	0.080	09:52
Air Blank	0.000	09:53
Control Test	0.080	09:53
Air Blank	0.000	09:54
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Operator's Signature

263