

INSTRUMENT PROCESSING SHEET

Agency Lynn Haven PD S/N 80-000840
 Date In 6/3/16 Date Out 6/9/16 Ship P/U H/D CMI EE

Intake Performed By PT

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By PW

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 206
 Flow Verification (L/s)
 Flow Column # AMP02
 32mm 148 (.139 - .169)
 36mm 167 (.156 - .190)
 53mm 234 (.228 - .278)
 103mm 492 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G2403	201507A 7/14/17
0.08	SD3964	201601F 1/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AG600504 1/5/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 40mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s) Program

Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 JUN 10 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By PW

Barometric Pressure
 ID# 28427 1013 Gauge
1014 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2880
Interferent	G2834
0.05	G2403
0.08	SD3964
0.20	R4444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other Form 40

Notes: QA/OC OK (PSPM)

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
 Quality Control Review

6/10/16
 Date

Stability Tests - Lynn Haven PD # 80-000840 6/2/16

BK

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
06/09/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:11
Control Test	0.050	10:11
Air Blank	0.000	10:12
Control Test	0.051	10:13
Air Blank	0.000	10:13
Control Test	0.051	10:14
Air Blank	0.000	10:14
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1355	

PWS

Operator's Signature

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
06/09/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:16
Control Test	0.080	10:17
Air Blank	0.000	10:17
Control Test	0.081	10:18
Air Blank	0.000	10:19
Control Test	0.081	10:19
Air Blank	0.000	10:20
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

PWS

Operator's Signature

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
06/09/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:59
Control Test	0.201	10:00
Air Blank	0.000	10:01
Control Test	0.202	10:01
Air Blank	0.000	10:02
Control Test	0.202	10:02
Air Blank	0.000	10:03
Control Test Stats		
Average	0.2017	
Std Dev	0.0006	
Rel Std Dev(%)	0.2863	

PWS

Operator's Signature

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
06/09/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:05
Control Test	0.079	10:05
Air Blank	0.000	10:06
Control Test	0.079	10:06
Air Blank	0.000	10:07
Control Test	0.078	10:07
Air Blank	0.000	10:07
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

DGS

PWS

Operator's Signature