

**INSTRUMENT PROCESSING SHEET**

Agency Blountstown PD S/N 80-000862

Date In 12/7/16 Date Out 12/8/16  Ship  P/U  H/D  CMI  EE

**Intake** Performed By (Signature)

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other \_\_\_\_\_

Notes: \_\_\_\_\_

**Quality Checks** Performed By (Signature)

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 200  
 Flow Verification (L/s)  
 Flow Column # ATP102  
 32mm .148 (.139 - .169)  
 36mm .167 (.156 - .190)  
 53mm .230 (.228 - .278)  
 103mm .492 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>SD1011</u>	<u>201601F</u> <u>1/26/18</u>
0.20	<u>SD1025</u>	<u>201604G</u> <u>4/5/18</u>
0.08 DGS	<u>N/A</u>	<u>AG636605</u> <u>9/22/18</u>

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 15mm  
 15L/min - 53mm  
 30L/min - 103mm  
 R-Value \_\_\_\_\_  
 Post Calibration Verification (L/gram)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

RECEIVED  
DEC 09 2016  
FDLE  
Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete  
 Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		<u>N/A</u>	<u>N/A</u>
0.040			
0.100			
0.200			
0.400			
0.080 DGS	<u>N/A</u>		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	<u>N/A</u>		

**Department Inspection** Performed By (Signature)

Barometric Pressure 1022 Gauge  
 ID# 28427 1015 Instrument

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>DR2035</u>
0.08	<u>SD1011</u>
0.20	<u>SD1025</u>

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration  
 Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: QA/QC OK SP

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kuitland  
Quality Control Review

12/9/16  
Date

BLOUNTSTOWN PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000862  
12/08/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:20
Control Test	0.050	09:21
Air Blank	0.000	09:22
Control Test	0.050	09:22
Air Blank	0.000	09:23
Control Test	0.050	09:24
Air Blank	0.000	09:24
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

  
-----  
Operator's Signature

BLOUNTSTOWN PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000862  
12/08/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:13
Control Test	0.081	09:14
Air Blank	0.000	09:14
Control Test	0.081	09:15
Air Blank	0.000	09:16
Control Test	0.081	09:16
Air Blank	0.000	09:17
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

  
-----  
Operator's Signature

SK

BLOUNTSTOWN PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000862  
12/08/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:30
Control Test	0.198	09:31
Air Blank	0.000	09:31
Control Test	0.197	09:32
Air Blank	0.000	09:32
Control Test	0.195	09:33
Air Blank	0.000	09:34
Control Test Stats		
Average	0.1967	
Std Dev	0.0015	
Rel Std Dev(%)	0.7767	

  
-----  
Operator's Signature

BLOUNTSTOWN PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000862  
12/08/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:41
Control Test	0.080	09:42
Air Blank	0.000	09:42
Control Test	0.081	09:42
Air Blank	0.000	09:43
Control Test	0.081	09:43
Air Blank	0.000	09:44
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

SP

DGS

  
-----  
Operator's Signature