



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Miami Dade Police Department S/N 80-000884

Date In 2/9/2016 Date Out 2/9/2016 Ship P/U H/D CMI EE

Intake Performed By DELL

- Registration
- Annual
- Return from CMI
- Return from Enforcement
- Electronics
- Other _____

Visual Inspection:

OK Case OK Handle

OK Dry Gas Holder OK Feet

OK Keyboard/Plug OK Back/Plugs

OK Screws tight OK Breath Hose

- Other Equipment:
- Power cord
 - Printer Cable
 - Other _____

Notes: _____

Quality Checks Performed By DELL

- Breath Tube Screen
 - Replace O-Rings
 - Instrument Set Up Verified
 - R-Value 192
 - Flow Verification (L/s)
- Flow Column # ATP 101
- 32mm 136 (.139 - .169)
- 36mm 156 (.156 - .190)
- 53mm 236 (.228 - .278)
- 103mm 511 (.447 - .547)

- Barometric Pressure Check
- Gauge ID # 68639
- Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3967	201507A 07/14/2017
0.08	SD2968	2015026 02/24/2017
0.20	SD3969	201505A 05/12/2017
0.08 DGS	N/A	AG507503 03/16/2017

Flow Calibration Performed By DELL

- Flow Calibration N/A
 - Flow Calibration Complete
 - Flow Column # ATP 104
 - 5L/min - 17mm
 - 15L/min - 53mm
 - 30L/min - 103mm
 - R-Value 192
 - Post Calibration Verification (U.S. Program)
- Flow Column # ATP 101
- 32mm 152 (.139 - .169)
- 36mm 167 (.156 - .190)
- 53mm 246 (.228 - .278)
- 103mm 511 (.447 - .547)

Maintenance Performed By _____

- Battery Replacement
- Dry Gas Regulator Replacement
- Breath Tube Replacement
- Other _____

Suggested Service _____

RECEIVED
FEB 26 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By DELL

- Optical Bench Calibration N/A
 - Optical Bench Calibration Complete
- Barometric Pressure Gauge 1018 ID# 28663

Simulator	Serial Number	Lot Number	Expiration
0.000	<u>2235</u>	N/A	N/A
0.040	<u>2108</u>	<u>15108</u>	<u>08-18-2017</u>
0.100	<u>2237</u>	<u>15001</u>	<u>05-20-2017</u>
0.200	<u>2238</u>	<u>15104</u>	<u>05-27-2017</u>
0.400	<u>2239</u>	<u>15105</u>	<u>06-10-2017</u>
0.080 DGS	N/A	<u>03415080A1</u>	<u>03-05-2017</u>

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05	<u>SD3967</u>	<u>201507A</u>	<u>07/14/2017</u>
0.08	<u>SD3968</u>	<u>2015026</u>	<u>02/24/2017</u>
0.20	<u>SD3969</u>	<u>201505A</u>	<u>05/12/2017</u>
0.08 DGS	N/A	<u>AG507503</u>	<u>03/16/2017</u>

Notes: E-MAILED APPROVED

Recalibrated to bring values closer to nominal Barometric Pressure.

Patrick Murphy
Quality Control Review

Department Inspection Performed By DELL

- Barometric Pressure 1018 Gauge
- ID# 68639 1017 Instrument

Mouth Alcohol Solution Lot # 2015-A

Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	<u>SD3965</u>
Interferent	<u>SD3966</u>
0.05	<u>SD3967</u>
0.08	<u>SD3968</u>
0.20	<u>SD3969</u>

Attachments

- Form 41
- Pre-Stability Tests
- Flow Calibration
- Optical Bench Cal
- Post-Stability Tests
- Other _____

- Instrument Complies with Chapter 11D-8, FAC
- Instrument Does Not Comply with Chapter 11D-8, FAC

- Return to/Place into Evidentiary Use
- Remain Out of Evidentiary Use
- Conduct an Agency Inspection Before Evidentiary Use

2/26/17
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Post Stabilities	80-000884	Miami Dade Police Department	02/09/2016	<i>Will</i>

Page 1

0.056/210L	0.086/210L	0.206/210L	DGS 0.086/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8605 SN 80-000884
02/09/2016
Software: 8100.27

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8605 SN 80-000884
02/09/2016
Software: 8100.27

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8605 SN 80-000884
02/09/2016
Software: 8100.27

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8605 SN 80-000884
02/09/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.009	11:52
Control Test	0.049	11:52
Air Blank	0.000	11:53
Control Test	0.050	11:54
Air Blank	0.000	11:54
Control Test	0.050	11:55
Air Blank	0.000	11:55
Control Test	0.000	11:55
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

Test	g/210L	Time
Air Blank	0.008	11:57
Control Test	0.078	11:58
Air Blank	0.000	11:58
Control Test	0.079	11:59
Air Blank	0.000	12:00
Control Test	0.079	12:00
Air Blank	0.000	12:01
Control Test	0.000	12:01
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Test	g/210L	Time
Air Blank	0.000	12:02
Control Test	0.201	12:03
Air Blank	0.000	12:04
Control Test	0.201	12:04
Air Blank	0.000	12:05
Control Test	0.201	12:05
Air Blank	0.000	12:06
Control Test	0.201	12:06
Average	0.2010	
Std Dev	0.0003	
Rel Std Dev(%)	0.9003	

Test	g/210L	Time
Air Blank	0.000	12:08
Control Test	0.080	12:08
Air Blank	0.000	12:09
Control Test	0.080	12:09
Air Blank	0.000	12:09
Control Test	0.080	12:09
Air Blank	0.000	12:10
Control Test	0.080	12:10
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Will
Operator's Signature

Will
Operator's Signature

Will
Operator's Signature

Will
Operator's Signature

1941-0926 99
 Model Analyzer - Alcoron Analyzer
 SN 80-000884
 Date: 8/10/16 10:51:27

***** CHANNEL 2 *****
 Sample % AOS (% AOS Ref)
 Sample #1 = 1.4956 (0.0113)
 Sample #2 = 1.4380 (0.0133)
 Sample #3 = 1.5770 (0.0083)
 Sample #4 = 1.4880 (0.0110)
 Avg % AOS = 1.4977 (0.0107)
 STD DEV = 0.0095 (0.0025)
 REL STD DEV = 0.635 (23.593)

***** CHANNEL 2 *****
 Sample % AOS (% AOS Ref)
 Sample #1 = 7.0070 (-0.0010)
 Sample #2 = 6.9520 (0.0540)
 Sample #3 = 6.9620 (0.0740)
 Sample #4 = 6.9860 (0.0830)
 Avg % AOS = 6.9133 (0.0703)
 STD DEV = 0.0344 (0.0148)
 REL STD DEV = 0.498 (21.195)

***** CHANNEL 2 *****
 Sample % AOS (% AOS Ref)
 Sample #1 = 6.6790 (-0.0050)
 Sample #2 = 6.6830 (0.0430)
 Sample #3 = 6.6230 (0.0820)
 Sample #4 = 6.6340 (0.0920)
 Avg % AOS = 6.6457 (0.0723)
 STD DEV = 0.0331 (0.0259)
 REL STD DEV = 0.499 (35.794)

***** CHANNEL 2 *****
 Sol Val = 0.0088 mg/l or 0.000 g/210L
 % AOS = 0.105
 Std Dev = 0.01 Rel Std Dev = 9.76
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % AOS = 1.498
 Std Dev = 0.01 Rel Std Dev = 0.63
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % AOS = 3.622
 Std Dev = 0.01 Rel Std Dev = 0.29
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % AOS = 6.313
 Std Dev = 0.03 Rel Std Dev = 0.50
 Sol Val = 1.9048 mg/l or 0.400 g/210L
 % AOS = 12.929
 Std Dev = 0.05 Rel Std Dev = 0.42
 Zero Order Coef = -109.10
 First Order Coef = 1295.06
 Second Order Coef = 14.41
 Standard Deviation = 25.725883

*** Channel 1 ***
 Fit Value = 0.008 g/210L
 Samples Taken = 4, Discarded = 1
 Sum 10 = 12272, Sum 10 = 14227

***** CHANNEL 1 *****
 Sample % AOS (% AOS Ref)
 Sample #1 = 1.8060 (-0.0110)
 Sample #2 = 1.8150 (0.0050)
 Sample #3 = 1.8310 (-0.0010)
 Sample #4 = 1.8070 (0.0370)
 Avg % AOS = 1.8177 (0.0137)
 STD DEV = 0.0122 (0.0204)
 REL STD DEV = 0.672 (199.479)

***** CHANNEL 1 *****
 Sample % AOS (% AOS Ref)
 Sample #1 = -0.0060 (-0.0080)
 Sample #2 = -0.0040 (-0.0200)
 Sample #3 = -0.0160 (-0.0240)
 Sample #4 = 0.0040 (-0.0440)
 Avg % AOS = -0.0053 (-0.0293)
 STD DEV = 0.0161 (0.0129)
 REL STD DEV = 188.746 (43.835)

***** CHANNEL 2 *****
 Sample % AOS (% AOS Ref)
 Sample #1 = 13.0520 (-0.0090)
 Sample #2 = 12.9820 (0.1070)
 Sample #3 = 12.9380 (0.1530)
 Sample #4 = 12.9280 (0.1680)
 Avg % AOS = 12.9293 (0.1433)
 STD DEV = 0.0520 (0.0321)
 REL STD DEV = 0.402 (22.416)

***** CHANNEL 2 *****
 Sol Val = 0.0088 mg/l or 0.000 g/210L
 % AOS = 0.105
 Std Dev = 0.01 Rel Std Dev = 9.76
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % AOS = 1.498
 Std Dev = 0.01 Rel Std Dev = 0.63
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % AOS = 3.622
 Std Dev = 0.01 Rel Std Dev = 0.29
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % AOS = 6.313
 Std Dev = 0.03 Rel Std Dev = 0.50
 Sol Val = 1.9048 mg/l or 0.400 g/210L
 % AOS = 12.929
 Std Dev = 0.05 Rel Std Dev = 0.42
 Zero Order Coef = -109.10
 First Order Coef = 1295.06
 Second Order Coef = 14.41
 Standard Deviation = 25.725883

*** Channel 2 ***
 Fit Value = 0.008 g/210L
 Samples Taken = 4, Discarded = 1
 Sum 10 = 12255, Sum 10 = 14221

***** CHANNEL 2 *****
 Sample % AOS (% AOS Ref)
 Sample #1 = 3.6320 (-0.0030)
 Sample #2 = 3.6310 (0.0070)
 Sample #3 = 3.6110 (0.0200)
 Sample #4 = 3.6250 (0.0180)
 Avg % AOS = 3.6223 (0.0150)
 STD DEV = 0.0103 (0.0070)
 REL STD DEV = 0.283 (46.667)

***** CHANNEL 2 *****
 Sample % AOS (% AOS Ref)
 Sample #1 = -0.0070 (0.0000)
 Sample #2 = -0.0040 (-0.0100)
 Sample #3 = -0.0020 (-0.0120)
 Sample #4 = 0.0070 (-0.0140)
 Avg % AOS = 0.0013 (-0.0143)
 STD DEV = 0.0059 (0.0055)
 REL STD DEV = 1757.842 (40.860)

***** CHANNEL 1 *****
 Sample % AOS (% AOS Ref)
 Sample #1 = 0.0010 (0.0000)
 Sample #2 = 0.0010 (0.0000)
 Sample #3 = 0.0010 (0.0000)
 Sample #4 = 0.0010 (0.0000)
 Avg % AOS = 0.0010 (0.0000)
 STD DEV = 0.0000 (0.0000)
 REL STD DEV = 0.0000 (0.0000)

***** CHANNEL 1 *****
 Sol Val = 0.0088 mg/l or 0.000 g/210L
 % AOS = 0.105
 Std Dev = 0.01 Rel Std Dev = 9.76
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % AOS = 1.498
 Std Dev = 0.01 Rel Std Dev = 0.63
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % AOS = 3.622
 Std Dev = 0.01 Rel Std Dev = 0.29
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % AOS = 6.313
 Std Dev = 0.03 Rel Std Dev = 0.50
 Sol Val = 1.9048 mg/l or 0.400 g/210L
 % AOS = 12.929
 Std Dev = 0.05 Rel Std Dev = 0.42
 Zero Order Coef = -109.10
 First Order Coef = 1295.06
 Second Order Coef = 14.41
 Standard Deviation = 25.725883

*** Channel 1 ***
 Fit Value = 0.008 g/210L
 Samples Taken = 4, Discarded = 1
 Sum 10 = 12255, Sum 10 = 14221

***** CHANNEL 1 *****
 Sample % AOS (% AOS Ref)
 Sample #1 = 3.5270 (-0.0090)
 Sample #2 = 3.5140 (0.0210)
 Sample #3 = 3.4820 (0.0450)
 Sample #4 = 3.4620 (0.0480)
 Avg % AOS = 3.4860 (0.0380)
 STD DEV = 0.0282 (0.0148)
 REL STD DEV = 0.752 (39.944)

***** CHANNEL 1 *****
 Sample % AOS (% AOS Ref)
 Sample #1 = 3.5270 (-0.0090)
 Sample #2 = 3.5140 (0.0210)
 Sample #3 = 3.4820 (0.0450)
 Sample #4 = 3.4620 (0.0480)
 Avg % AOS = 3.4860 (0.0380)
 STD DEV = 0.0282 (0.0148)
 REL STD DEV = 0.752 (39.944)

***** CHANNEL 1 *****
 Sample % AOS (% AOS Ref)
 Sample #1 = 3.5270 (-0.0090)
 Sample #2 = 3.5140 (0.0210)
 Sample #3 = 3.4820 (0.0450)
 Sample #4 = 3.4620 (0.0480)
 Avg % AOS = 3.4860 (0.0380)
 STD DEV = 0.0282 (0.0148)
 REL STD DEV = 0.752 (39.944)

***** CHANNEL 1 *****
 Sol Val = 0.0088 mg/l or 0.000 g/210L
 % AOS = 0.105
 Std Dev = 0.01 Rel Std Dev = 9.76
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % AOS = 1.498
 Std Dev = 0.01 Rel Std Dev = 0.63
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % AOS = 3.622
 Std Dev = 0.01 Rel Std Dev = 0.29
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % AOS = 6.313
 Std Dev = 0.03 Rel Std Dev = 0.50
 Sol Val = 1.9048 mg/l or 0.400 g/210L
 % AOS = 12.929
 Std Dev = 0.05 Rel Std Dev = 0.42
 Zero Order Coef = -109.10
 First Order Coef = 1295.06
 Second Order Coef = 14.41
 Standard Deviation = 25.725883

*** Channel 1 ***
 Fit Value = 0.008 g/210L
 Samples Taken = 4, Discarded = 1
 Sum 10 = 12255, Sum 10 = 14221

***** CHANNEL 1 *****
 Sample % AOS (% AOS Ref)
 Sample #1 = 0.7500 (-0.0080)
 Sample #2 = 0.7550 (0.0200)
 Sample #3 = 0.7740 (0.0170)
 Sample #4 = 0.7540 (0.0370)
 Avg % AOS = 0.7610 (0.0247)
 STD DEV = 0.0113 (0.0108)
 REL STD DEV = 1.481 (43.726)

***** CHANNEL 1 *****
 Sample % AOS (% AOS Ref)
 Sample #1 = 3.5270 (-0.0090)
 Sample #2 = 3.5140 (0.0210)
 Sample #3 = 3.4820 (0.0450)
 Sample #4 = 3.4620 (0.0480)
 Avg % AOS = 3.4860 (0.0380)
 STD DEV = 0.0282 (0.0148)
 REL STD DEV = 0.752 (39.944)

***** CHANNEL 1 *****
 Sample % AOS (% AOS Ref)
 Sample #1 = 3.5270 (-0.0090)
 Sample #2 = 3.5140 (0.0210)
 Sample #3 = 3.4820 (0.0450)
 Sample #4 = 3.4620 (0.0480)
 Avg % AOS = 3.4860 (0.0380)
 STD DEV = 0.0282 (0.0148)
 REL STD DEV = 0.752 (39.944)

***** CHANNEL 1 *****
 Sol Val = 0.0088 mg/l or 0.000 g/210L
 % AOS = 0.105
 Std Dev = 0.01 Rel Std Dev = 9.76
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % AOS = 1.498
 Std Dev = 0.01 Rel Std Dev = 0.63
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % AOS = 3.622
 Std Dev = 0.01 Rel Std Dev = 0.29
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % AOS = 6.313
 Std Dev = 0.03 Rel Std Dev = 0.50
 Sol Val = 1.9048 mg/l or 0.400 g/210L
 % AOS = 12.929
 Std Dev = 0.05 Rel Std Dev = 0.42
 Zero Order Coef = -109.10
 First Order Coef = 1295.06
 Second Order Coef = 14.41
 Standard Deviation = 25.725883

Optical Calibration
 SN: 80-000884
 Agency: Miami Dade P. D.
 Date: 02/09/2016
 Quadratic Fit: +/-0.002g/210L
 By: *[Signature]*

[Signature]

Sci Value = 0.090 g/210L ***
Fit Value = 0.3815 mg/l 222%
Samples Taken = 4, Discarded = 1


GGM

***** CHANNEL 1 *****
Sample #1 = 3218.00
Sample #2 = 3230.00
Sample #3 = 3243.00
Sample #4 = 3190.00
Average Result = 3221.0000
STD DEU = 27.6225
REL STD DEU = 0.856

***** CHANNEL 2 *****
Sample #1 = 3411.00
Sample #2 = 3429.00
Sample #3 = 3442.00
Sample #4 = 3437.00
Average Result = 3436.0000
STD DEU = 6.5574
REL STD DEU = 0.191

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1917
3 um H2O Adjust (mg/1*10,000) = 588 *
9 um H2O Adjust (mg/1*10,000) = 373
**** AUTO CAL PASS

AK

Optical Calibration Cont
SN: 80-000884
Agency: Miami Dade P.D.
Date: 02/09/2016
Quadratic Fit: +/-0.002g/210L
By: 

Simulator 400 was not sealed properly. Resealed and ran the test again.
--

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000884	Miami Dade Police Department	02/09/2016	<i>YALL</i>

YALL

15

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model: 8300 SN: 89-008884
02/09/2016
Software: 8100.27

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model: 8300 SN: 89-008884
02/09/2016
Software: 8100.27

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model: 8300 SN: 89-008884
02/09/2016
Software: 8100.27

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model: 8300 SN: 89-008884
02/09/2016
Software: 8100.27

Test 9:21:06 Time

Air Blank 0.000 09:18
Control Test 2.051 09:19
Air Blank 0.000 09:19
Control Test 0.050 09:20
Air Blank 0.000 09:20
Control Test 0.050 09:21
Air Blank 0.000 09:22
Control Test Stats
Average 0.0503
Std Dev 0.0005
Rel Std Dev(%) 1.1471

Test 9:21:06 Time

Air Blank 0.000 09:23
Control Test 0.081 09:24
Air Blank 0.000 09:24
Control Test 0.061 09:25
Air Blank 0.000 09:26
Control Test 0.081 09:26
Air Blank 0.000 09:27
Control Test Stats
Average 0.0610
Std Dev 0.0005
Rel Std Dev(%) 0.0000

Test 9:21:06 Time

Air Blank 0.000 09:28
Control Test 0.205 09:29
Air Blank 0.000 09:30
Control Test 0.206 09:30
Air Blank 0.000 09:31
Control Test 0.205 09:32
Air Blank 0.000 09:32
Control Test Stats
Average 0.2055
Std Dev 0.0005
Rel Std Dev(%) 0.2812

Test 9:21:06 Time

Air Blank 0.000 09:34
Control Test 0.082 09:34
Air Blank 0.000 09:35
Control Test 0.082 09:35
Air Blank 0.000 09:36
Control Test 0.082 09:36
Air Blank 0.000 09:36
Control Test Stats
Average 0.0820
Std Dev 0.0000
Rel Std Dev(%) 0.0000

YALL
Operator's Signature

YALL
Operator's Signature

YALL
Operator's Signature

YALL
Operator's Signature

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Flow calibration	80-000884	Miami Dade Police Department	2/9/2016	<i>[Signature]</i>

[Handwritten Signature]

MIAMI-DADE PD
 Intoxilizer - alcohol Analyzer
 Model 8000 SN 80-000884
 02/09/2016
 Software: 8100.27

Flow Rate Calibration*****
 1: Rate (liters/min) = 5
 SQR(DIFF) = 5.781
 2: Rate (liters/min) = 15
 SQR(DIFF) = 11.746
 3: Rate (liters/min) = 30
 SQR(DIFF) = 20.637
 Dependent Data Scale Factor = 10000 L/min
 Independent Data Scale Factor = 256
 Rounded Slope = 699
 Rounded Intercept = -572787
 Correlation = 0.99889

[Handwritten Mark]