

INSTRUMENT PROCESSING SHEET

Agency Chattahoochee PD S/N 80-000887
 Date In 6/6/16 Date Out 6/9/16 Ship P/U H/D CMI EE

Intake Performed By DR

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By PWS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 175
 Flow Verification (L/s)
 Flow Column # AP102
 32mm 144 (.139 - .169)
 36mm 1120 (.156 - .190)
 53mm 230 (.228 - .278)
 103mm 488 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G2403	2015034 7/14/17
0.08	SD3964	201601F 1/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AG1000504 1/5/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verifications (Alcohol Testing Program)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED

JUN 10 2016

FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By PWS

Barometric Pressure 1013 Gauge
 ID# 28427 1015 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2880
Interferent	G2834
0.05	G2403
0.08	SD3964
0.20	G4444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other Form 40

Notes: QA/QC OK DSGM

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
 Quality Control Review

6/10/16
 Date

Stability Tests - Chattahoochee 3D

#SU-000887

6/9/16

BK

CHATTAHOOCHEE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000887
06/09/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:14
Control Test	0.051	11:15
Air Blank	0.000	11:15
Control Test	0.049	11:16
Air Blank	0.000	11:17
Control Test	0.050	11:17
Air Blank	0.000	11:18
Control Test Stats		
Average	0.0500	
Std Dev	0.0010	
Rel Std Dev(%)	2.0000	

[Signature]

Operator's Signature

CHATTAHOOCHEE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000887
06/09/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:58
Control Test	0.079	10:58
Air Blank	0.000	10:59
Control Test	0.079	11:00
Air Blank	0.000	11:00
Control Test	0.079	11:01
Air Blank	0.000	11:01
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Signature]

Operator's Signature

CHATTAHOOCHEE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000887
06/09/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:03
Control Test	0.202	11:04
Air Blank	0.000	11:04
Control Test	0.202	11:05
Air Blank	0.000	11:06
Control Test	0.200	11:06
Air Blank	0.000	11:07
Control Test Stats		
Average	0.2013	
Std Dev	0.0012	
Rel Std Dev(%)	0.5735	

[Signature]

Operator's Signature

CHATTAHOOCHEE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000887
06/09/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:09
Control Test	0.079	11:09
Air Blank	0.000	11:10
Control Test	0.078	11:10
Air Blank	0.000	11:11
Control Test	0.079	11:11
Air Blank	0.000	11:11
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

[Signature]

Operator's Signature

[Handwritten mark]