

Agency Pinellas County SO SN 20-000889

JUL 11 2016

Date In 6/29/16 Date Out 7/8/16

Ship P/U H/D CMI EE FDLE

Intake Performed By DO

- Registration
- Annual
- Return from CMI
- Return from Enforcement Electronics
- Other _____

Visual Inspection:

OK Case OK Handle

OK Dry Gas Holder OK Feet

OK Keyboard/Plug OK Back/Plugs

OK Screws tight OK Breath Hose

- Other Equipment:
- Power cord
 - Printer Cable
 - Other _____

Notes: _____

Quality Checks Performed By DMB

- Breath Tube Screen
- Replace O-Rings
- Instrument Set Up Verified
- R-Value 187
- Flow Verification (L/s)

Flow Column # ATP.102

32mm 0.152 (.139 - .169)

36mm 0.167 (.156 - .190)

53mm 0.242 (.228 - .278)

103mm 0.515 (.447 - .547)

- Barometric Pressure Check
- Stability Checks

| Simulator | Serial # | Lot #/Exp |
|-----------|----------|---------------------|
| 0.05 | SD1018 | 201507A 7/14/17 |
| 0.08 | SD1011 | 201601F 1/26/19 |
| 0.20 | SD1025 | 201505A 5/12/17 |
| 0.08 DGS | N/A | AGW05301 2/22/18 |

Flow Calibration Performed By _____

- Flow Calibration N/A
 - Flow Calibration Complete
- Flow Column # _____
- 5L/min - 17mm
 - 15L/min - 53mm
 - 30L/min - 103mm
- R-Value _____
- Post Calibration Verification (L/s)
- Flow Column # _____
- 32mm _____ (.139 - .169)
- 36mm _____ (.156 - .190)
- 53mm _____ (.228 - .278)
- 103mm _____ (.447 - .547)

Maintenance Performed By _____

- Battery Replacement
- Dry Gas Regulator Replacement
- Breath Tube Replacement
- Other _____

Suggested Service

Optical Bench Calibration Performed By _____

- Optical Bench Calibration N/A
- Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000 | | N/A | N/A |
| 0.040 | | | |
| 0.100 | | | |
| 0.200 | | | |
| 0.400 | | | |
| 0.080 DGS | N/A | | |

- Post Calibration Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05 | | | |
| 0.08 | | | |
| 0.20 | | | |
| 0.08 DGS | N/A | | |

Notes: QA/QC OK GDM

Department Inspection Performed By DMB

- Barometric Pressure ID# 28427 1017 Gauge 1013 Instrument

Mouth Alcohol Solution Lot # 2015-A

Acetone Stock Solution Lot # 2016-B

| Simulator | Serial Number |
|-------------|---------------|
| 0.00 | SD1019 |
| Interferent | SD1021 |
| 0.05 | SD1018 |
| 0.08 | SD1011 |
| 0.20 | SD1025 |

Attachments

- Form 41
- Pre-Stability Tests
- Flow Calibration
- Optical Bench Cal
- Post-Stability Tests
- Other _____

- Instrument Complies with Chapter 11D-8, FAC
- Instrument Does Not Comply with Chapter 11D-8, FAC
- Return to/Place Into Evidentiary Use
- Remain Out of Evidentiary Use
- Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland

Quality Control Review

7/11/16

Date

Stability Checks 80-000889 Pinellas County SO. 7/6/16 *RMS* BK

RMS

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/06/2016
Software: 8100.27

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/06/2016
Software: 8100.27

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/06/2016
Software: 8100.27

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/06/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 16:21 |
| Control Test | 0.049 | 16:22 |
| Air Blank | 0.000 | 16:23 |
| Control Test | 0.049 | 16:23 |
| Air Blank | 0.000 | 16:24 |
| Control Test | 0.049 | 16:25 |
| Air Blank | 0.000 | 16:25 |
| Control Test Stats | | |
| Average | 0.0490 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

RMS

RMS
Operator's Signature

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 16:16 |
| Control Test | 0.079 | 16:17 |
| Air Blank | 0.000 | 16:18 |
| Control Test | 0.078 | 16:18 |
| Air Blank | 0.000 | 16:19 |
| Control Test | 0.079 | 16:20 |
| Air Blank | 0.000 | 16:20 |
| Control Test Stats | | |
| Average | 0.0787 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7339 | |

RMS
Operator's Signature

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 16:32 |
| Control Test | 0.202 | 16:33 |
| Air Blank | 0.000 | 16:33 |
| Control Test | 0.200 | 16:34 |
| Air Blank | 0.000 | 16:35 |
| Control Test | 0.199 | 16:35 |
| Air Blank | 0.000 | 16:36 |
| Control Test Stats | | |
| Average | 0.2003 | |
| Std Dev | 0.0015 | |
| Rel Std Dev(%) | 0.7625 | |

RMS
Operator's Signature

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 16:28 |
| Control Test | 0.079 | 16:28 |
| Air Blank | 0.000 | 16:29 |
| Control Test | 0.079 | 16:29 |
| Air Blank | 0.000 | 16:30 |
| Control Test | 0.079 | 16:30 |
| Air Blank | 0.000 | 16:30 |
| Control Test Stats | | |
| Average | 0.0790 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

RMS
Operator's Signature

INSTRUMENT PROCESSING SHEET

RECEIVED

Agency Pinellas County SO S/N 80-000889 JUL 1 1 2016
 Date In 4/29/16 Date Out 6/2/16 Ship P/U H/D CMI FDLE

Alcohol Testing Program

| Intake <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input checked="" type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____ | Quality Checks Performed By <u>RBS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>185</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>0.156</u> (.139 - .169) 36mm <u>0.167</u> (.156 - .190) 53mm <u>0.238</u> (.228 - .278) 103mm <u>0.511</u> (.447 - .547) <input type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>G2403</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD3964</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>G4444</td> <td>201503A 5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG60504 01/05/18</td> </tr> </tbody> </table> | Simulator | Serial # | Lot #/Exp | 0.05 | G2403 | 201507A 7/14/17 | 0.08 | SD3964 | 201601F 1/26/18 | 0.20 | G4444 | 201503A 5/12/17 | 0.08 DGS | N/A | AG60504 01/05/18 | Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) |
|--|---|--|----------|-----------|------|-------|--------------------|------|--------|--------------------|------|-------|--------------------|----------|-----|---------------------|--|
| Simulator | Serial # | Lot #/Exp | | | | | | | | | | | | | | | |
| 0.05 | G2403 | 201507A 7/14/17 | | | | | | | | | | | | | | | |
| 0.08 | SD3964 | 201601F 1/26/18 | | | | | | | | | | | | | | | |
| 0.20 | G4444 | 201503A 5/12/17 | | | | | | | | | | | | | | | |
| 0.08 DGS | N/A | AG60504 01/05/18 | | | | | | | | | | | | | | | |
| | | Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____ _____ | | | | | | | | | | | | | | | |

| Optical Bench Calibration Performed By _____ <input type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ | | | | | | | | | | | | | | | | | | | | | | | |
|--|---------------|------------|------------|-----------|---------------|------------|------------|------|--|--|--|------|--|--|--|------|--|--|--|----------|-----|--|--|
| Simulator | Serial Number | Lot Number | Expiration | | | | | | | | | | | | | | | | | | | | |
| 0.000 | | N/A | N/A | | | | | | | | | | | | | | | | | | | | |
| 0.040 | | | | | | | | | | | | | | | | | | | | | | | |
| 0.100 | | | | | | | | | | | | | | | | | | | | | | | |
| 0.200 | | | | | | | | | | | | | | | | | | | | | | | |
| 0.400 | | | | | | | | | | | | | | | | | | | | | | | |
| 0.080 DGS | N/A | | | | | | | | | | | | | | | | | | | | | | |
| <input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> | | | | Simulator | Serial Number | Lot Number | Expiration | 0.05 | | | | 0.08 | | | | 0.20 | | | | 0.08 DGS | N/A | | |
| Simulator | Serial Number | Lot Number | Expiration | | | | | | | | | | | | | | | | | | | | |
| 0.05 | | | | | | | | | | | | | | | | | | | | | | | |
| 0.08 | | | | | | | | | | | | | | | | | | | | | | | |
| 0.20 | | | | | | | | | | | | | | | | | | | | | | | |
| 0.08 DGS | N/A | | | | | | | | | | | | | | | | | | | | | | |

| Department Inspection Performed By _____ <input type="checkbox"/> Barometric Pressure Gauge ID# _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.05</td> <td></td> </tr> <tr> <td>0.08</td> <td></td> </tr> <tr> <td>0.20</td> <td></td> </tr> </tbody> </table> | | Simulator | Serial Number | 0.00 | | Interferent | | 0.05 | | 0.08 | | 0.20 | |
|---|---------------|-----------|---------------|------|--|-------------|--|------|--|------|--|------|--|
| Simulator | Serial Number | | | | | | | | | | | | |
| 0.00 | | | | | | | | | | | | | |
| Interferent | | | | | | | | | | | | | |
| 0.05 | | | | | | | | | | | | | |
| 0.08 | | | | | | | | | | | | | |
| 0.20 | | | | | | | | | | | | | |
| Attachments <input type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input checked="" type="checkbox"/> Other <u>Form 40</u> | | | | | | | | | | | | | |

Notes: Upon quality checks, determined instrument could be better calibrated - returned to Enforcement Electronics for calibration to bring values closer to nominal. @RBS

| | |
|---|----------------------------|
| <input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use | Compliance not determined. |
|---|----------------------------|

JUL 11 2016



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Pinellas County SO

S/N 80-000889

Date In 1/7/16

Date Out 1/26/16

Ship P/U H/D CMI FDLE Alcohol Testing Program

| Intake <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____ | Quality Checks Performed By <u>DUB</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>189</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>0.152</u> (.139 - .169) 36mm <u>0.160</u> (.156 - .190) 53mm <u>0.230</u> (.228 - .278) 103mm <u>0.507</u> (.447 - .547) <input type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1024</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>DR2035</td> <td>201502G 2/24/17</td> </tr> <tr> <td>0.20</td> <td>DR3856</td> <td>201505A 5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG511701 4/27/17</td> </tr> </tbody> </table> | Simulator | Serial # | Lot #/Exp | 0.05 | SD1024 | 201507A 7/14/17 | 0.08 | DR2035 | 201502G 2/24/17 | 0.20 | DR3856 | 201505A 5/12/17 | 0.08 DGS | N/A | AG511701 4/27/17 | Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) |
|---|--|---|----------|-----------|------|--------|--------------------|------|--------|--------------------|------|--------|--------------------|----------|-----|---------------------|--|
| Simulator | Serial # | Lot #/Exp | | | | | | | | | | | | | | | |
| 0.05 | SD1024 | 201507A 7/14/17 | | | | | | | | | | | | | | | |
| 0.08 | DR2035 | 201502G 2/24/17 | | | | | | | | | | | | | | | |
| 0.20 | DR3856 | 201505A 5/12/17 | | | | | | | | | | | | | | | |
| 0.08 DGS | N/A | AG511701 4/27/17 | | | | | | | | | | | | | | | |
| | | Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____ | | | | | | | | | | | | | | | |

| Optical Bench Calibration #1 Performed By <u>SOMB</u> <input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1004</u> ID# <u>28427</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>DR1075</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>G2882</td> <td>13108</td> <td>8/18/17</td> </tr> <tr> <td>0.100</td> <td>G2078</td> <td>15001</td> <td>5/20/17</td> </tr> <tr> <td>0.200</td> <td>G2408</td> <td>14104</td> <td>6/25/16</td> </tr> <tr> <td>0.400</td> <td>G5358</td> <td>15105</td> <td>6/10/17</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>09014080A1</td> <td>5/1/16</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1024</td> <td>201507A</td> <td>7/14/17</td> </tr> <tr> <td>0.08</td> <td>DR2035</td> <td>201502G</td> <td>2/24/17</td> </tr> <tr> <td>0.20</td> <td>DR3856</td> <td>201505A</td> <td>5/12/17</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG511701</td> <td>4/27/17</td> </tr> </tbody> </table> | Simulator | Serial Number | Lot Number | Expiration | 0.000 | DR1075 | N/A | N/A | 0.040 | G2882 | 13108 | 8/18/17 | 0.100 | G2078 | 15001 | 5/20/17 | 0.200 | G2408 | 14104 | 6/25/16 | 0.400 | G5358 | 15105 | 6/10/17 | 0.080 DGS | N/A | 09014080A1 | 5/1/16 | Simulator | Serial Number | Lot Number | Expiration | 0.05 | SD1024 | 201507A | 7/14/17 | 0.08 | DR2035 | 201502G | 2/24/17 | 0.20 | DR3856 | 201505A | 5/12/17 | 0.08 DGS | N/A | AG511701 | 4/27/17 |
|---|---------------|---------------|------------|------------|-------|--------|-----|-----|-------|-------|-------|---------|-------|-------|-------|---------|-------|-------|-------|---------|-------|-------|-------|---------|-----------|-----|------------|--------|-----------|---------------|------------|------------|------|--------|---------|---------|------|--------|---------|---------|------|--------|---------|---------|----------|-----|----------|---------|
| Simulator | Serial Number | Lot Number | Expiration | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.000 | DR1075 | N/A | N/A | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.040 | G2882 | 13108 | 8/18/17 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.100 | G2078 | 15001 | 5/20/17 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.200 | G2408 | 14104 | 6/25/16 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.400 | G5358 | 15105 | 6/10/17 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.080 DGS | N/A | 09014080A1 | 5/1/16 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Simulator | Serial Number | Lot Number | Expiration | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.05 | SD1024 | 201507A | 7/14/17 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.08 | DR2035 | 201502G | 2/24/17 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.20 | DR3856 | 201505A | 5/12/17 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 0.08 DGS | N/A | AG511701 | 4/27/17 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |

| Department Inspection Performed By _____ <input type="checkbox"/> Barometric Pressure _____ Gauge ID# _____ Instrument Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot.# _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.05</td> <td></td> </tr> <tr> <td>0.08</td> <td></td> </tr> <tr> <td>0.20</td> <td></td> </tr> </tbody> </table> | Simulator | Serial Number | 0.00 | | Interferent | | 0.05 | | 0.08 | | 0.20 | |
|---|---------------|---------------|------|--|-------------|--|------|--|------|--|------|--|
| Simulator | Serial Number | | | | | | | | | | | |
| 0.00 | | | | | | | | | | | | |
| Interferent | | | | | | | | | | | | |
| 0.05 | | | | | | | | | | | | |
| 0.08 | | | | | | | | | | | | |
| 0.20 | | | | | | | | | | | | |
| Attachments <input type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Optical Bench Cal <u>x2</u> <input checked="" type="checkbox"/> Post-Stability Tests <u>x2</u> <input type="checkbox"/> Other _____ | | | | | | | | | | | | |

Notes: Calibrated twice to bring values closer to nominal. While values are within range - suggest sending to repair for further evaluation.
QC-TSK

| | |
|---|----------------------------------|
| <input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use | <i>Compliance not determined</i> |
|---|----------------------------------|

Quality Control Review

Date



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

RECEIVED

Agency Pinellas County SO

S/N 80-000889

JUL 11 2016

Date In _____ Date Out _____

Ship P/U H/D CM FDLE

Alcohol Testing Program

Intake Performed By _____

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 _____ Case _____ Handle
 _____ Dry Gas Holder _____ Feet
 _____ Keyboard/Plug _____ Back/Plugs
 _____ Screws tight _____ Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By _____

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value _____
 Flow Verification (L/s)
 Flow Column # _____

| | | |
|-------|-------|---------------|
| 32mm | _____ | (.139 - .169) |
| 36mm | _____ | (.156 - .190) |
| 53mm | _____ | (.228 - .278) |
| 103mm | _____ | (.447 - .547) |

Barometric Pressure Check
 Gauge ID # _____

Stability Checks

| Simulator | Serial # | Lot #/Exp |
|-----------|----------|-----------|
| 0.05 | | |
| 0.08 | | |
| 0.20 | | |
| 0.08 DGS | N/A | |

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____

5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____

Post Calibration Verification (L/s)
 Flow Column # _____

| | | |
|-------|-------|---------------|
| 32mm | _____ | (.139 - .169) |
| 36mm | _____ | (.156 - .190) |
| 53mm | _____ | (.228 - .278) |
| 103mm | _____ | (.447 - .547) |

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration #2 Performed By DMB

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge 1030 ID# 26930

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000 | DR1275 | N/A | N/A |
| 0.040 | SD3962 | 15108 | 8/18/17 |
| 0.100 | G2078 | 15001 | 5/20/17 |
| 0.200 | G2408 | 4104 | 6/25/16 |
| 0.400 | SD3964 | 15105 | 6/10/17 |
| 0.080 DGS | N/A | 09014080A1 | 5/1/16 |

Post Calibration Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05 | SD1018 | 201507A | 7/14/17 |
| 0.08 | SD1011 | 201502G | 2/24/17 |
| 0.20 | SD1025 | 201505A | 5/12/17 |
| 0.08 DGS | N/A | AG511701 | 4/27/17 |

Department Inspection Performed By _____

Barometric Pressure _____ Gauge
 ID# _____ Instrument

Mouth Alcohol Solution Lot # _____
 Acetone Stock Solution Lot # _____

| Simulator | Serial Number |
|-------------|---------------|
| 0.00 | |
| Interferent | |
| 0.05 | |
| 0.08 | |
| 0.20 | |

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: _____

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

Pre-Cal
Stability Checks

80-000889 Pinellas County S.O. 1/20/16 *PMB*

80

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000889
01/20/2016
Software: 8100.27

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000889
01/20/2016
Software: 8100.27

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000889
01/20/2016
Software: 8100.27

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000889
01/20/2016
Software: 8100.27

| Test | 9/21/0L | Time |
|--------------------|---------|-------|
| Air Blank | 0.000 | 10:41 |
| Control Test | 0.053 | 10:41 |
| Air Blank | 0.000 | 10:42 |
| Control Test | 0.051 | 10:43 |
| Air Blank | 0.000 | 10:43 |
| Control Test | 0.052 | 10:44 |
| Air Blank | 0.000 | 10:45 |
| Control Test Stats | | |
| Average | 0.0520 | |
| Std Dev | 0.0010 | |
| Rel Std Dev(%) | 1.9231 | |

PMB

Operator's Signature

| Test | 9/21/0L | Time |
|--------------------|---------|-------|
| Air Blank | 0.000 | 10:46 |
| Control Test | 0.083 | 10:46 |
| Air Blank | 0.000 | 10:47 |
| Control Test | 0.082 | 10:48 |
| Air Blank | 0.000 | 10:48 |
| Control Test | 0.081 | 10:49 |
| Air Blank | 0.000 | 10:50 |
| Control Test Stats | | |
| Average | 0.0820 | |
| Std Dev | 0.0010 | |
| Rel Std Dev(%) | 1.2195 | |

PMB

Operator's Signature

| Test | 9/21/0L | Time |
|--------------------|---------|-------|
| Air Blank | 0.000 | 10:52 |
| Control Test | 0.207 | 10:53 |
| Air Blank | 0.000 | 10:54 |
| Control Test | 0.205 | 10:54 |
| Air Blank | 0.000 | 10:55 |
| Control Test | 0.205 | 10:56 |
| Air Blank | 0.000 | 10:56 |
| Control Test Stats | | |
| Average | 0.2057 | |
| Std Dev | 0.0012 | |
| Rel Std Dev(%) | 0.5614 | |

PMB

Operator's Signature

| Test | 9/21/0L | Time |
|--------------------|---------|-------|
| Air Blank | 0.000 | 10:59 |
| Control Test | 0.080 | 10:59 |
| Air Blank | 0.000 | 11:00 |
| Control Test | 0.081 | 11:00 |
| Air Blank | 0.000 | 11:00 |
| Control Test | 0.080 | 11:01 |
| Air Blank | 0.000 | 11:01 |
| Control Test Stats | | |
| Average | 0.0803 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7187 | |

PMB

Operator's Signature

Pinellas County So. Intox. User: SN 80-000889 12:27:42

Channel 2 % Abs % Abs Ref

Sample #1 = 1.522 (-0.0160)
Sample #2 = 1.486 (-0.0200)
Sample #3 = 1.453 (-0.0460)
Sample #4 = 1.464 (-0.0490)
Aug % Abs = 1.4677 (0.0383)
STD DEV = 0.0168 (0.0159)
REL STD DEV = 1.145 (41.603)

*Auto Cal Data Channel 1
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.112
Std Dev = 0.01 Rel Std Dev = 8.22
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.817
Std Dev = 0.02 Rel Std Dev = 2.65
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.908
Std Dev = 0.01 Rel Std Dev = 0.27
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.589
Std Dev = 0.01 Rel Std Dev = 0.36
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 6.879
Std Dev = 0.02 Rel Std Dev = 0.35
Zero Order Coef = -296.81
First Order Coef = 2631.63
Second Order Coef = 26.30
Standard Deviation = 38.949104

Table with columns: Act, g/210L, Fit, Residual, g/210L. Values range from 0.000 to 0.400.

Auto Calibration Max Power Res Value = 60 Auto Range Res Value = 47
Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12296, Sum Io = 14287
Channel 1 >>>>
Sample % Abs (% Abs Ref) (-0.0100)
Sample #1 = 0.1560 (0.0390)
Sample #2 = 0.1070 (0.0850)
Sample #3 = 0.1230 (0.0880)
Sample #4 = 0.1070 (0.0707)
Aug % Abs = 0.1123 (0.0707)
STD DEV = 0.0092 (0.0275)
REL STD DEV = 8.223 (38.866)

Channel 2 % Abs (% Abs Ref) (-0.0220)
Sample #1 = 1.9770 (0.0580)
Sample #2 = 1.9070 (0.0820)
Sample #3 = 1.9140 (0.0840)
Sample #4 = 1.9040 (0.0747)
Aug % Abs = 1.9083 (0.0145)
STD DEV = 0.0151 (0.0145)
REL STD DEV = 0.269 (19.377)

Channel 2 % Abs (% Abs Ref) (-0.0238)
Sample #1 = 7.0920 (0.1670)
Sample #2 = 6.9070 (0.2620)
Sample #3 = 6.8670 (0.2740)
Sample #4 = 6.8630 (0.2343)
Aug % Abs = 6.8790 (0.0586)
STD DEV = 0.0243 (0.0566)
REL STD DEV = 0.354 (25.016)

Channel 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.092
Std Dev = 0.01 Rel Std Dev = 8.33
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.468
Std Dev = 0.02 Rel Std Dev = 1.14
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.539
Std Dev = 0.01 Rel Std Dev = 0.21
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.726
Std Dev = 0.02 Rel Std Dev = 0.30
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 12.725
Std Dev = 0.04 Rel Std Dev = 0.33
Zero Order Coef = -120.59
First Order Coef = 1345.63
Second Order Coef = 12.65
Standard Deviation = 25.113581

Table with columns: Act, g/210L, Fit, Residual, g/210L. Values range from 0.000 to 0.400.

Channel 2 >>>>
Sample % Abs (% Abs Ref) (-0.0140)
Sample #1 = 0.0940 (0.0140)
Sample #2 = 0.0900 (0.0440)
Sample #3 = 0.0850 (0.0390)
Sample #4 = 0.1000 (0.0323)
Aug % Abs = 0.0917 (0.0161)
STD DEV = 0.0076 (0.0161)
REL STD DEV = 8.332 (49.710)

Channel 2 >>>>
Sample % Abs (% Abs Ref) (-0.0160)
Sample #1 = 3.6660 (0.0980)
Sample #2 = 3.5310 (0.1220)
Sample #3 = 3.5420 (0.1280)
Sample #4 = 3.5393 (0.1161)
Aug % Abs = 3.5393 (0.0159)
STD DEV = 0.0074 (0.0159)
REL STD DEV = 0.208 (13.685)

Channel 2 >>>>
Sample % Abs (% Abs Ref) (-0.0160)
Sample #1 = 13.1030 (0.3350)
Sample #2 = 12.7730 (0.4570)
Sample #3 = 12.7020 (0.4630)
Sample #4 = 12.6990 (0.4250)
Aug % Abs = 12.7247 (0.0750)
STD DEV = 0.0419 (0.0750)
REL STD DEV = 0.329 (18.593)

Channel 2 >>>>
Sol Value = 0.000 g/210L ***
Fit Value = 0.3810 mg/l 2222
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 3041.00
Sample #2 = 2984.00
Sample #3 = 3003.00
Sample #4 = 2966.00
Average Result = 2984.3333
STD DEV = 18.5023
REL STD DEV = 0.620
***** CHANNEL 2
Sample #1 = 3434.00
Sample #2 = 3470.00
Sample #3 = 3491.00
Sample #4 = 3455.00
Average Result = 3472.0000
STD DEV = 18.0831
REL STD DEV = 0.521

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1004
3 um H2O Adjust (mg/l*10,000) = 825
9 um H2O Adjust (mg/l*10,000) = 337
**** AUTO CAL PASS

Table with columns: Act, g/210L, Fit, Residual, g/210L. Values range from 0.000 to 0.400.

Channel 2 >>>>
Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12279, Sum Io = 14274
Channel 1 >>>>
Sample % Abs (% Abs Ref) (-0.0140)
Sample #1 = 0.8690 (0.0150)
Sample #2 = 0.8400 (0.0430)
Sample #3 = 0.7970 (0.0570)
Sample #4 = 0.8140 (0.0383)
Aug % Abs = 0.8170 (0.0214)
STD DEV = 0.0217 (0.0214)
REL STD DEV = 2.651 (55.788)

Channel 1 >>>>
Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12265, Sum Io = 14266
Channel 1 >>>>
Sample % Abs (% Abs Ref) (-0.0090)
Sample #1 = 3.6500 (0.0910)
Sample #2 = 3.6020 (0.1290)
Sample #3 = 3.5680 (0.1410)
Sample #4 = 3.5760 (0.1207)
Aug % Abs = 3.5887 (0.0207)
STD DEV = 0.0730 (0.0207)
REL STD DEV = 0.363 (21.092)

Channel 1 >>>>
Sol Value = 0.400 g/210L ***
Fit Value = 1.9048 mg/l 2222
Samples Taken = 4, Discarded = 1
Sum Io = 12259, Sum Io = 14258
Channel 1 >>>>
Sample % Abs (% Abs Ref) (-0.0238)
Sample #1 = 7.0920 (0.1670)
Sample #2 = 6.9070 (0.2620)
Sample #3 = 6.8670 (0.2740)
Sample #4 = 6.8630 (0.2343)
Aug % Abs = 6.8790 (0.0586)
STD DEV = 0.0243 (0.0566)
REL STD DEV = 0.354 (25.016)

Channel 1 >>>>
Sol Value = 0.800 g/210L ***
Fit Value = 0.3810 mg/l 2222
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 3041.00
Sample #2 = 2984.00
Sample #3 = 3003.00
Sample #4 = 2966.00
Average Result = 2984.3333
STD DEV = 18.5023
REL STD DEV = 0.620
***** CHANNEL 2
Sample #1 = 3434.00
Sample #2 = 3470.00
Sample #3 = 3491.00
Sample #4 = 3455.00
Average Result = 3472.0000
STD DEV = 18.0831
REL STD DEV = 0.521

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1004
3 um H2O Adjust (mg/l*10,000) = 825
9 um H2O Adjust (mg/l*10,000) = 337
**** AUTO CAL PASS

Table with columns: Act, g/210L, Fit, Residual, g/210L. Values range from 0.000 to 0.400.

EMS

Post Card

Stability Checks 80-00889 Pinellas County SO. 1/22/16 RMB

GB

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-00889
01/22/2016
Software: 8100.27

| Test | g/210L | Time |
|----------------|--------|-------|
| Air Blank | 0.000 | 14:54 |
| Control Test | 0.052 | 14:54 |
| Air Blank | 0.000 | 14:55 |
| Control Test | 0.052 | 14:56 |
| Air Blank | 0.000 | 14:57 |
| Control Test | 0.052 | 14:58 |
| Air Blank | 0.000 | 14:59 |
| Control Test | 0.052 | 14:59 |
| Average | 0.0520 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

RMB

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-00889
01/22/2016
Software: 8100.27

| Test | g/210L | Time |
|----------------|--------|-------|
| Air Blank | 0.000 | 14:59 |
| Control Test | 0.082 | 15:00 |
| Air Blank | 0.000 | 15:00 |
| Control Test | 0.082 | 15:01 |
| Air Blank | 0.000 | 15:02 |
| Control Test | 0.082 | 15:02 |
| Air Blank | 0.000 | 15:03 |
| Control Test | 0.082 | 15:03 |
| Average | 0.0820 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

RMB

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-00889
01/22/2016
Software: 8100.27

| Test | g/210L | Time |
|----------------|--------|-------|
| Air Blank | 0.000 | 15:05 |
| Control Test | 0.210 | 15:05 |
| Air Blank | 0.000 | 15:06 |
| Control Test | 0.208 | 15:07 |
| Air Blank | 0.000 | 15:07 |
| Control Test | 0.207 | 15:08 |
| Air Blank | 0.000 | 15:09 |
| Control Test | 0.000 | 15:09 |
| Average | 0.2083 | |
| Std Dev | 0.0015 | |
| Rel Std Dev(%) | 0.7332 | |

RMB

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-00889
01/22/2016
Software: 8100.27

| Test | g/210L | Time |
|----------------|--------|-------|
| Air Blank | 0.000 | 15:11 |
| Control Test | 0.081 | 15:11 |
| Air Blank | 0.000 | 15:12 |
| Control Test | 0.081 | 15:12 |
| Air Blank | 0.000 | 15:13 |
| Control Test | 0.081 | 15:13 |
| Air Blank | 0.000 | 15:14 |
| Control Test | 0.000 | 15:14 |
| Average | 0.0810 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

RMB

Operator's Signature

Calibration #2 Data 80-000889 Pinellas County S.O. 1/26/16 OMS

PINELAS COUNTY, FL
Integrator Model: 80-000889
Date: 12-26-2016
Auto Calibration
Max Power Res Value = 60
Auto Range Res Value = 46

Sample % Abs (% Abs Ref)
Sample #1 = 1.5490 (-0.0180)
Sample #2 = 1.5100 (0.0100)
Sample #3 = 1.5430 (-0.0090)
Sample #4 = 1.5250 (-0.0120)
Avg % Abs = 1.5260 (-0.0077)
STD DEV = 0.0165 (0.0119)
REL STD DEV = 1.083 (325.373)

Soil Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Sum Io = 12152, Sum Io = 14205
Samples Taken = 4, Discarded = 1
Sum Io = 12152, Sum Io = 14205
Sample % Abs (% Abs Ref)
Sample #1 = 1.9960 (-0.0250)
Sample #2 = 1.9570 (-0.0340)
Sample #3 = 1.9250 (-0.0490)
Sample #4 = 1.9160 (-0.0640)
Avg % Abs = 1.9327 (-0.0490)
STD DEV = 0.0215 (0.0150)
REL STD DEV = 1.115 (330.612)

Sample % Abs (% Abs Ref)
Sample #1 = 0.0530 (-0.0160)
Sample #2 = 0.1140 (-0.0230)
Sample #3 = 0.1120 (-0.0280)
Sample #4 = 0.0900 (-0.0200)
Avg % Abs = 0.1053 (-0.0237)
STD DEV = 0.0133 (0.0040)
REL STD DEV = 12.642 (17.077)

Soil Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Sum Io = 12155, Sum Io = 14202
Samples Taken = 4, Discarded = 1
Sum Io = 12155, Sum Io = 14202
Sample % Abs (% Abs Ref)
Sample #1 = 0.8570 (-0.0160)
Sample #2 = 0.8560 (-0.0090)
Sample #3 = 0.8680 (-0.0050)
Sample #4 = 0.8500 (-0.0030)
Avg % Abs = 0.8580 (-0.0057)
STD DEV = 0.0092 (0.0031)
REL STD DEV = 1.368 (53.913)

Sample % Abs (% Abs Ref)
Sample #1 = 7.0010 (0.0060)
Sample #2 = 6.7990 (0.2080)
Sample #3 = 6.7770 (0.2460)
Sample #4 = 6.7790 (0.2420)
Avg % Abs = 6.7850 (0.2320)
STD DEV = 0.0122 (0.0209)
REL STD DEV = 0.179 (9.000)

Soil Value = 0.400 g/210L ***
Fit Value = 1.9048 mg/l %%%
Sum Io = 12143, Sum Io = 14190
Samples Taken = 4, Discarded = 1
Sum Io = 12143, Sum Io = 14190
Sample % Abs (% Abs Ref)
Sample #1 = 7.1450 (-0.0220)
Sample #2 = 6.9370 (0.2110)
Sample #3 = 7.0420 (0.1470)
Sample #4 = 6.9860 (0.2030)
Avg % Abs = 6.9883 (0.1870)
STD DEV = 0.0525 (0.0249)
REL STD DEV = 0.752 (18.648)

Sample % Abs (% Abs Ref)
Sample #1 = 13.2820 (-0.0180)
Sample #2 = 12.8700 (0.4060)
Sample #3 = 13.0150 (0.2830)
Sample #4 = 12.8950 (0.3910)
Avg % Abs = 12.9267 (0.3600)
STD DEV = 0.0775 (0.0671)
REL STD DEV = 0.500 (18.640)

Soil Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Sum Io = 12150, Sum Io = 14199
Samples Taken = 4, Discarded = 1
Sum Io = 12150, Sum Io = 14199
Sample % Abs (% Abs Ref)
Sample #1 = 3.7510 (-0.0070)
Sample #2 = 3.6300 (0.1050)
Sample #3 = 3.6190 (0.1340)
Sample #4 = 3.6230 (0.1380)
Avg % Abs = 3.6240 (0.1260)
STD DEV = 0.0056 (0.0174)
REL STD DEV = 0.154 (13.838)

Soil Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.136
Std Dev = 0.01 Rel Std Dev = 5.73
Soil Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.858
Std Dev = 0.01 Rel Std Dev = 1.07
Soil Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.933
Std Dev = 0.02 Rel Std Dev = 1.11
Soil Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.624
Std Dev = 0.01 Rel Std Dev = 0.15
Soil Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 6.988
Std Dev = 0.05 Rel Std Dev = 0.75
Zero Order Coef = -383.68
First Order Coef = 2656.73
Second Order Coef = 17.87
Standard Deviation = 37.794987

Soil Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.105
Std Dev = 0.01 Rel Std Dev = 12.64
Soil Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.526
Std Dev = 0.02 Rel Std Dev = 1.08
Soil Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.571
Std Dev = 0.01 Rel Std Dev = 0.41
Soil Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.785
Std Dev = 0.01 Rel Std Dev = 0.18
Soil Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 12.927
Std Dev = 0.08 Rel Std Dev = 0.60
Zero Order Coef = -166.74
First Order Coef = 1352.76
Second Order Coef = 10.37
Standard Deviation = 28.749342

Soil Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 2936.00
Sample #2 = 2972.00
Sample #3 = 3071.00
Sample #4 = 3034.00
Average Result = 3025.6667
STD DEV = 50.0233
REL STD DEV = 1.653
***** CHANNEL 2 *****
Sample #1 = 3425.00
Sample #2 = 3472.00
Sample #3 = 3521.00
Sample #4 = 3505.00
Average Result = 3499.3333
STD DEV = 24.9867
REL STD DEV = 0.714
***** CHANNEL 1 *****
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1018
3 um H2O Adjust (mg/l*10,000) = 784
9 um H2O Adjust (mg/l*10,000) = 310
***** AUTO CAL PASS *****

Soil Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 0.8570 (-0.0160)
Sample #2 = 0.8560 (-0.0090)
Sample #3 = 0.8680 (-0.0050)
Sample #4 = 0.8500 (-0.0030)
Avg % Abs = 0.8580 (-0.0057)
STD DEV = 0.0092 (0.0031)
REL STD DEV = 1.368 (53.913)

Solution State Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0005
0.040 0.040 -0.0001
0.100 0.101 -0.0012
0.200 0.199 0.0009
0.400 0.400 -0.0002

Solution State Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.001 0.0005
0.040 0.040 -0.0004
0.100 0.101 -0.0007
0.200 0.199 0.0007
0.400 0.400 -0.0001

Soil Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 2936.00
Sample #2 = 2972.00
Sample #3 = 3071.00
Sample #4 = 3034.00
Average Result = 3025.6667
STD DEV = 50.0233
REL STD DEV = 1.653
***** CHANNEL 2 *****
Sample #1 = 3425.00
Sample #2 = 3472.00
Sample #3 = 3521.00
Sample #4 = 3505.00
Average Result = 3499.3333
STD DEV = 24.9867
REL STD DEV = 0.714
***** CHANNEL 1 *****
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1018
3 um H2O Adjust (mg/l*10,000) = 784
9 um H2O Adjust (mg/l*10,000) = 310
***** AUTO CAL PASS *****

Soil Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 0.8570 (-0.0160)
Sample #2 = 0.8560 (-0.0090)
Sample #3 = 0.8680 (-0.0050)
Sample #4 = 0.8500 (-0.0030)
Avg % Abs = 0.8580 (-0.0057)
STD DEV = 0.0092 (0.0031)
REL STD DEV = 1.368 (53.913)

Post Cal #2

Stability Checks 80-000889 Pinellas County SO 1/26/16 DUB

DUB

PINELLAS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-010889
 01/26/2016
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 14:39 |
| Control Test | 0.052 | 14:40 |
| Air Blank | 0.000 | 14:41 |
| Control Test | 0.050 | 14:41 |
| Air Blank | 0.000 | 14:42 |
| Control Test | 0.052 | 14:43 |
| Air Blank | 0.000 | 14:43 |
| Control Test Stats | | |
| Average | 0.0513 | |
| Std Dev | 0.0012 | |
| Rel Std Dev(%) | 2.2494 | |

DUB

Operator's Signature

PINELLAS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-010889
 01/26/2016
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 14:44 |
| Control Test | 0.002 | 14:45 |
| Air Blank | 0.000 | 14:46 |
| Control Test | 0.000 | 14:46 |
| Air Blank | 0.000 | 14:47 |
| Control Test | 0.001 | 14:48 |
| Air Blank | 0.000 | 14:48 |
| Control Test Stats | | |
| Average | 0.0010 | |
| Std Dev | 0.0010 | |
| Rel Std Dev(%) | 1.2346 | |

DUB

Operator's Signature

PINELLAS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-010889
 01/26/2016
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 14:50 |
| Control Test | 0.209 | 14:50 |
| Air Blank | 0.000 | 14:51 |
| Control Test | 0.207 | 14:52 |
| Air Blank | 0.000 | 14:52 |
| Control Test | 0.206 | 14:53 |
| Air Blank | 0.000 | 14:54 |
| Control Test Stats | | |
| Average | 0.2073 | |
| Std Dev | 0.0015 | |
| Rel Std Dev(%) | 0.7368 | |

DUB

Operator's Signature

PINELLAS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-010889
 01/26/2016
 Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 14:55 |
| Control Test | 0.078 | 14:55 |
| Air Blank | 0.000 | 14:55 |
| Control Test | 0.079 | 14:56 |
| Air Blank | 0.000 | 14:56 |
| Control Test | 0.079 | 14:57 |
| Air Blank | 0.000 | 14:57 |
| Control Test Stats | | |
| Average | 0.0787 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7339 | |

DUB

Operator's Signature

Stability Checks
1/26/16
80-00839
Pinellas County S.O.

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000889
01/26/2016
Software: 8100.27

| Test | °g/210L | Time |
|--------------------|---------|-------|
| Air Blank | 0.000 | 15:15 |
| Control Test | 0.208 | 15:15 |
| Air Blank | 0.000 | 15:16 |
| Control Test | 0.207 | 15:17 |
| Air Blank | 0.000 | 15:17 |
| Control Test | 0.207 | 15:18 |
| Air Blank | 0.000 | 15:19 |
| Control Test | 0.207 | 15:19 |
| Air Blank | 0.000 | 15:20 |
| Control Test | 0.207 | 15:21 |
| Air Blank | 0.000 | 15:21 |
| Control Test | 0.209 | 15:22 |
| Air Blank | 0.000 | 15:23 |
| Control Test | 0.206 | 15:23 |
| Air Blank | 0.000 | 15:24 |
| Control Test | 0.208 | 15:25 |
| Air Blank | 0.000 | 15:25 |
| Control Test | 0.207 | 15:26 |
| Air Blank | 0.000 | 15:27 |
| Control Test | 0.207 | 15:27 |
| Air Blank | 0.000 | 15:28 |
| Control Test | 0.207 | 15:29 |
| Air Blank | 0.000 | 15:29 |
| Control Test | 0.207 | 15:30 |
| Air Blank | 0.000 | 15:31 |
| Control Test | 0.206 | 15:32 |
| Air Blank | 0.000 | 15:32 |
| Control Test | 0.207 | 15:33 |
| Air Blank | 0.000 | 15:34 |
| Control Test | 0.207 | 15:34 |
| Air Blank | 0.000 | 15:35 |
| Control Test | 0.207 | 15:36 |
| Air Blank | 0.000 | 15:36 |
| Control Test | 0.207 | 15:37 |
| Air Blank | 0.000 | 15:38 |
| Control Test | 0.207 | 15:38 |
| Air Blank | 0.000 | 15:39 |
| Control Test | 0.207 | 15:40 |
| Air Blank | 0.000 | 15:40 |
| Control Test | 0.207 | 15:41 |
| Air Blank | 0.000 | 15:42 |
| Control Test Stats | | |
| Average | 0.2071 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.3094 | |



Operator's Signature