



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FWC

S/N 80-000891

Date In 10/27/16

Date Out 11/3/16

Ship P/U H/D CMI EE

Intake Performed By <u>DR</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Dry Gas (Wicklin)</u> Notes: _____ _____ _____	Quality Checks Performed By _____ <input type="checkbox"/> Breath Tube Screen <input type="checkbox"/> Replace O-Rings <input type="checkbox"/> Instrument Set Up Verified <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Flow Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) <input type="checkbox"/> Barometric Pressure Check Gauge ID # _____ <input type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05			0.08			0.20			0.08 DGS	N/A		Flow Calibration Performed By _____ <input type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 30mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (Program) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05																	
0.08																	
0.20																	
0.08 DGS	N/A																
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____															

RECEIVED
NOV 04 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By <u>SPM</u> <input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1023</u> ID# <u>26932</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G4444</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD1024</td> <td>16101</td> <td>2/2/18</td> </tr> <tr> <td>0.100</td> <td>SD1013</td> <td>16001</td> <td>5/8/18</td> </tr> <tr> <td>0.200</td> <td>SD1012</td> <td>16103</td> <td>6/14/18</td> </tr> <tr> <td>0.400</td> <td>G6621</td> <td>16102</td> <td>3/22/18</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>03415080A1</td> <td>3/5/17</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A</td> <td>7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F</td> <td>1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C</td> <td>4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG619605</td> <td>7/14/18</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	G4444	N/A	N/A	0.040	SD1024	16101	2/2/18	0.100	SD1013	16001	5/8/18	0.200	SD1012	16103	6/14/18	0.400	G6621	16102	3/22/18	0.080 DGS	N/A	03415080A1	3/5/17	Simulator	Serial Number	Lot Number	Expiration	0.05	SD1018	201507A	7/14/17	0.08	SD1011	201601F	1/26/18	0.20	SD1025	201604C	4/5/18	0.08 DGS	N/A	AG619605	7/14/18	Department Inspection Performed By <u>SPM</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1024</u> Gauge ID# <u>28427</u> <u>1024</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Post-Stability Tests <input checked="" type="checkbox"/> Other <u>EXPERIMENTAL BREATH TESTS</u>	Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	G4444	N/A	N/A																																																										
0.040	SD1024	16101	2/2/18																																																										
0.100	SD1013	16001	5/8/18																																																										
0.200	SD1012	16103	6/14/18																																																										
0.400	G6621	16102	3/22/18																																																										
0.080 DGS	N/A	03415080A1	3/5/17																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05	SD1018	201507A	7/14/17																																																										
0.08	SD1011	201601F	1/26/18																																																										
0.20	SD1025	201604C	4/5/18																																																										
0.08 DGS	N/A	AG619605	7/14/18																																																										
Simulator	Serial Number																																																												
0.00	SD1019																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	SD1025																																																												

Notes: CALIBRATED BY AGENCY REQUEST TO BRING VALUES CLOSER TO NOMINAL.
QC: SP Quality checks done in September

Brett Henderson
Quality Control Review

 Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use
 Date 11/4/16

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000891
11/03/2016 09:14:33

Auto Calibration
Max Power Res Value = 31
Auto Range Res Value = 8

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12963, Sum Io = 13844

Channel 1 data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1280 (-0.0190)
Sample #2 = 0.0980 (0.0250)
Sample #3 = 0.1070 (0.0850)
Sample #4 = 0.0790 (0.1210)
Avg % Abs = 0.0947 (0.0770)
STD DEV = 0.0143 (0.0485)
REL STD DEV = 15.100 (62.984)

Channel 2 data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1730 (-0.0080)
Sample #2 = 0.1760 (0.0010)
Sample #3 = 0.1580 (0.0290)
Sample #4 = 0.1690 (0.0400)
Avg % Abs = 0.1677 (0.0233)
STD DEV = 0.0091 (0.0201)
REL STD DEV = 5.412 (86.177)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12949, Sum Io = 13839

Channel 1 data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.8020 (0.0080)
Sample #2 = 0.7860 (0.0480)
Sample #3 = 0.7910 (0.0500)
Sample #4 = 0.8020 (0.0700)
Avg % Abs = 0.7930 (0.0560)
STD DEV = 0.0082 (0.0122)
REL STD DEV = 1.032 (21.724)

Channel 2 data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.5730 (0.0040)
Sample #2 = 1.5690 (0.0240)
Sample #3 = 1.5650 (0.0190)
Sample #4 = 1.5640 (0.0320)
Avg % Abs = 1.5660 (0.0250)
STD DEV = 0.0026 (0.0066)
REL STD DEV = 0.169 (26.230)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12891, Sum Io = 13817

Channel 1 data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.8760 (-0.0180)
Sample #2 = 1.8060 (0.0400)
Sample #3 = 1.8120 (0.0670)
Sample #4 = 1.8470 (0.0610)
Avg % Abs = 1.8217 (0.0560)
STD DEV = 0.0221 (0.0142)
REL STD DEV = 1.216 (25.317)

Channel 2 data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.6720 (-0.0210)
Sample #2 = 3.6390 (0.0080)
Sample #3 = 3.6780 (0.0150)
Sample #4 = 3.6710 (0.0200)
Avg % Abs = 3.6627 (0.0143)
STD DEV = 0.0208 (0.0060)
REL STD DEV = 0.568 (42.054)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12909, Sum Io = 13821

Channel 1 data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.5170 (-0.0080)
Sample #2 = 3.4770 (0.0390)
Sample #3 = 3.5180 (0.0500)
Sample #4 = 3.5080 (0.0390)
Avg % Abs = 3.5010 (0.0427)
STD DEV = 0.0214 (0.0064)
REL STD DEV = 0.611 (14.885)

Channel 2 data:
Sample % Abs (% Abs Ref)
Sample #1 = 7.0210 (-0.0100)
Sample #2 = 6.9820 (0.0140)
Sample #3 = 7.0270 (0.0150)
Sample #4 = 7.0280 (0.0090)
Avg % Abs = 7.0123 (0.0127)
STD DEV = 0.0263 (0.0032)
REL STD DEV = 0.375 (25.378)

Sol Value = 0.400 g/210L ***
Fit value = 1.9048 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12865, Sum Io = 13804

Channel 1 data:
Sample % Abs (% Abs Ref)
Sample #1 = 6.6060 (-0.0070)
Sample #2 = 6.6240 (0.0000)
Sample #3 = 6.6020 (0.0280)
Sample #4 = 6.6200 (0.0100)
Avg % Abs = 6.6153 (0.0127)
STD DEV = 0.0117 (0.0142)
REL STD DEV = 0.177 (112.020)

Channel 2 data:
Sample % Abs (% Abs Ref)
Sample #1 = 12.9520 (-0.0140)
Sample #2 = 12.9670 (-0.0080)
Sample #3 = 12.9570 (0.0100)
Sample #4 = 12.9670 (-0.0100)
Avg % Abs = 12.9637 (-0.0027)
STD DEV = 0.0058 (0.0110)
REL STD DEV = 0.045 (413.068)

Auto Cal Data:
Channel 1 data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.095
Std Dev = 0.01 Rel Std Dev = 15.10
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.793
Std Dev = 0.01 Rel Std Dev = 1.03
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.822
Std Dev = 0.02 Rel Std Dev = 1.22
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.501
Std Dev = 0.02 Rel Std Dev = 0.61
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 6.615
Std Dev = 0.01 Rel Std Dev = 0.18
Zero Order Coef = -236.98
First Order Coef = 2659.92
Second Order Coef = 38.47
Standard Deviation = 19.477907

Channel 2 data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.168
Std Dev = 0.01 Rel Std Dev = 5.41
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.566
Std Dev = 0.00 Rel Std Dev = 0.17
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.663
Std Dev = 0.02 Rel Std Dev = 0.57
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 7.012
Std Dev = 0.03 Rel Std Dev = 0.37
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 12.964
Std Dev = 0.01 Rel Std Dev = 0.04
Zero Order Coef = -175.46
First Order Coef = 1280.59
Second Order Coef = 15.54
Standard Deviation = 40.372635

SP

BSK

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0003
0.040	0.040	0.0002
0.100	0.099	0.0005
0.200	0.200	-0.0005
0.400	0.400	0.0001

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0008
0.040	0.039	0.0008
0.100	0.099	0.0008
0.200	0.201	-0.0009
0.400	0.400	0.0002

Sol Value = 0.080 g/210L ***
 Fit value = 0.3818 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1
 Sample #1 = 2963.00
 Sample #2 = 2985.00
 Sample #3 = 2928.00
 Sample #4 = 2980.00
 Average Result = 2964.3333
 STD DEV = 31.5647
 REL STD DEV = 1.065

 ***** CHANNEL 2
 Sample #1 = 3141.00
 Sample #2 = 3166.00
 Sample #3 = 3154.00
 Sample #4 = 3173.00
 Average Result = 3164.3333
 STD DEV = 9.6090
 REL STD DEV = 0.304

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1024
 3 um H2O Adjust (mg/l*10,000) = 845
 9 um H2O Adjust (mg/l*10,000) = 645
 **** AUTO CAL PASS

SP

FWC
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000891
 11/03/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:57
Control Test	0.050	09:58
Air Blank	0.000	09:58
Control Test	0.050	09:59
Air Blank	0.000	10:00
Control Test	0.049	10:00
Air Blank	0.000	10:01
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

P. Murphy
 Operator's Signature

FWC
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000891
 11/03/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:07
Control Test	0.195	10:08
Air Blank	0.000	10:08
Control Test	0.197	10:09
Air Blank	0.000	10:09
Control Test	0.197	10:10
Air Blank	0.000	10:11
Control Test Stats		
Average	0.1963	
Std Dev	0.0012	
Rel Std Dev(%)	0.5881	

P. Murphy
 Operator's Signature

BSK

FWC
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000891
 11/03/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:02
Control Test	0.079	10:03
Air Blank	0.000	10:03
Control Test	0.080	10:04
Air Blank	0.000	10:05
Control Test	0.079	10:05
Air Blank	0.000	10:06
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

P. Murphy
 Operator's Signature

FWC
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000891
 11/03/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:12
Control Test	0.081	10:13
Air Blank	0.000	10:13
Control Test	0.080	10:13
Air Blank	0.000	10:14
Control Test	0.080	10:14
Air Blank	0.000	10:14
Control Test	0.080	10:14
Air Blank	0.000	10:15
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DGS
P. Murphy
 Operator's Signature

EXPERIMENTAL BREATH TESTS, JARROD MOLINAR PRESENT

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000891
10/28/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:49
Subject Sample #1	0.007	06:51
Air Blank	0.000	06:51

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000891
10/28/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:58
Subject Sample #1	0.000	06:59
Air Blank	0.000	06:59

P Murphy

Operator's Signature

P Murphy

Operator's Signature

SP

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000891
10/28/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:53
Subject Sample #1	0.000	06:54
Air Blank	0.000	06:54

P Murphy

Operator's Signature

ASK



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FWC

S/N 80-000891

Date In 10/27/16

Date Out _____

Ship P/U H/D CMI EE

Intake Performed By <u>DB</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Dry gas (walked in)</u> Notes: _____ _____ _____		Quality Checks Performed By _____ <input type="checkbox"/> Breath Tube Screen <input type="checkbox"/> Replace O-Rings <input type="checkbox"/> Instrument Set Up Verified <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Flow Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) <input type="checkbox"/> Barometric Pressure Check Gauge ID # _____ <input type="checkbox"/> Stability Checks <u>PSM 11/3/16</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG619605</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AG619605	Flow Calibration Performed By _____ <input type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)	
Simulator	Serial #	Lot #/Exp																		
0.05	SD1018	201507A 7/14/17																		
0.08	SD1011	201601F 1/26/18																		
0.20	SD1025	201604C 4/5/18																		
0.08 DGS	N/A	AG619605																		
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____																		

RECEIVED
NOV 04 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____ <input type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G4444</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>SD1024</td> <td>16101</td> <td>2/2/18</td> </tr> <tr> <td>0.100</td> <td>SD1013</td> <td>16001</td> <td>5/8/18</td> </tr> <tr> <td>0.200</td> <td>SD1012</td> <td>16103</td> <td>6/14/18</td> </tr> <tr> <td>0.400</td> <td>G6621</td> <td>16102</td> <td>3/22/18</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>03415080A1</td> <td>3/5/17</td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A</td> <td>7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F</td> <td>1/26/2018</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C</td> <td>04/05/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG619605</td> <td>7/14/18</td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	G4444	N/A	N/A	0.040	SD1024	16101	2/2/18	0.100	SD1013	16001	5/8/18	0.200	SD1012	16103	6/14/18	0.400	G6621	16102	3/22/18	0.080 DGS	N/A	03415080A1	3/5/17	Simulator	Serial Number	Lot Number	Expiration	0.05	SD1018	201507A	7/14/17	0.08	SD1011	201601F	1/26/2018	0.20	SD1025	201604C	04/05/18	0.08 DGS	N/A	AG619605	7/14/18	Department Inspection Performed By _____ <input type="checkbox"/> Barometric Pressure _____ Gauge ID# _____ Instrument Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.05</td> <td></td> </tr> <tr> <td>0.08</td> <td></td> </tr> <tr> <td>0.20</td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	0.00		Interferent		0.05		0.08		0.20	
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000	G4444	N/A	N/A																																																												
0.040	SD1024	16101	2/2/18																																																												
0.100	SD1013	16001	5/8/18																																																												
0.200	SD1012	16103	6/14/18																																																												
0.400	G6621	16102	3/22/18																																																												
0.080 DGS	N/A	03415080A1	3/5/17																																																												
Simulator	Serial Number	Lot Number	Expiration																																																												
0.05	SD1018	201507A	7/14/17																																																												
0.08	SD1011	201601F	1/26/2018																																																												
0.20	SD1025	201604C	04/05/18																																																												
0.08 DGS	N/A	AG619605	7/14/18																																																												
Simulator	Serial Number																																																														
0.00																																																															
Interferent																																																															
0.05																																																															
0.08																																																															
0.20																																																															
Attachments <input type="checkbox"/> Form 41 <input type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____		<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																													

Notes: VOIDED

Quality Control Review

Date



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FWC S/N 80-000891

Date In 9/28/16 Date Out 9/28/16 Ship P/U H/D CMI EE

Intake Performed By PWS

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By PWS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 202
 Flow Verification (L/s)
 Flow Column # APP102
 32mm 144 (.139 - .169)
 36mm 164 (.156 - .190)
 53mm 234 (.228 - .278)
 103mm 503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1013	201507A 7/14/17
0.08	DR1279	201601F 1/26/18
0.20	DR3856	201604C 4/5/18
0.08 DGS	N/A	AG162405 5/3/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____
 Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
SEP 29 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By PWS

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge 1013 ID # 26932

Simulator	Serial Number	Lot Number	Expiration
0.000	G2835	N/A	N/A
0.040	SD1024	16101	2/2/18
0.100	SD1012	16001	5/8/18
0.200	G2407	16103	6/14/18
0.400	DR3855	16102	3/22/18
0.080 DGS	N/A	03415080A1	3/5/17

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05	SD1013	201507A	7/14/17
0.08	DR1279	201601F	1/26/18
0.20	DR3856	201604C	4/5/18
0.08 DGS	N/A		

Department Inspection Performed By PWS

Barometric Pressure 1014 Gauge
 ID# 28427 1012 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1016
Interferent	SD1022
0.05	SD1013
0.08	DR1279
0.20	DR3856

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: All stabilities within range. Calibrated Optical Bench to bring values closer to nominal. (PWS)

QC: SP

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
Quality Control Review

9/29/16
Date

Stability Tests - FWC #80-000891 9/28/16
 Pre-Calibration

FWC
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000891
 09/28/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:34
Control Test	0.048	09:35
Air Blank	0.000	09:35
Control Test	0.047	09:36
Air Blank	0.000	09:37
Control Test	0.048	09:37
Air Blank	0.000	09:38
Control Test Stats		
Average	0.0477	
Std Dev	0.0006	
Rel Std Dev(%)	1.2112	

[Signature]
 Operator's Signature

FWC
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000891
 09/28/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:52
Control Test	0.078	09:52
Air Blank	0.000	09:53
Control Test	0.077	09:53
Air Blank	0.000	09:54
Control Test	0.077	09:55
Air Blank	0.000	09:55
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

[Signature]
 Operator's Signature

FWC
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000891
 09/28/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:43
Control Test	0.195	09:44
Air Blank	0.000	09:44
Control Test	0.194	09:45
Air Blank	0.000	09:45
Control Test	0.195	09:46
Air Blank	0.000	09:47
Control Test Stats		
Average	0.1947	
Std Dev	0.0006	
Rel Std Dev(%)	0.2956	

[Signature]
 Operator's Signature

FWC
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000891
 09/28/2016
 Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:59
Control Test	0.077	09:59
Air Blank	0.000	10:00
Control Test	0.079	10:00
Air Blank	0.000	10:01
Control Test	0.077	10:01
Air Blank	0.000	10:02
Control Test Stats		
Average	0.0777	
Std Dev	0.0012	
Rel Std Dev(%)	1.4867	

SP
 78K

[Signature]
 Operator's Signature

Optical Bench
Calibration

FWE
#20000894
9/28/16

Sol Value = 0.0000 mg/l
Fit Value = 0.0000 mg/l
Samples Taken = 4, Discarded = 1
Sum To = 12599, Sum To = 13628
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.7680 (-0.0110)
Sample #2 = 0.7740 (0.0030)
Sample #3 = 0.7840 (0.0410)
Sample #4 = 0.7660 (0.0750)
Avg % Abs = 0.7747 (0.0397)
STD DEV = 0.0090 (0.0360)
REL STD DEV = 1.164 (90.803)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.5380 (-0.0130)
Sample #2 = 1.5110 (0.0010)
Sample #3 = 1.5220 (0.0220)
Sample #4 = 1.4940 (0.0380)
Avg % Abs = 1.5090 (0.0283)
STD DEV = 0.0141 (0.0186)
REL STD DEV = 0.935 (91.260)

Intoxilizer - Alcohol Analyzer
Model 8000
SN 80-000891
09/28/2016 11:16:27

Auto Calibration
Max Power Res Value = 31
Auto Range Res Value = 7

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %
Samples Taken = 4, Discarded = 1
Sum To = 12552, Sum To = 13623
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 1.7940 (-0.0130)
Sample #2 = 1.8200 (-0.0120)
Sample #3 = 1.7990 (0.0190)
Sample #4 = 1.8330 (0.0280)
Avg % Abs = 1.8173 (0.0117)
STD DEV = 0.0172 (0.0210)
REL STD DEV = 0.944 (179.864)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 3.5710 (-0.0020)
Sample #2 = 3.5870 (-0.0060)
Sample #3 = 3.5940 (0.0020)
Sample #4 = 3.5960 (0.0190)
Avg % Abs = 3.5923 (0.0050)
STD DEV = 0.0035 (0.0045)
REL STD DEV = 0.099 (105.45)

Sol Value = 0.0000 mg/l
Fit Value = 0.0000 mg/l
Samples Taken = 4, Discarded = 1
Sum To = 12573, Sum To = 13628
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 3.4780 (-0.0380)
Sample #2 = 3.5020 (-0.0110)
Sample #3 = 3.4750 (0.0210)
Sample #4 = 3.4680 (0.0390)
Avg % Abs = 3.4817 (0.0163)
STD DEV = 0.0180 (0.0253)
REL STD DEV = 0.516 (155.048)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 6.8640 (-0.0070)
Sample #2 = 6.9270 (-0.0150)
Sample #3 = 6.9250 (0.0010)
Sample #4 = 6.9310 (0.0140)
Avg % Abs = 6.9277 (0.0000)
STD DEV = 0.0031 (0.0145)
REL STD DEV = 0.044 (467910416.008)

Sol Value = 0.400 g/210L ***
Fit Value = 1.9048 mg/l %
Samples Taken = 4, Discarded = 1
Sum To = 12579, Sum To = 13628
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 6.5240 (-0.0140)
Sample #2 = 6.5740 (0.0000)
Sample #3 = 6.5690 (-0.0110)
Sample #4 = 6.5910 (0.0210)
Avg % Abs = 6.5780 (0.0030)
STD DEV = 0.0115 (0.0157)
REL STD DEV = 0.175 (523.874)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 12.7660 (-0.0280)
Sample #2 = 12.8240 (0.0020)
Sample #3 = 12.8330 (0.0000)
Sample #4 = 12.8600 (0.0090)
Avg % Abs = 12.8657 (0.0037)
STD DEV = 0.0035 (0.0045)
REL STD DEV = 0.027 (28.69)

Sol Value = 0.0000 mg/l
Fit Value = 0.0000 mg/l
Samples Taken = 4, Discarded = 1
Sum To = 12573, Sum To = 13628
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.03 Rel Std Dev = 117.65
Sample #2 = 0.1905 mg/l or 0.040 g/210L
Sample #3 = 0.775
Sample #4 = 0.200
Sample #5 = 0.400
Avg % Abs = 0.01 Rel Std Dev = 1.16
STD DEV = 1.817
REL STD DEV = 0.02 Rel Std Dev = 0.94
Sample #1 = 0.9524 mg/l or 0.200 g/210L
Sample #2 = 3.482
Sample #3 = 0.02 Rel Std Dev = 0.52
Sample #4 = 1.9048 mg/l or 0.400 g/210L
Sample #5 = 6.578
Sample #6 = 0.01 Rel Std Dev = 0.18
Zero Order Coef = -65.97
First Order Coef = 2579.84
Second Order Coef = 50.10
Standard Deviation = 25.99811

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 0.400 g/210L ***
Sample #2 = 3033.00
Sample #3 = 3104.00
Sample #4 = 3128.00
Average Result = 3088.3333
STD DEV = 49.3997
REL STD DEV = 1.600
Channel 2
Sample #1 = 3122.00
Sample #2 = 3126.00
Sample #3 = 3134.00
Sample #4 = 3187.00
Average Result = 3149.0000
STD DEV = 33.1512
REL STD DEV = 1.053

Sol Value = 0.0000 mg/l
Fit Value = 0.0000 mg/l
Samples Taken = 4, Discarded = 1
Sum To = 12573, Sum To = 13628
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.01 Rel Std Dev = 510.51
Sample #2 = 0.1905 mg/l or 0.040 g/210L
Sample #3 = 1.509
Sample #4 = 0.01 Rel Std Dev = 0.93
Sample #5 = 0.4762 mg/l or 0.100 g/210L
Sample #6 = 3.592
Sample #7 = 0.00 Rel Std Dev = 0.13
Sample #8 = 0.9524 mg/l or 0.200 g/210L
Sample #9 = 6.928
Sample #10 = 0.00 Rel Std Dev = 0.04
Sample #11 = 1.9048 mg/l or 0.400 g/210L
Sample #12 = 12.846
Sample #13 = 0.03 Rel Std Dev = 0.23
Sample #14 = -6.56
Zero Order Coef = 1254.89
First Order Coef = 17.76
Second Order Coef = 21.040243
Standard Deviation = 21.040243

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 12.7660 (-0.0280)
Sample #2 = 12.8240 (0.0020)
Sample #3 = 12.8330 (0.0000)
Sample #4 = 12.8600 (0.0090)
Avg % Abs = 12.8657 (0.0037)
STD DEV = 0.0035 (0.0045)
REL STD DEV = 0.027 (28.69)

Soil Residual

AC	g/210L	Residual
0.000	0.000	0.0006
0.040	0.041	-0.0008
0.100	0.100	-0.0001
0.200	0.200	0.0004
0.400	0.400	-0.0001

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
0.000	0.000	0.0001
0.040	0.040	-0.0005
0.100	0.099	0.0007
0.200	0.200	-0.0003
0.400	0.400	0.0001

Sol Value = 0.000 g/210L ***
Fit Value = 0.3810 mg/l %
Samples Taken = 4, Discarded = 1
Channel 1
Sample #1 = 2970.00
Sample #2 = 3033.00
Sample #3 = 3104.00
Sample #4 = 3128.00
Average Result = 3088.3333
STD DEV = 49.3997
REL STD DEV = 1.600
Channel 2
Sample #1 = 3122.00
Sample #2 = 3126.00
Sample #3 = 3134.00
Sample #4 = 3187.00
Average Result = 3149.0000
STD DEV = 33.1512
REL STD DEV = 1.053

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1013
3 um H2O Adjust (mg/l*10,000) = 721
9 um H2O Adjust (mg/l*10,000) = 660
**** AUTO CAL PRESS

SL
BK

Stability Tests
Post-Calibration

FWC #80-000891

7/28/16

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000891
09/28/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:58
Control Test	0.052	11:59
Air Blank	0.000	11:59
Control Test	0.051	12:00
Air Blank	0.000	12:00
Control Test	0.052	12:01
Air Blank	0.000	12:02
Control Test Stats	0.0517	
Average	0.0006	
Std Dev	1.1175	
Rel Std Dev(%)		

[Signature]

Operator's Signature

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000891
09/28/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:05
Control Test	0.080	12:06
Air Blank	0.000	12:07
Control Test	0.080	12:07
Air Blank	0.000	12:08
Control Test	0.080	12:08
Air Blank	0.000	12:09
Control Test Stats	0.0800	
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Signature]

Operator's Signature

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000891
09/28/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:10
Control Test	0.197	12:11
Air Blank	0.000	12:11
Control Test	0.198	12:12
Air Blank	0.000	12:12
Control Test	0.198	12:13
Air Blank	0.000	12:14
Control Test Stats	0.1977	
Average	0.0006	
Std Dev	0.2921	
Rel Std Dev(%)		

[Signature]

Operator's Signature

FWC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000891
09/28/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:14
Control Test	0.080	12:15
Air Blank	0.000	12:15
Control Test	0.080	12:16
Air Blank	0.000	12:16
Control Test	0.080	12:16
Air Blank	0.000	12:17
Control Test Stats	0.0800	
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Signature]

Operator's Signature

[Handwritten initials]